MIL-R-36025B
17 September 1976
SUPERSEDING
MIL-R-36025A
5 September 1967

MILITARY SPECIFICATION

ROD, IRRIGATOR SUPPORTING

This specification is approved for use by all Departments and Agencies of the Department of Defense

1. SCOPE

1 1 Scope. This specification covers one type of lightweight, telescoping, metal, supporting rod, fitted with a clamp

2 APPLICABLE DOCUMENTS

2 1 Issues of documents The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein

SPECIFICATIONS

FEDERAL

TT-E-485	- Enamel, Semigloss, Rust-Inhibiting
PPP-B-566	- Boxes, Folding, Paperboard.
PPP-B-636	- Boxes, Shipping, Fiberboard.

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to Headquarters, Defense Personnel Support Center, ATTN Directorate of Medical Materiel, Code ATT, 2800 South 20th Street, Philadelphia, Pa. 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter

FSC 6530

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T07:58Z
Check the source to verify that this is the current version before use.

MILITARY

MIL-C-5541 - Chemical Films for Aluminum and Aluminum Alloys.

STANDARDS

FEDERAL

FED-TEST METHOD STD-151 - Metals, Test Methods.
FED-STD-595 - Colors.

MILITARY

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 - Marking for Shipment and Storage.

DRAWINGS

Defense Personnel Support Center (DPSC)

20076 Rod, Irrigator Supporting, Telescoping Type.

(Copies of specifications, standards, drawings, and publications required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following document forms a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposal shall apply.

American Welding Society

Welders Handbook, 6th Edition

(Application for copies may be addressed to the American Welding Society, 345 East 47th Street, New York, NY 10017.)

3. REQUIREMENTS

3.1 First article. When specified in the procurement document, the contractor shall furnish sample unit(s) for the first article inspection and approval (see 4.1.2.1 and 6.2).

3.2 Materials. Materials shall be as specified in DPSC Drawing 20076, of the highest quality throughout, and suitable in all respects for the purpose intended.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T07:58Z
Check the source to verify that this is the current version before use.

* 3.3 Construction. The irrigator supporting rod shall be constructed as shown on DPSC Drawing 20076 and as specified herein. The rod shall be adjustable from approximately 18 inches to 43 inches.

3.4 Finish. All aluminum alloy parts shall be chemically or otherwise treated in accordance with MIL-C-5541 to form a suitable base for subsequent paint finishes. All exposed parts of the supporting rod, except threads, shall be given a coat of primer followed by a coat of semi-gloss olive drab enamel conforming to TT-E-489 and to color 24087 of FED-STD-595. Requirement for QPL source for paint and primer shall not apply.

* 3.5 Identification marking. One face of the clamp jaw of the irrigator supporting rod shall be permanently and legibly marked with the contract or purchase order number and the name or registered trademark of the manufacturer.

* 3.6 Performance. The rod shall have a minimum range of adjustment from 18 to 43 inches. With the thumbscrews not engaged, the upper rod and extension tube shall slide smoothly, free of binding, during the entire length of travel. The thumbscrews shall securely lock the telescoping components in any set position. The stops on the telescoping components shall prevent removal from the assembly.

3.7 Workmanship. Workmanship shall be first class throughout. Irrigator supporting rods shall be free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T07:58Z
Check the source to verify that this is the current version before use.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such a visual and auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2.1 Classification of inspection The inspection requirements specified herein are classified as follows.

(a) First article inspection.
(b) Quality conformance inspection.

4.1.2.1.1 First article inspection First article inspection shall be performed on the rods when a first article sample is required (see 3.1). This inspection shall include examination of 4.3 and tests of 4.4 except chemical composition test, 4.4.1.

4.1.2.1.2 Quality conformance inspection Quality conformance inspection shall include the examination of 4.3 and tests of 4.4.

4.1.3 Certificates of quality Certificates of quality, supplied by the manufacturer of the materials may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demostrated to the Government The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the Item Identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling

4.2.1 For examination Sampling for examination shall be conducted in accordance with MIL-STD-105, and table I. The unit of product for sampling purposes shall be the rod

TABLE I. Sampling for examination

	Inspection level	AQL (defect per hundred units)
For visual examination		
Major defects	II	1.0
Minor defects	II	4.0
For dimensional examination	S-1	2.5

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T07:58Z
Check the source to verify that this is the current version before use.

* 4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and table II. Sample for analysis shall be as specified in 4.2 3.

TABLE II Sampling for tests

Component (Unit of procuct)	Characteristic	Requirement	Test procedure
Material	Chemical composition	3.2	4.4.1
Rod	Performance:		
"	Range of adjustment	3.6	4.4.2.1
"	Smoothness of operation	"	4.4.2.2
"	Locking capability	"	4.4.2.3
"	Prevention of rod and extension tube removal	"	4.4.2.4

4.2 3 Raw materials. A sufficient quantity of each lot of each metal used in the manufacture of the rods shall be selected for chemical analysis.

4.3 Examination The rods shall be examined to determine compliance with all requirements contained in this specification.

4.3.1 Classification of defects Examination shall be conducted in accordance with the classification of defects listed in table III. Examination shall not be restricted to the classified possible defects listed below.

TABLE III Classification of defects

Major	
101	Rod assembly not complete with all specified component parts.
102	Edges and corners of component parts not free of burrs or sharpness.
103	Minimum range of adjustment not 18 to 43 inches.
104	Telescoping component parts 1 and 2 fail to slide easily for entire length of their travel (with thumbscrews in loosened position).
105	Thumbscrews 7 and 8 fail to securely lock telescoping components in any set position.
106	Stops on telescoping components fail to prevent removal from the assembly.
107	Component parts not free of cracks or ruptures.
108	Component parts in assembly not properly fitted and securely fastened.
109	Threaded components 11, 7 and 8 fail to function without seizing or stripping
110	End of handwheel screw not fully recessed below the surfaces of serrated jaw
111	Jaw points not present and not force fitted into jaw to prevent accidental removal.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T07:58Z
Check the source to verify that this is the current version before use.

TABLE III. Classification of defects (continued)

Minor:	
201	Serrated jaw does not contain a minimum of five teeth.
202	Identification marking not complete, permanent and legible, or not present.
203	Enameled finish not free of blisters, peeling, or bare areas

4.3.2 Dimensional examination The rod shall be examined for defects in dimensions. Any dimension not within the tolerances specified herein shall be classified as a defect.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Chemical composition. Four ounces from each lot of material shall be tested in accordance with FED-TEST METHOD STD - 151.

4.4.2 Performance With the rod completely assembled, the following function operations shall be performed.

4.4.2.1 Range of adjustment. Unloosen the thumbscrews allowing the upper rod and extension tube to slide Move the upper rod and extension tube to the unextended position Measure the length of the rod at this adjustment Move the upper rod and extension tube to the fully extended position. Measure the length of the rod at this adjustment

4.4.2.2 Smoothness of operation Disengage the thumbscrews allowing the telescoping components to slide upward and downward Any binding during upper rod or extension tube travel shall be noted

4.4.2.3 Locking capability The telescoping components shall be raised 6 inches from the unextended position The thumbscrews shall be tightened to secure the components in position A 5 pound downward force shall be exerted on the top of the rod The rod shall then be examined for slippage after this test.

4.4.2.4 Prevention of rod and extension tube removal Disengage the thumbscrews allowing the telescoping components to slide upward and downward Grasp the stationary tube and exert a 5 pound upward force on the upper rod by pulling upward on the upper rod Examine the rod for failure of the collars to prevent removal of either the upper rod or extension tube from the assembly

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T07:58Z
Check the source to verify that this is the current version before use.

4.5 Packaging inspection. The inspection of the preservation -packaging and interior package marking shall be in accordance with the group A and B quality conformance inspection requirements of MIL-P-116 The inspection of the packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129

5. PACKAGING

5.1 Preservation and packaging Preservation and packing shall be Level A or C as specified (see 6.2).

5.1.1 Unit of issue. One rod as specified, constitutes one unit of issue.

5.1.2 Level A.

5.1.2.1 Cleaning. Rods shall be cleaned in accordance with MIL-P-116, method C-1.

5.1.2.2 Drying Rods shall be dried in accordance with MIL-P-116

5.1.2.3 Preservative application Preservatives shall not be applied to bars

5.1.2.4 Unit packaging Each rod shall be disassembled and the disassembled components wrapped in kraft paper and packaged in accordance with MIL-P-116, method III in a box conforming to PPP-B-566, 3 flute corrugated paperboard or PPP-B-636, class domestic Closure shall be as specified in the appendix of the box specification.

5.1.3 Level C Each rod shall be packaged in a manner that will afford adequate protection against corrosion, deterioration and physical damage during shipment from the supply source to the first receiving activity

5.2 Packing Packing shall be Level A, B or C as specified (see 6.2).

5.2.1 Level A Twenty-four unit packages shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-636, class weather-resistant Closure, strapping and waterproofing shall be as specified in the appendix of the box specification

5.2.2 Level B Twenty-four packages shall be packed in an exterior container designed for a type 2 load and conforming to PPP-B-636, class domestic Closure shall be as specified in the appendix of the box specification

5.2.3 Level C. The packaged rods shall be packed in shipping containers that will afford adequate protection against damage during direct shipment from the supply source to the first receiving activity These packs shall conform to the applicable carrier's rules and regulations

5.3 Marking. Each unit package, and exterior container shall be marked as specified in MIL-STD 129.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T07:58Z
Check the source to verify that this is the current version before use.

6. NOTES

6.1 Intended use. This rod is intended for use in supporting intravenous solution containers. The rod is equipped with a clamp for securing it to stretchers and beds.

6.2 Ordering data. Purchasers should select the preferred options permitted herein and include the following information in procurement documents:

(a) Title, number and date of this specification.
(b) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).
(c) When a first article is required (see 3.1, 4.1.2.1 and 6.3).

6.3 First article. When a first article is required, it shall be tested and approved under the appropriate provisions of 7-104.55 of the Armed Services Procurement Regulation. The contract or purchase order specifies the quantity, procedure for delivery and other specific instructions for inspection and approval of the first article.

6.4 This specification covers the following item appearing in the Federal Supply Catalog:

National Stock Number	Item Identification
6530-00-792-6000	ROD, IRRIGATOR SUPPORTING, Telescoping Type

6.5 Changes from previous issue. The margins of this specification are marked with an asterisk to indicate where changes (additions, modifications, corrections, deletions) from the previous issue were made. This was done as a convenience only and the Government assumes no liability whatsoever for any inaccuracies in these notations. Bidders and contractors are cautioned to evaluate the requirements of this document based on the entire content irrespective of the marginal notations and relationship to the last previous issue.

MILITARY CUSTODIANS
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
DSA-DM

Project No. 6530-1390

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T07:58Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No 22-R255

INSTRUCTIONS The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay or will otherwise enhance use of the document DoD contractors government activities or manufacturers vendors who are prospective suppliers of the product are invited to submit comments to the government Fold on lines on reverse side staple in corner, and send to preparing activity Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements Attach any pertinent data which may be of use in improving this document If there are additional papers attach to form and place both in an envelope addressed to preparing activity

DOCUMENT IDENTIFIER AND TITLE
MIL-R-36025B ROD, IRRIGATOR SUPPORTING

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A
☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1 HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?
A GIVE PARAGRAPH NUMBER AND WORDING

B RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2 COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3 IS THE DOCUMENT RESTRICTIVE?
☐ YES ☐ NO *(If "Yes", in what way?)*

4 REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO

DATE

DD FORM 1 JAN 72 1426 REPLACES EDITION OF 1 JAN 68 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T07:58Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, PA 19101

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T07:58Z
Check the source to verify that this is the current version before use.