MIL-R-0036037A(DM)
20 August 1964
USED IN LIEU OF
MIL-R-36037
21 June 1960

MILITARY SPECIFICATION

RACKS, DRYING, CLINICAL UTENSILS

This limited coordination Military specification has been prepared by the Defense Medical Supply Center based upon currently available technical information, but it has not been approved for promulgation as a coordinated revision of Military Specification MIL-R-36037. It is subject to modification. However, pending its promulgation as a coordinated Military specification, it may be used in procurement.

1. SCOPE

1.1 Scope. This specification covers field-type clinical-utensil-drying racks, for use in drying bedpans, urinals, pitchers, basins, and other utensils after they have been washed and sterilized.

1.2 Classification. Utensil-drying racks covered by this specification shall be of the following styles, as specified (see 6.1):

Style No. 1.

Style No. 2.

6530

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:00Z
Check the source to verify that this is the current version before use.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

QQ-A-274	Aluminum Alloy Bars, Rods, and Structural and Special Shaped Sections, Extruded, 6063, (63S).
QQ-A-318	Aluminum Alloy Plate and Sheet 5052.
PPP-B-585	Boxes; Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.
PPP-B-676	Boxes, Set-Up, Paperboard.
PPP-T-76	Tape, Pressure-Sensitive Adhesive Paper, Water Resistant (for Carton Sealing).

Military

MIL-A-8625	Anodic Coatings, for Aluminum and Aluminum Alloys.
MIL-L-10547	Liners, Case, and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.
MS20257	Hinge, Continuous.
MS20450	Rivet, Tubular.

Air Force - Navy Aeronautical

AN960	Washer, Flat.

DRAWING

Defense Medical Supply Center

6020	Racks, Drying, Clinical Utensils.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:00Z
Check the source to verify that this is the current version before use.

(Copies of specifications, standards, and drawings required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Materials. All materials used in the fabrication of the utensil-drying racks shall conform to the applicable specification or standard as indicated in Table I.

3.2 Style and design. The utensil-drying racks shall be folding-type, aluminum racks. The racks shall consist of an inverted V-type frame with crossbars and a folding-type stretcher at the base.

3.2.1 Style No. 1. Style No. 1 shall conform strictly to the dimensions, design, fabrication, and materials of No. 1 rack, as specified in Drawing 6020.

3.2.2 Style No. 2. Style No. 2 shall conform strictly to the dimensions, design, fabrication, and materials of No. 2 rack, as specified in Drawing 6020.

3.3 Finish. After assembly, the racks shall be alkaline cleaned, anodized by the chromic-acid process in accordance with MIL-A-8625, and hot-water sealed. The resultant anodic film shall be clear to translucent, and colorless.

3.4 Identification marking. Each utensil-drying rack shall be marked with the following information:

Stock number

Manufacturer's name or registered trademark

Contract or order number

U.S.

The marking shall be legible and permanent and shall be accomplished by die stamping or by the use of nameplate(s) permanently attached. Marking secured by adhesives shall not be employed.

3.5 Workmanship. Workmanship shall be first class throughout. The utensil-drying racks shall be free from defects which detract from their appearance or may impair their serviceability.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:00Z
Check the source to verify that this is the current version before use.

TABLE I - MATERIALS

PART NO.	NOMENCLATURE	RACK 1 REQUIRED	RACK 2 REQUIRED	COMMERCIAL DESIGNATION	SPECIFICATION
1	Cross brace member, narrow frame	3	3	Alum. 5052-H34	QQ-A-318
2	Cross brace member, wide frame	5	4	Alum. 5052-H34	QQ-A-318
3	Cross brace member, narrow frame (Rack No. 2)	-	1	Alum. 5052-H34	QQ-A-318
4	Side member, wide frame	2	2	Alum. 6063-T5	QQ-A-274
5	Top member, narrow frame	1	1	Alum. 6063-T5	QQ-A-274
6	Top member, wide frame	1	1	Alum. 6063-T5	QQ-A-274
7	Bottom member, narrow frame	1	1	Alum. 6063-T5	QQ-A-274
8	Bottom member, wide frame	1	1	Alum. 6063-T5	QQ-A-274
9	Divider, narrow	-	2	Alum. 5052-H34	QQ-A-318
10	Divider, wide	-	1	Alum. 5052-H34	QQ-A-318
11	Brace, rack	4	4	Alum. 5052-H34	QQ-A-318
12	Bracket, brace	2	2	Alum. 6063-T5	QQ-A-274
13	Hook, frame	4	4	Alum. 5052-H34	QQ-A-318
14	Side member, narrow frame (as shown)	1	1	Alum. 6063-T5	QQ-A-274
15	Side member, narrow frame (opp. hd.)	1	1	Alum. 6063-T5	QQ-A-274
-	Rivet, tubular, oval head 13/32 lg.	2	2	Steel, AISI 1010 or 1015	MS 20450
-	Rivet, tubular, oval head 11/32 lg.	2	2	Steel, AISI 1010 or 1015	MS 20450
-	Rivet, tubular, oval head 9/32 lg.	2	2	Steel, AISI 1010 or 1015	MS 20450
-	Washers	8	8	Corrosion resisting steel	AN960-C4
-	Hinge, continuous	1	1	Alum. Alloy	MS 20257-2-1100

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:00Z
Check the source to verify that this is the current version before use.

3.5.1 Welding. All welding shall be done with suitable equipment and by skilled welders, thoroughly experienced in the operations of aluminum welding. Spotwelds shall be sound welds of sufficient strength to tear the spot out of the parent metal, when pulled apart, when tested as specified in 4.4.1. All exposed spotwelds shall be ground smooth of flash and burs.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual, auditory, etc., investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105.

	Inspection level	AQL (percent defective)
For visual examination		
Major defects	II	1.0
Total (major and minor combined) defects	II	4.0
For dimensional examination	S-3	2.5

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:00Z
Check the source to verify that this is the current version before use.

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, and as follows:

Component	Characteristic	Req.	Test procedure	Insp. level	AQL
Material (aluminum alloy)	Chemical analysis Physical analysis	3.1	QQ-A-274 or QQ-A-318		
Rack	Welds	3.5.1	4.4.1	S-1	*

* Acceptance number shall be zero (0) for all sample sizes. Not less than two samples shall be taken regardless of lot size.

4.2.3 Components. Sufficient samples of the components shall be selected from each lot of material used in the manufacture of the racks to perform tests as required by the applicable subsidiary specifications or by tests specified herein. When sample size for components covered by the subsidiary specifications is not specified, the sample size for each test shall be as indicated above. No test result shall fall below the minimum or exceed the maximum values specified. The sample unit for testing shall be the quantity necessary to perform all required tests one time each. Not less than three samples shall be taken regardless of lot size. Lot sizes are expressed as follows:

Rivets. . . . . .each
Washers . . . . . .each
Hinges. . . . . .each

4.2.4 Where possible, the same sample unit shall be used for the determination of two or more test characteristics.

4.3 Examination. The racks shall be examined to determine compliance with all requirements contained in this specification.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:00Z
Check the source to verify that this is the current version before use.

4.3.1 Classification of defects. Examination shall be conducted in accordance with the following classification of defects:

Categories and defects*

Major

101	Rack is not style and design specified.
102	Component parts improperly aligned.
103	Rack cannot be readily set up.
104	Rack does not fold flat.
105	Rack is not free of cracks, tears, ruptures, and deformed areas.
106	Acessible edges, corners, and surfaces are not free of burs, sharpness, and sharp projections.
107	Component parts are not securely spotwelded.
108	Spotwelds are not free of burs, porosity, and flash.

Minor

251	Hinge binds in operation.
252	Hinge pins are not secured at each end of the hinge.
253	Rivets are not properly fastened.
254	Finish is not an anodized finish as specified in 3.3.
255	Finish is not uniform, and is not free from coarse grind marks and scratched areas.
256	Identification marking is not present, complete, legible, correct, or permanently affixed.

* Examination shall not be restricted to the classified possible defects listed above.

4.3.2 Dimensional examination. The racks shall be examined for defects in dimensions. Any dimension not within the tolerances specified herein shall be classified as a defect.

4.4 Tests. Tests shall be conducted **to determine compliance with specification requirements.**

4.4.1 Spotwelds. Spotwelds shall be tested by exerting such force on the welded members that they will be lifted or pried apart. **Welded** areas shall be examined to determine compliance with requirements specified in 3.5.1.

4.4.2 Materials. Materials shall be tested in accordance with the quality assurance provisions of their respective specifications.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:00Z
Check the source to verify that this is the current version before use.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One (1) rack, as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in table I shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table I. Packaging Quantities

Unit package	Intermediate package	Exterior container
1 unit	Not required	18 units

5.1.2.1 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Unit package. Unit package shall be a carton of appropriate size constructed in accordance with PPP-B-676, except that commercial colors shall be acceptable, or PPP-B-636. Carton shall be designed with suitable interior fittings or shall contain sufficient cushioning material to insure against shifting of, or damage to, the contents. Closure shall be adequate to prevent spillage of contents under normal handling.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for type 2 load, and constructed in accordance with PPP-B-585, style 3, for class 3 use; PPP-B-601, table II, using type III, class 1 plywood; PPP-B-621, table III; or PPP-B-636, type I, class 2.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:00Z
Check the source to verify that this is the current version before use.

5.2.1.2 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes.

5.2.1.3 Closure. Closure of wood boxes shall be in accordance with appendix of applicable box specification. Closure of each fiberboard box shall be as specified in the appendix of PPP-B-636. In addition, all seams, including manufacturer's joint, shall be closed by tape conforming to the requirements of PPP-T-76. The tape shall be not less than 2 inches wide. Three strips shall be applied to both top and bottom so that all seams and joints shall be covered their full length. Tape applied over the lengthwise seams shall extend a minimum of 2 inches onto each of the end panels. Tape applied over the end edge seams shall extend a minimum of 2 inches along the side and top, and side and bottom panels.

5.2.1.4 Strapping. Strapping, when required, shall be in accordance with appendix of applicable box specification.

NOTE: Strapping shall not be required for shipments forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall conform to PPP-B-601, table I, using type III, class 1 plywood; or shall be designed for type 2 load, and constructed in accordance with PPP-B-585, style 3, for class 2 use; PPP-B-621, table I; or PPP-B-626, type I, class 1.

5.3 Marking.

5.3.1 Interior packages. Each unit package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129. In addition, the letter "L" shall precede the Federal stock number. One space shall be allowed between the letter "L" and the first digit of the stock number. The terms "Stock number" or "FSN" shall not be indicated.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:00Z
Check the source to verify that this is the current version before use.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.

(b) Style required (see 1.2).

(c) Selection of applicable levels of packing (see 5.1 and 5.2).

6.2 This specification does not include all types, classes, grades, sizes, etc., of the commodity indicated by the title of this specification, or which are commercially available, but is intended to cover the types, etc., which are normally procured to meet Military requirements.

Custodians:
Army. - MD
Navy. - MS
Air Force. - 03

Review activities:
Army. - MD
Navy. - MS
Air Force. - 03

Preparing activity:
Defense Supply Agency. - DM

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:00Z
Check the source to verify that this is the current version before use.

CAUTION

NOTICE TO BIDDERS/OFFERORS

DO NOT CONDITION OR BASE YOUR BID/OFFER ON ANY CURRENT PROCUREMENT ON THE INFORMATION SUBMITTED ON THIS FORM SINCE ANY CHANGES OR DELETIONS IN THE SPECIFICATIONS MAY RENDER YOUR BID/OFFER NON-RESPONSIVE IN WHICH CASE IT CANNOT BE CONSIDERED FOR AWARD.

FOLD UNDER

DEFENSE SUPPLY AGENCY
DEFENSE MEDICAL SUPPLY CENTER
3RD AVENUE AND 29TH STREET
BROOKLYN, N.Y., 11232

OFFICIAL BUSINESS

POSTAGE FEES PAID
DEFENSE SUPPLY AGENCY

DEFENSE SUPPLY AGENCY

DEFENSE MEDICAL SUPPLY CENTER

ATTN: DIRECTORATE OF TECHNICAL OPERATIONS (CODE TM5)

3RD AVENUE AND 29TH STREET

BROOKLYN, N.Y., 11232

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:00Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

FORM APPROVED BUDGET BUREAU NO. 119-R004

INSTRUCTIONS

This sheet is to be filled out by personnel either Government or contractor, involved in the use of the specification in procurement of products for ultimate use of the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity.

SPECIFICATION MILITARY SPECIFICATION MIL-R-0036037A(DM), dated 20 August 1964, RACKS, DRYING, CLINICAL UTENSILS

ORGANIZATION	CITY AND STATE

CONTRACT NO.	QUANTITY OF ITEMS PROCURED	DOLLAR AMOUNT
		$

MATERIAL PROCURED UNDER A ☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?
   A. GIVE PARAGRAPH NUMBER AND WORDING

   B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?
   ☐ YES ☐ NO IF "YES" IN WHAT WAY?

4. REMARKS (ATTACH ANY PERTINENT DATA WHICH MAY BE OF USE IN IMPROVING THIS SPECIFICATION. IF THERE ARE ADDITIONAL PAPERS, ATTACH TO FORM AND PLACE BOTH IN AN ENVELOPE ADDRESSED TO PREPARING ACTIVITY)

SUBMITTED BY (PRINTED OR TYPED NAME AND ACTIVITY)	DATE

DD FORM 1426

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:00Z
Check the source to verify that this is the current version before use.