MIL-R-36037B
25 October 1967
SUPERSEDING
MIL-R-0036037A(DM)
20 August 1964 and
MIL-R-36037
20 June 1960

MILITARY SPECIFICATION

RACKS, DRYING, CLINICAL UTENSILS

This specification is mandatory for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers field-type clinical utensil drying racks; for use in drying bedpans, urinals, pitchers, basins, and other utensils after they have been washed and sterilized.

1.2 Classification. Utensil-drying racks covered by this specification shall be of the following styles, as specified (see 6.1).

Style No. 1
Style No. 2

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of the specification to the extent specified herein:

SPECIFICATIONS

Federal

QQ-A-200/9B	Aluminum Bars, Rods, and Structural and Special Shaped Sections, Extruded 6063/63S.
QQ-A-250/8C	Aluminum Alloy Plate and Sheet 5052.
PPP-B-566	Boxes, Folding, Paperboard.
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.

FSC 6530

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:00Z
Check the source to verify that this is the current version before use.

PPP-B-676	Box, Setup.
PPP-T-76	Tape, Pressure-Sensitive Adhesive Paper, (For Carton Sealing).

Military

MIL-A-8625	Anodic Coatings for Aluminum and Aluminum Alloys.
MIL-L-10547	Liners, Case, and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.
MS16535	Rivet, Tubular, Oval Head
MS20257	Hinge, Continuous.

Air Force-Navy Aeronautical

AN960	Washer, Flat.

DRAWING

Defense Personnel Support Center (DPSC)

6020	Racks, Drying, Clinical Utensils.

(Copies of specifications, standards, and drawings required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.

3. REQUIREMENTS

3.1 First Article. Unless otherwise specified, prior to the start of production of the contract quantity, one sample of the rack shall be assembled, completely finished, and tested by the contractor. The sample shall be submitted, together with the contractor's inspection and test reports, to the procuring contracting officer for examination and testing to determine compliance with the requirements of this specification. Written

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:00Z
Check the source to verify that this is the current version before use.

approval of the materials, construction, and performance of the sample shall be obtained from the procuring contracting officer, or his authorized representative, prior to the manufacture and assembly of the remaining quantity to be manufactured. Pending approval of the first article, acquisition of materials, components, or commencement of production is at the sole risk of the contractor. Accepted sample shall be returned to the cognizant Quality Assurance Representative for retention and use as a guide for inspection during production.

3.2 Material. All material used in the fabrication of the utensil drying racks shall conform to the applicable specification or standard as indicated in Table I.

3.3 Style and design. The utensil-drying racks shall be folding type racks fabricated from aluminum. The racks shall consist of an inverted V-type frame with crossbars and a folding-type stretcher at the base. The racks shall be as shown in drawing DPSC 6020 and as hereinafter specified.

3.3.1 Style No. 1. Style No. 1 shall conform strictly to the dimensions, style, design, fabrication, and materials of No. 1 rack as specified in the drawing.

3.3.2 Style 2. Style 2 rack shall be in strict accordance with the drawing for design, dimensions, material, and fabrication.

3.4 Finish. After assembly the racks shall be alkaline cleaned, anodized by the chromic acid process in accordance with MIL-A-8625, and hot-water sealed. The resultant anodic film shall be clear to translucent, and colorless.

3.5 Identification marking. Each utensil drying rack shall be marked with the following information:

Stock number
Manufacturer's name or registered trademark.
Contract or order number.
U.S.

The marking shall be legible and permanent, and shall be accomplished by die stamping or by the use of nameplates permanently attached. Marking secured by adhesives shall not be employed.

3.6 Workmanship. Workmanship shall be first class throughout. The utensil drying racks shall be free from defects which detract from their appearance or may impair their serviceability.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:00Z
Check the source to verify that this is the current version before use.

Any exposed or accessible edge, corner, or surface not true at

Table I. Materials

PART NO.	NOMENCLATURE	RACK 1 REQUIRED	RACK 2 REQUIRED	COMMERCIAL DESIGNATION	SPECIFICATION
1	Cross brace member, narrow frame	3	3	Alum. 5052-H34	QQ-A-250/8B
2	Cross brace member, wide frame	5	4	Alum. 5052-H34	QQ-A-250/8B
3	Cross brace member, narrow frame (Rack No. 2)	-	1	Alum. 5052-H34	QQ-A-250/8B
4	Side member, wide frame	2	2	Alum. 6063-T5	QQ-A-200/9A
5	Top member, narrow frame	1	1	Alum. 6063-T5	QQ-A-200/9A
6	Top member, wide frame	1	1	Alum. 6063-T5	QQ-A-200/9A
7	Bottom member, narrow frame	1	1	Alum. 6063-T5	QQ-A-200/9A
8	Bottom member, wide frame	1	1	Alum. 6063-T5	QQ-A-200/9A
9	Divider, narrow	-	2	Alum. 5052-H34	QQ-A-250/8B
10	Divider, wide	-	1	Alum. 5052-H34	QQ-A-250/8B
11	Brace, rack	4	4	Alum. 5052-H34	QQ-A-250/8B
12	Bracket, brace	2	2	Alum. 6063-T5	QQ-A-200/9A
13	Hook, frame	4	4	Alum. 5052-H34	QQ-A-250/8B
14	Side member, narrow frame (as shown)	1	1	Alum. 6063-T5	QQ-A-200/9A
15	Side member, narrow frame (opp. hd.)	1	1	Alum. 6063-T5	QQ-A-200/9A
-	Rivet, tubular, oval head 13/32 lg.	2	2	Steel, Carbon	MS16535
-	Rivet, tubular, oval head 11/32 lg.	2	2	Steel, carbon	MS16535
-	Rivet, tubular, oval head 9/32 lg.	2	2	Steel, carbon	MS16535
-	Washers	8	8	Corrosion resisting steel	AN960-C4
-	Hinge, continuous	1	1	Alum. Alloy	MS20257

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:00Z
Check the source to verify that this is the current version before use.

3.6.1 Welding. All welding shall be done with suitable equipment and by skilled welders, thoroughly experienced in the operations of aluminum welding. Spotwelds shall be sound welds of sufficient strength to tear the spot out of the parent metal, when pulled apart, when tested as specified in 4.4.1. All exposed spotwelds shall be ground smooth of flash and burrs.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual and auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105.

	Inspection level	AQL (percent defective)
For visual examination		
Major defects	II	1.0
Total (major and minor combined) defects	II	4.0
For dimensional examination	S-3	2.5

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:00Z
Check the source to verify that this is the currentversion before use.

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, and as follows:

Component	Characteristic	Req.	Test procedure	Insp. level	AQL
Material (aluminum alloy)	Chemical analysis Physical analysis	3.1	QQ-A-250/8B		
Rack	Welds	3.5.1	4.4.1	S-1	*

* Acceptance number shall be zero for all sample sizes. Not less than two samples shall be taken regardless of lot size.

4.2.3 Components. Sufficient samples of the components shall be selected from each lot of material used in the manufacture of the racks to perform tests as required by the applicable subsidiary specifications or by tests specified herein. When sample size for components covered by the subsidiary specifications is not specified, the sample size for each test shall be as indicated above. No. test result shall fall below the minimum or exceed the maximum values specified. The sample unit for testing shall be the quantity necessary to perform all required tests one time each. Not less than three samples shall be taken regardless of lot size. Lot sizes are expressed as follows:

Rivets.....each
Washers.....each
Hinges.....each

4.2.4 Where possible, the same sample unit shall be used for the determination of two or more test characteristics.

4.3 Examination. The racks shall be examined to determine compliance with all requirements contained in this specification.

4.3.1 Classification of defects. Examination shall be conducted in accordance with the following classification of defects:

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:00Z
Check the source to verify that this is the current version before use.

Categories and defects*

Major

101	Rack is not style and design specified.
102	Component parts improperly aligned.
103	Rack cannot be readily set up.
104	Rack does not fold flat.
105	Rack is not free of cracks, tears, ruptures, and deformed areas.
106	Accessible edges, corners, and surfaces are not free of burs, sharpness, and sharp projections.
107	Component parts are not securely spotwelded.
108	Spotwelds are not free of burs, porosity, and flash.

Minor

251	Hinge binds in operation.
252	Hinge pins are not secured at each end of the hinge.
253	Rivets are not properly fastened.
254	Finish is not an anodized finish as specified in 3.3.
255	Finish is not uniform, and is not free from coarse grind marks and scratched areas.
256	Identification marking is not present, complete, legible, correct, or permanently affixed.

* Examination shall not be restricted to the classified possible defects listed above.

4.3.2 Dimensional examination. The racks shall be examined for defects in dimensions. Any dimension not within the tolerance specified herein shall be classified as a defect.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Spotwelds. Spotwelds shall be tested by exerting such force on the welded members that they will be lifted or pried apart. Welded areas shall be examined to determine compliance with requirements specified in 3.6.1.

4.4.2 Materials. Materials shall be tested in accordance with the quality assurance provisions of their respective specifications.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:00Z
Check the source to verify that this is the current version before use.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One rack, as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in the following table shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table Packaging Quantities

Unit package	Intermediate package	Exterior container
1 unit	Not required	18 units

5.1.2.1 Packing variation permitted. If the required number of units in the entire shipment is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Unit package. Unit package shall be a box of appropriate size and design constructed in accordance with PPP-B-636, type CF, class weather-resistant. Boxes shall be closed by taping the top and bottom flaps with 2 inch wide tape conforming to PPP-T-76. Closing of bottom flap with adhesive conforming to MMM-A-250 or metal stitches is permitted.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

Source: https://assist.dla.mil -- Downloaded: 20261O-21T08:00Z
Check the source to verify that this is the current version before use.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for type 2 load, and constructed in accordance with PPP-B-585, style, 3, for class 3 use; PPP-B-601, table II, using type I, class 1 plywood; PPP-B-621, table III; or PPP-B-636, type CF, class weather-resistant. Closure and strapping shall be in accordance with appendix of applicable box specification.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall conform to PPP-B-601, table I, using type III, class 1 plwyood; or shall be designed for type 2 load and constructed in accordance with PPP-B-585, style 3, for class 2 use; PPP-B-621, table I; or PPP-B-636, type CF, class domestic. Closure shall be in accordance with appendix of applicable box specification.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Interior packages. Each unit package and each intermediate package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:00Z
Check the source to verify that this is the current version before use.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.

(b) Style required (see 1.2).

(c) Selection of applicable levels of packing (see 5.1 and 5.2).

(d) When First Article is not required (see 3.1).

(e) Where First Article shall be sent (see 3.1).

6.2 This specification does not include all types, classes, grades, sizes, of the commodity indicated by the title of this specification, or which are commercially available, but is intended to cover the types, which are normally procured to meet Military requirements.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Category N.

Preparing activity:
Defense Supply Agency - DM

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T08:00Z
Check the source to verify that this is the current version before use.

CAUTION

NOTICE TO BIDDERS/OFFERORS

DO NOT CONDITION OR BASE YOUR BID/OFFER ON ANY CURRENT PROCUREMENT ON THE INFORMATION SUBMITTED ON THIS FORM SINCE ANY CHANGES OR DELETIONS IN THE SPECIFICATIONS MAY RENDER YOUR BID/OFFER NON-RESPONSIVE IN WHICH CASE IT CANNOT BE CONSIDERED FOR AWARD.

Fold

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

DEFENSE SUPPLY AGENCY
OFFICIAL BUSINESS

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

Fold

DPSC FORM 2418
AUG 66

Source: https://assist.dla.mil -- Downloaded: 202640-21T08:00Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 22-R255

INSTRUCTIONS: This sheet is to be filled out by personnel, either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments and suggestions submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or serve to amend contractual requirements.

SPECIFICATION

MIL-R-36037B

ORGANIZATION

CITY AND STATE

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO *(If "yes", in what way?)*

4. REMARKS *(Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)*

SUBMITTED BY *(Printed or typed name and activity - Optional)*

DATE

DD FORM 1 JAN 66 1426

REPLACES EDITION OF 1 OCT 64, WHICH MAY BE USED.

Source: https://assist.dla.mil -- Downloaded: 202610-21T08:00Z
Check the source to verify that this is the current version before use.