MILITARY SPECIFICATION

STRAP, WEBBING, LITTER CARRYING, OLIVE DRAB

This specification has been approved by the Department of Defense and is mandatory for use by the Departments of the Army, the Navy, and the Air Force.

1. SCOPE

1.1 This specification covers one type of webbing strap, designed to aid the litter bearer in carrying standard field litters (see 6.1).

2. APPLICABLE DOCUMENTS

2.1 The following documents, of the issue in effect on date of invitation for bids, form a part of this specification.

SPECIFICATIONS

FEDERAL

V-T-276 — Thread, Cotton.

CCC-T-191 — Textile Test Methods.

DDD-S-751 — Stitches; Seams; and Stitching

PPP-B-566 — Boxes, Folding, Paperboard

PPP-B-585 — Boxes, Wood, Wirebound.

PPP-B-601 — Boxes, Wood, Cleated-Plywood.

PPP-B-621 — Boxes, Wood, Nailed and Lock-Corner.

PPP-B-676 — Boxes, Set-Up, Paperboard.

MILITARY

MIL-W-530 — Webbing, Textile, Cotton, General Purpose, Natural or in Colors

MIL-T-3530 — Treatment, Mildew-Resistant, for Thread and Twine.

MIL-L-3851 — Loops, Slide (for Equipage).

MIL-B-10377 — Box, Wood, Cleated, Veneer, Paper Overlaid.

MIL-L-10547 — Liners, Case, Waterproof.

MIL-L-11075 — Loops, Strap.

STANDARDS

FEDERAL

Fed Test Method Std. No. 151 — Metals; Test Methods.

MILITARY

MIL-STD-105 — Sampling Procedures and Tables for Inspection by Attributes

MIL-STD-129 — Marking for Shipment and Storage.

DRAWING

MILITARY MEDICAL SUPPLY AGENCY

1873 — Strap, Webbing, Litter Carrying.

FSC 6530

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T07:58Z
Check the source to verify that this is the current version before use.

(Copies of specifications, standards, and drawings required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer).

2.2 **Other publication.** The following document forms a part of this specification. Unless otherwise indicated, the issue in effect on date of invitation for bids shall apply.

COLOR ASSOCIATION OF THE UNITED STATES, INC.

United States Army Standard Color Card for the Official Standardized Shades of Sewing Threads.

(Application for copies should be addressed to the Color Association of the United States, Inc., 200 Madison Avenue, New York 16, New York)

3. REQUIREMENTS

3.1 Material.

3.1.1 *Cotton webbing* Webbing shall conform to type IIa class 4, 2-inch width, vat-dyed olive drab No. 7, of Specification MIL-W-530.

3.1.2 *Cotton thread* Thread for all stitching shall conform to type I, 4-ply, ticket No. 12, of Specification V-T-276. Type IB3 shall be used for the needle, and type IA3 may be used for the bobbin.

3.1.2.1 *Color and treatment* The thread shall be dyed to match olive drab shade S-1 (cable No. 66022) of the United States Army Standard Color Card for the Official Standardized Shades of Sewing Threads, and shall be mildew-resistant in accordance with class 1 treatment of Specification MIL-T-3530. The color of the thread after mildew-resistant treatment will be considered acceptable.

3.1.3 *Slide loop* Slide loop shall conform to size 2 inch, and may be class 1, style 1 or 2; class 2, style 1; or class 3, style 2, of Specification MIL-L-3851. Preproduction samples are not required.

3.1.4 *Strap loop.* Strap loop shall conform to class 1 or 3, 2 by 1/4 inches, symbol E, of Specification MIL-L-11075. Preproduction samples are not required.

3.1.5 *Rivets* Rivets shall be flathead tubular brass rivets with caps, of a size suitable for the purpose intended. Rivets shall be given a dark oxidized finish.

3.2 **Construction.** Litter carrying strap shall be constructed in accordance with Drawing 1873 and this specification. The strap shall be made of two pieces of webbing, and shall be adjustable from 42 inches to not less than 72 inches.

3.2.1 *Stitching* All stitching shall be machine stitching conforming to type 301 of Specification DDD-S-751. There shall be eight to ten stitches to the inch. When machine stitching, type 301, is not a continuous thread, it shall be backstitched not less than 1 inch at each break. All stitching shall be backstitched at the ends to prevent raveling, except where ends are turned under in a hem or held down by other stitching. All seams shall be straight and evenly spaced. All raw ends of webbing shall be turned under and stitched with a double row of stitching.

3.3 **Identification marking.** The manufacturer's name or registered trademark, of appropriate size, shall be stenciled or printed with black waterproof stencil paint or black waterproof ink on the main body portion of each strap in a conspicuous location. Decalcomanias shall not be employed.

3.4 **Workmanship.** Workmanship shall be first class throughout. The finished straps shall be clean, well made and free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 **Supplier responsibilities for inspection.** Such examinations and tests as are set forth in this specification or as shall otherwise be appropriate or necessary to insure that supplies conform to specification requirements, shall be performed by and at the expense of

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T07:58Z
Check the source to verify that this is the current version before use.

the supplier. Suppliers who do not have facilities adequate for such tests shall arrange for the use of test facilities acceptable to the Government. Records of examinations and tests performed by the supplier shall be maintained by the supplier and made available to the Government, upon the Government's request, at any time, or from time to time, for a period of 2 years after delivery of the supplies to which such records relate

4.1.1 Inspection, as used in this specification, is defined as both examination (such as visual, auditory, etc, investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2 *Certificates of quality.* When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.2 Sampling. Sampling shall be conducted in accordance with Standard MIL-STD-105 and Appendix thereto

	Inspection level	AQL (percent defective)
For visual examination		
Major defects	II	1.0
Total (major and minor combined) defects	II	4.0
For dimensional examination of strap and metal loops	L5	2.5

Sampling for materials testing

Component	Characteristic	Req	Test procedure	Rqmt: Indiv unit	Appl: lot avg	Sample unit	Insp * level
Webbing	Mildew inhibitor content	3.1.1	MIL-W-530	X		1 yd.	L2
	Mildew resistance	3.1.1	MIL-W-530	X		1 yd.	L2
Thread	Mildew inhibitor content	3.1.2.1	MIL-T-3530	X		1 cone or spool	L2
	Mildew resistance	3.1.2.1	MIL-T-3530	X		1 cone or spool	L2
Metals (slide loops & strap loops)	Chemical analysis	MIL-L-11075 and MIL-L-3851			X	4 oz. each lot of each metal	
	Finish	MIL-L-11075 and MIL-L-3851		X		each	L3

* Minimum sample size shall be two (2) regardless of lot size. Acceptance number shall be zero (0)

4.2.1 Where possible, the same sample unit shall be used for the determination of two or more test characteristics.

4.3 Examination. The straps shall be examined to determine compliance with all requirements contained in this specification. Nonconformance with these requirements will be permitted to the extent indicated in 4.2.

4.3.1 *Classification of defects.* Examination shall be conducted in accordance with the following classification of defects:

Categories and defects**

Major

101 Strap not adjustable within specified limits.
102. Loops not present.
103. Securing of strap ends not as specified.
104 Repair of stitch defects by backstitching not as specified.
105 Stitches not type 301.

Minor

201. Stitches per inch not within specified limits.
202. Identification marking not legible or otherwise not as specified.
203. Raw ends not stitched as specified.

** Examination shall not be restricted to the classified possible defects listed above.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T07:58Z
Check the source to verify that this is the current version before use.

4.3.2 *Dimensional examination.* The straps shall be examined for defects in dimensions. Any dimension not within the tolerances specified herein shall be classified as a defect.

4.4 **Tests.** Tests shall be conducted, as necessary, to determine compliance with specification requirements.

4.4.1 *Materials.*

4.4.1.1 *Chemical analysis of metals.* Tests for chemical analysis shall be conducted in accordance with method 111 or 112 of Test Method Standard 151.

4.4.1.2 *Webbing* The webbing shall be tested in accordance with the test procedures of Specification MIL–W–530, except that the performance of the tests for mildew inhibitor content and mildew resistance shall be conducted under the following conditions

(a) If the material conforms to the requirements for mildew inhibitor content, the mildew resistance tests shall not be performed.

(b) If the mildew inhibitor content exceeds the maximum requirement, the material shall be rejected

(c) If the mildew inhibitor content is found to be below the minimum requirement, the material may be accepted provided it passes the mildew resistance test

4.4.1.2.1 *Mildew resistance* The finished webbing shall be tested in accordance with method 4762 as specified in the test section of Specification MIL–W–530.

4.4.1.2.2 *Water absorption* The finished webbing shall be tested for water absorption in accordance with method 5500 as specified in the test section of Specification MIL–W–530, except that the size of the specimen and the size of the blotting paper shall conform to method 5502 of Specification CCC–T–191.

4.4.1.3 *Thread.* The untreated thread shall be tested in accordance with the test procedures of Specification V–T–276 (see 3.1.2) The treated thread shall be tested for copper analysis as specified in the test section of Specification MIL–T–3530, except that the performance of the tests for mildew inhibitor content and mildew resistance shall be conducted under the following conditions:

(a) If the material conforms to the requirements for mildew inhibitor content, the mildew resistance test shall not be performed, and the material shall be considered as complying with mildew resistance requirements

(b) If the mildew inhibitor content exceeds the maximum requirement, the material shall be rejected.

(c) If the mildew inhibitor content is found to be below the minimum requirement, the material may be accepted with respect to mildew resistance provided it passes the mildew resistance test The mildew resistance shall be determined by either the soil burial method (20 percent loss in tensile strength, 14 days) or the direct inoculation method The leaching process on the test specimens shall be omitted.

5. PREPARATION FOR DELIVERY

5.1 **Packaging.**

5.1.1 *Unit of issue* One (1) strap, as specified, constitutes one unit

5.1.2 *Packaging quantities* The number of units of issue indicated in table I shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

TABLE I *Packaging quantities*

Federal stock number	Unit package	Intermediate package	Exterior container
8530–784–3120	1 unit	6 units	72 units

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T07:58Z
Check the source to verify that this is the current version before use.

5.1.2.1 *Packing variation permitted.* If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which will insure safe delivery to destination.

5.1.3 *Level A.*

5.1.3.1 *Unit package.* Each unit shall be neatly rolled and then packaged in a carton of appropriate size constructed in accordance with Specification PPP-B-566 or PPP-B-676. Closure shall be adequate to prevent spillage of contents under normal handling.

5.1.3.2 *Intermediate package.* Intermediate package shall be a carton of appropriate size, or a complete wrapping of kraft paper of suitable weight. Closure shall be adequate to prevent spillage of contents under normal handling.

5.2 Packing.

5.2.1 *Level A.*

5.2.1.1 *Exterior container.* Exterior container shall be designed for type 2 load and constructed in accordance with Specification PPP-B-585, style 3, for class 3 use; PPP-B-601, table II, using type III, class 1 plywood; PPP-B-621, table III; or MIL-B-10377, table II.

5.2.1.2 *Case liner.* Each level A box shall be lined with a waterproof case liner conforming to Specification MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto.

5.2.1.3 *Strapping.* Strapping, when required, shall be in accordance with applicable box specification.

NOTE: Strapping will not be required for shipment forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

5.2.2 *Level B.*

5.2.2.1 *Exterior container.* Exterior container shall conform to Specification PPP-B-601, table I, using type III, class 1 plywood, or shall be designed for type 2 load and constructed in accordance with Specification PPP-B-585, style 3, for class 2 use; PPP-B-621, table I; or MIL-B-10377, table I.

5.3 Marking.

5.3.1 *Unit cartons.* Each unit carton shall be marked with stock number, item identification, unit, quantity, and name of contractor.

5.3.2 *Intermediate package.* Each intermediate package shall be marked in accordance with section 5 of Standard MIL-STD-129.

5.3.3 *Exterior container.* Exterior container shall be marked in accordance with Standard MIL-STD-129, except that 6.2(c), 6.2(d), 6.5, and 7.5.1 shall not be applicable. Service color and service symbol markings are required only on direct shipments from the contractor's plant to overseas activities.

6. NOTES

6.1 Intended use. The strap covered by this specification is intended for use with Federal stock number 6530-788-7905, LITTER, FOLDING, RIGID POLE, Aluminum Pole.

6.2 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.

(b) Selection of applicable level of packing (see 6.2).

6.3 This specification covers the following item appearing in the Armed Services Medical Stock List:

6530-784-8120 STRAP, WEBBING, Litter Carrying, Olive Drab.

6.4 This specification does not include all types, classes, grades, sizes, etc., of the com-

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T07:58Z
Check the source to verify that this is the current version before use.

modity indicated by the title of this specification, or which are commercially available, but is intended to cover the types, etc., which are normally procured to meet Military requirements.

Notice. When Government drawings, specifications, or other data are used for any purpose other than in connection with a definitely related Government procurement operation, the United States Government thereby incurs no responsibility nor any obligation whatsoever, and the fact that the Government may have formulated, furnished, or in any way supplied the said drawings, specifications, or other data is not to be regarded by implication or otherwise as in any manner licensing the holder or any other person or corporation, or conveying any rights or permission to manufacture, use, or sell any patented invention that may in any way be related thereto.

Custodians:

Army—Army Medical Service
Navy—Bureau of Medicine and Surgery
Air Force

Preparing activity:

Navy—Bureau of Medicine and Surgery
(MMSA)

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T07:58Z
Check the source to verify that this is the current version before use.