COMMERCIAL ITEM DESCRIPTION

GAGE, ANEROID, SPHYGMOMANOMETER

The General Services Administration has authorized the use of this commercial item description

This commercial item description covers the requirements for a gage suitable for use with an aneriod type sphygmomanometer

Salient characteristics:

Shall be a monometer gage suitable for the determining pressure measurements in the range of 0 to 300 millimeters of mercury when used with an areroid sphygmomanometer

The complete assembly shall essentially consist of an aneroid gage, a glass dial face, with retaining ring and a spring clip and pin assembly

The dial of the gage shall be made of metal or plastic The dial shall have a pressure scale clearly and permanently marked extending from zero to not less than 300 millimeters of mercury The graduations shall cover the entire range of the scale in 2 millimeter divisions, with each 10 millimeter division longer than the intervening marks, and each 20 millimeter division marked with Arabic numbers, indicating the pressure value There shall be a zero indicating mark or marks, within the limits of which the pointer shall rest when there is no pressure upon the diaphragm or bellows

No zero stop pin shall be used on the face The pointer of the gage shall be light and of sufficient fineness at the point to allow accurate reading of the pressure There shall be no internal resistance of the pointer movement within 25 angular degrees of the zero indicating mark The housing of the gage shall have a spring, clip and pin assembly suitable for holding the gage on the cuff when the instrument is in use

No screw holding any of the inner mechanism shall be accessible from the outside of the instrument

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A Approved for public release, distribution is unlimited

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:30Z
Check the source to verify that this is the current version before use.

A tapered male adapter on the bottom of the gage shall adapt to blood pressure apparatus tubing to form a snug fit The gage shall be compatible with a commercially available aneriod sphygmomanometer

Accuracy The error at any point shall not exceed 3 millimeters of Hg The motion of the pointer shall be free from backlash and irregular motion, when the pressure is changed uniformly After a sudden jar, the pointer shall return to the point previously occupied Excessive irregular motion of the pointer shall be considered a sufficient cause for rejection

Workmanship The gage shall be free from defects which detract from its appearance or impair its serviceability

Visual and dimensional examination The gage shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements

Unit Each (EA) One gage, as specified, constitutes one unit

QUALITY ASSURANCE PROVISIONS

Responsibility for inspection Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein Except otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements

Records Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate

Inspection Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means and chemical properties) of the item

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:30Z
Check the source to verify that this is the current version before use.

Tests Tests shall be conducted to determine compliance with specification requirements Where feasible, the same sample shall be used for the determination of two or more test characteristics

Sampling for tests Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein The unit of product for test purposes shall be one gage

Contractor certification The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers'own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract

Metric products Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the lastest revision of Federal Standard No 376, and all other requirements of this document are met

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the products is acceptable

The contracting officer has the option of accepting or rejecting the product

Regulatory requirements

Federal Food, Drug, and Cosmetic Act If the product covered by this document has been determined by the U S Food and Drug Administration to be under its jurisdiction , the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug Cosmetic Act, as amended, and regulations promulgated thereunder In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal State, and local statutes, ordinances, and regulations

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:30Z
Check the source to verify that this is the current version before use.

Recovered materials The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23 4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc , involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and /or order

NOTE The following National Stock Number is covered by this document

NSN	ITEM IDENTIFICATION
6515-01-219-3670	GAGE, ANERIOD, Sphygmomanometer

Ordering data (Intermediate/exterior package quantities and labeling ard marking must be specified ir the contract and/or order)

MILITARY INTERESTS·

Custodians

Army - MD
Navy - MS
Air Force - 03

Preparing Activity

DoD-MB

Agent

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES·

VA-OSS
PHS
FDA-MPQAS

Project No 6515-4841

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:30Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 Alexandria, VA

POSTAGE WILL BE PAID BY

Commander
Defense Personnel Support Center
ATTN: DPSC-RST
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

NO POSTAGE
NECESSARY
IF MAILED
IN THE
UNITED STATES

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:30Z
Check the source to verify that this is the current version before use.

(TO DETACH THIS FORM CUT ALONG THIS LINE.)

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions - Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-54170	GAGE, ANEROID, SPHYGMOMANOMETER

3a NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER *(Last, First, MI — Optional)*	b WORK TELEPHONE NUMBER *(Include Area Code) — Optional*
c MAILING ADDRESS *(Street, City, State, ZIP Code — Optional)*	8 DATE OF SUBMISSION *(YYMMDD)*

DD FORM 1426

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:30Z
Check the source to verify that this is the current version before use.