COMMERCIAL ITEM DESCRIPTION

BAG, COMPRESSION, SPHYGMOMANOMETER

The General Services Administration has authorized the use of this commercial item description

This commercial item description covers the requirements for a nonconductive compression bag suitable for inflating sphygmomanometer cuff

Salient characteristics

Material The bag shall be fabricated of rubber, natural rubber, or plastic suitable for inflating sphygmomanometer cuff The material used need not be conductive The tubing shall be constructed of suitable plastic or rubber Both bag and the tubing shall be free of stiffening, softening or cracking The bag shall be of seamless construction and shall withstand an internal pressure of 15 pounds per square inch The metal parts shall be fabricated form corrosion-resisting steel, aluminum, or nickel-plated brass

Style and design The bag shall nominally measure 23 centimeters (9-1/8 inches) in length and 12 centimeters in width Two lengths of tubing with nominal inside diameter of 5/32 inches (40mm) and lengths of 8 inches (20mm) and 18 inches (46cm) shall be permanently affixed on one long edge of the bag, approximately 2-1 2 inches (6 5 cm) from the short axis of the bag

One tube shall be fitted with a metal female Luer connector The Luer connector shall be in accordance with ANSI\HIMA MD 70 1

The bag shall be capable of repetitive autoclaving

Performance

Autoclaving The bag shall not lose shape or decompose when autoclaved three times at 15 psig and 250 degrees F for 15 minutes and then allowed to cool at room temperature (68 72-F)

AMSC N/A FSC 6515
DISTRIBUTION STATEMENT A Approved for public release, distribution is unlimited

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:30Z
Check the source to verify that this is the current version before use.

Leakage of compression bag. Leakage of the compression bag shall not exceed 10 millimeters of mercury per minute when subjected to a pressure of 300 millimeters of mercury. The bag shall be secured from ballooning by the use of screening of sufficient strength to withstand the pressure. The tests shall be conducted both before and after autoclaving.

Identification marking. The compression bag shall be legibly and permanently marked with the contractor's name, registered trademark, and date of manufacturing (month and year).

Workmanship. The bag shall be free from defects which detract from its appearance or impair its serviceability.

Visual and dimensional examination. The bag shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements.

Material. The supplier shall be responsible for performing tests to assure that materials are in accordance with requirements.

Unit. Each (Ea). One sphygmomanometer compression bag, as specified, constitutes one unit.

QUALITY ASSURANCE PROVISIONS

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:30Z
Check the source to verify that this is the current version before use.

Inspection. Inspection as used herein is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means and chemical properties) of the item

Tests Tests shall be conducted to determine compliance with specification requirements Where feasible, the same sample shall be used for the determination of two or more test characteristics

Sampling for tests Sampling for tests shall be conducted in accordance with MIL STD-105 and as specified herein The unit of product for test purposes shall be one sphygmomanometer compression bag

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the lastest revision of Federal Standard No 376, and all other requirements of this document are met

If a product is manufactured to metric dimensions and the dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the products is acceptable

The contracting officer has the option of accepting or rejecting the product

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:30Z
Check the source to verify that this is the current version before use.

A-A-5419

Regulatory requirements

Federal Food, Drug, and Cosmetic Act If the product covered by this document has been determined by the U S Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug Cosmetic Act, as amended, and regulations promulgated thereunder In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal State, and local statutes, ordinances, and regulations

Recovered materials The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 as amended, and as implemented by Federal Acquisition Regulation Subpart 23 4 to the maximum extent practical

Preservation, packaging, packing, labeling, and marking Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc , involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and /or order

NOTE The following National Stock Number is covered by this document

NSN	ITEM IDENTIFICATION
6515-01-117-0506	BAG, COMPRESSION, Sphygmomanometer

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order)

MILITARY INTERESTS:

Custodians

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent

DLA-DM

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:30Z
Check the source to verify that this is the current version before use.

CIVIL AGENCY COORDINATING ACTIVITIES

VA-OSS
PHS
FDA-MFQAS

Project No 6515-4842

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:30Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS. In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4900 Alexandria, VA

POSTAGE WILL BE PAID BY

Commander
Defense Personnel Support Center
ATTN: DPSC-RSTS
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:30Z
Check the source to verify that this is the current version before use.

(TO DETACH THIS FORM CUT ALONG THIS LINE.)

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-54191	BAG, COMPRESSION, SPHYGMOMANOMETER

3a NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b ADDRESS *(Street City State ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5 PROBLEM AREAS

a Paragraph Number and Wording

b Recommended Wording

c Reason/Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER *(Last First MI)* – Optional	b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c MAILING ADDRESS *(Street City State ZIP Code)* – Optional	8 DATE OF SUBMISSION *(YYMMDD)*

DD FORM 1426

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:30Z
Check the source to verify that this is the current version before use.