A-A-54193
1 July 1990

COMMERCIAL ITEM DESCRIPTION

PAD, POST-SURGICAL-OBSTRETRICAL, BELTLESS, POSTPARTUM OR POSTOPERATIVE

The General Services Administration has authorized the use of this Commercial Item Description.

This commercial item description covers a post-partum, post-operative, beltless pad.

Salient characteristics.

Shall be a clean, beltless, extra absorbent pereneal pad having a minimum length of 13 inches and a minimum width of 3 inches. The pad shall be suitable for post-partum patients and post-operative vaginal/rectal patients.

The filler shall be absorbent cellulose fibers. The cellulose fibers (fluff) shall be bleached white chemical wood pulp which shall be relatively free from impurities and foreign matter. Other extra absorbent fibers and bleached absorbent creped cellulose may be used in combination with the cellulose fibers. The absorbent fibers shall be contained so that they shall not penetrate the pad covering. The entire filler shall be free of any water soluable coloring matter. The filler shall not bunch or separate during use.

The pad shall be completely wrapped and completely sealed with a soft nonwoven polypropylene based fiber fabric with the fiber having a minimum denier of four. The filler shall be completely contained within the pad. The dry breaking strength of the nonwoven fabric cover shall be a minimum of 5 pounds in the machine direction and a minimum of 3 pounds in the cross direction when tested by ASTM D2208-80. The minimum weight of the nonwoven fabric cover shall be 0.70 ounces per square yard. The side of the pad opposite the patient side shall contain a sheet of soft plastic just under the nonwoven fabric cover and the sheet of soft plastic shall be a barrier against strike-through drainage.

AMSC/NA FSC 6510

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:30Z
Check the source to verify that this is the current version before use.

The soft plastic sheet shall run the full length and width of the pad and shall be fastened in place. The side of the pad opposite the patient side shall have a pressure sensitive adhesive strip(s) which shall run almost the entire length of the pad on the outside of the pad. The adhesive strip(s) shall hold the pad securely in place during use and the pads shall be easily detached for disposal. The adhesive area shall be protected by an easily removable paper facing.

The finished pad shall be non-irritating and shall be non-adhering to skin, sutures or to the surgical site. Both ends of the finished pad shall be rounded. The finished pad shall not ravel, tear or bunch during use. The finished pad shall have a minimum weight of 24 grams and the pad shall absorb a minimum of 8 times its weight of water.

Workmanship. The pad shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). Fifteen pads shall be packaged in a bag capable of maintaining cleanliness of the pads. Each pad shall be supplied in a plastic disposal bag, or, as an alternate, 15 plastic disposal bags shall be included with the pads. One package containing 12 bags of 15 pads each (180 pads), as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and test performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:30Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item. Testing and examination shall be conducted to determine compliance with all specification requirements. Test methods shall be suitable, accurate and reproducible.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for examination shall be one pad. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5 maximum percent defective.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tablets contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:30Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6510-01-293-5596	PAD, POST-SURGICAL-OBSTRETRICAL, BELTLESS, POSTPARTUM OR POSTOPERATIVE, 180s

Ordering data. Inspection/test dating required, intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:30Z
Check the source to verify that this is the current version before use.

A-A-54193

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS
PHS
FDA-MPQAS

PGC 71156

Project No. 6510-0769

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:30Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 8, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 8 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC- RSIE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:30Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-54193	PAD, POST-SURGICAL-OBSTRETRICAL, BELTLESS, POSTPARTUM OR POSTOPERAT

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)*:

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:30Z
Check the source to verify that this is the current version before use.