A-A-54195
18 June 1990

COMMERCIAL ITEM DESCRIPTION

STOPPER SET, BOTTLE

The General Services Administration has authorized the use of this Commercial Item Description in preference to Stopper Set requirements of Military Specification MIL-S-36662 which is cancelled.

This Commercial Item Description covers an assortment of perforated bottle stoppers with removable plugs, for use in a chemical or clinical laboratory.

Salient Characteristics:

Shall be an assortment of thirty-three (33) white color rubber stoppers for use as closures on bottles used in laboratory procedures.

Shall be suitable for use as both a solid stopper and a perforated stopper.

Shall all be fabricated from natural rubber or synthetic rubber or from a mixture of natural and synthetic rubber.

Shall be supplied with stoppers having full length perforations (holes) that are countersunk on the top surface, and are sealed on the bottom surface with protruding and removable nodules (plugs). Each perforation (hole) shall have a millimeter (mm) diameter of 5 mm +/- 0.5 mm. The number of perforations (holes) for each required designated size stopper shall be as hereinafter indicated.

Shall include the following number of stopper sizes with pertinent dimensions and number of perforations (holes)as indicated:

DIMENSIONS OF STOPPERS (MM)

Quantity Required	Stopper Number Size	Number of Holes	Diameter of Top	Diameter of Bottom	Length
9	3	2	24	18	25
8	4	2	26	20	25
6	5	3	27	23	25

AMSC N/A FSC 6640

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:30Z
Check the source to verify that this is the currentversion before use.

DIMENSIONS OF STOPPERS (MM)

Quantity Required	Stopper Number Size	Number of Holes	Diameter of Top	Diameter of Bottom	Length
4	6	3	32	26	25
2	7	3	37	30	25
2	8	3	41	33	25
1	9	3	45	37	25
1	10	3	50	42	25

A +/- 1 mm tolerance shall be applicable to the dimensional requirements for the diameters (top and bottom) and lengths of the stoppers.

Shall have each stopper supplied with its stopper number size identified in the form of a raised number on the top side.

Shall have the bottom side of each perforation (in a stopper) supplied with an inserted protruding nodule (plug) that can be removed by twisting off with finger pressure. The removal of each nodule (plug) shall result in an unobstructed full length opening that permits the full insertion of laboratory glass tubing.

Shall be supplied in a unit package (see "Unit" requirements) that includes a label printed with the legend "CAUTION-Stoppers are rubber based and may become tacky when used with organic solvents".

Workmanship. The bottle stoppers shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Package (PG). One package containing a thirty-three stopper set of the stopper sizes specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:30Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspections and testing. Inspections and testing shall be conducted on the bottle stoppers in accordance with commercial practice. Sampling for inspections and testing shall be in accordance with commercial practice.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:30Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. (Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.)

NOTE: The following National Stock Number is covered by this document.

NSN	Item Identification
6640-00-935-1122	STOPPER, SET, BOTTLE, White Rubber, Assorted Sizes, Set of 33

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project Number: 6640-1638

Location: ENABLE/STOPSETC/S31

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:30Z
Check the source to verify that this is the current version before use.