A-A-54196
25 June 1990

COMMERCIAL ITEM DESCRIPTION

PIPET, SEROLOGICAL (Plastic, Disposable)

The General Services Administration has authorized the use of this commercial item description in preference to the type II pipets of Military Specification MIL-P-0036599 (DSA-DM) which is cancelled.

This commercial item description covers disposable sterile serological pipets with plugged ends, for use in laboratory procedures.

Salient characteristics.

Shall be disposable plastic serological pipets that are calibrated "to deliver" (TD) rated capacity by blowing out the last drop.

Shall be plugged at the proximal end with natural or synthetic fibrous material.

Shall be supplied in a sterile condition.

Shall have the following rated capacities and graduation markings (milliliters):

Specification designated size	Rated Capacity	Graduation Increments	Numbered Increments	Numbered Range
Size 1	1.0	0.1	0.1	0 to 0.9
Size 2	1.0	0.01	0.1	0 to 0.9
Size 3	5.0	0.1	1.0	0 to 4.0
Size 4	10.0	0.1	1.0	0 to 9.0

Shall be designed to deliver the rated capacity with a volumetric accuracy of ± 3%.

Shall conform to the requirements of ASTM Designation E934.

Shall be tested for volumetric accuracy to meet the requirements of the College of American Pathologists (see Quality Assurance Provisions).

AMSC N/A FSC 6640

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:30Z
Check the source to verify that this is the current version before use.

Workmanship. The pipets shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Package (PG). For size 1 and size 2, one package containing one hundred pipet of one size only, as specified, constitutes one unit.

For size 3 and size 4, one package containing fifty pipet of one size only, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspections and tests. Sampling for inspections and tests shall be conducted in accordance with commercial practice.

Inspections and tests. Inspections and tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for one or more tests.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:30Z
Check the source to verify that this is the current version before use.

Volumetric accuracy testing. To meet requirements for certification of the College of American Pathologists, the pipets shall be tested in accordance with ASTM Designation E 934 to determine conformance to calibration and capacity tolerance requirements. A Certificate of Quality may be furnished in lieu of actual performance of individual testing, provided that lot/sample identities can be maintained. The Certificate of Quality shall include the lot number, lot size, sample size and test method.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/ contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:30Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Other publications. The following documents form a part of this specification to the exent that is specified herein. Unless otherwise indicated, the effect on date of invitation for bids or request for proposals shall apply.

AMERICAN SOCIETY FOR TESTING AND MATERIALS (ASTM)

ASTM E 934 Standard Specification for Serological Pipet, Disposable Plastics

(Application for copies should be addressed to American Society for Testing and Materials, 1916 Race Street, Philadelphia, PA 19103)

NOTE: The following National Stock Numbers are covered by this document:

NSN	CAPACITY	SPECIFICATION DESIGNATED SIZE	ITEM IDENTIFICATION
6640-00-241-6520	1.0 ml	1	PIPET, SEROLOGICAL, Plastic, Disposable, 1.0 ml, 100's
6640-00-533-8634	10.0 ml	4	PIPET, SEROLOGICAL, Plastic, Disposable, 10.0 ml, 50's
6640-00-728-7267	5.0 ml	3	PIPET, SEROLOGICAL, Plastic, Disposable, 5.0 ml, 50's
6640-00-829-4023	1.0 ml	2	PIPET, SEROLOGICAL, Plastic, Disposable, 1.0 ml, 100's

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:30Z
Check the source to verify that this is the current version before use.

Ordering data. (Designated size required, intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6640-1615

Location: CID/PIPET/SEROLOGICAL/S31/12

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:30Z
Check the source to verify that this is the current version before use.