A-A-54208
18 July 1990

COMMERCIAL ITEM DESCRIPTION

SYRINGE, IRRIGATING (Metal, Ear, 2 oz)

The General Services Administration has authorized the use of this commercial item description as a replacement for Federal Specification GG-S-00928C(DSA-DM) which is cancelled

This commercial item description covers a metal irrigating syringe suitable for irrigating the human ear

Salient characteristics

General The ear syringe shall consist of the following components: Barrel, cap, plunger rod assembly, guard assembly, and a bulbous ear tip having a working capacity of 2 fluid ounces The cap shall be readily removable and readily assembled to the proximal end of the barrel The plunger assembly shall accurately fit into the barrel providing for free plunger travel throughout the full excursion of the barrel External threads of the syringe tip and guard assembly shall match the barrel and guard assembly

Plunger assembly. The Type I plunger proper shall be of heavy wall and cup shaped construction The piston component shall be sized to conform to the inside diameter of the barrel The distal edge of the piston shall be slightly beveled to permit ready insertion in the barrel The plunger rod proper shall be securely attached to the thumb rest at its proximal end

The Type II plunger proper shall be of the "O" ring design Two extra "O"- rings shall be furnished with each syringe that contains plunger assemblies of the Type II design. "O" rings shall be coated with a light film of an appropriate lubricant, and shall be furnished with instructions as specified. The piston component shall be sized to conform to the inside diameter of the barrel The plunger rod proper shall be securely attached or threaded to the piston component at its distal end, and securely attached or threaded to the thumb rest at its proximal end The piston shall be recessed to properly accommodate an "O" ring of adequate size

AMSC N/A FSC 6515
DISTRIBUTION STATEMENT A Approved for public release. distribution is unlimited

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:30Z
Check the source to verify that this is the current version before use.

Material

Metal parts Metal parts shall be fabricated from chromium plated over nickel plated brass As an alternate, the inside surfaces of the barrel and the plunger proper shall be nickel plated only As a further alternate, metal parts may be fabricated from Class 3 austenitic stainless steel in accordance with ASTM Designation F899

Sealing gaskets and 'O' Rings Gaskets that are placed on the syringe tip, the proximal end of the guard assembly, and the 'O' ring contained on the plunger assembly to Type II design shall be fabricated from natural or synthetic rubber.

Style, design and dimensions The syringe shall be in accordance with Defense Personnel Support Center (DPSC) Drawing No 24244

Hardness Metal parts fabricated from Class 3 steel scale (HRB) of 70 to 90 when tested in accordance with ASTM Designation E 18 or E 92 whichever is applicable Conversion of hardness values shall be in accordance with ASTM Designation E 140

Finish Metal parts fabricated from Class 3 steel shall be mirror satin, or matte

Corrosion resistance. The plated brass components shall show no signs of corrosion when salt spray tested for 24 hours in accordance with ASTM Designation B117 Class 3 Stainless steel components shall show no signs of corrosion when boil tested in accordance with ASTM Designation F1089

Performance

Fluid tightness

Barrel and plunger. The barrel and plunger shall be tested for fluid tightness as specified When pressure is applied as specified there shall not be leakage exceeding 3 cubic centimeters of water.

Tips and guard. The tips shall be tested for fluid tightness as specified When pressure is applied as specified, there shall not be water leakage at the threaded end of the tips exceeding one falling drop

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:30Z
Check the source to verify that this is the current version before use.

Identification marking and labeling Syringes shall be permanently marked with the manufacturer's name or registered trademark and the word 'stainless' if applicable In addition, syringes of foreign manufacture shall bear the name of the country of origin.

Instruction Sheet The following shall be furnished with each syringe containing Type II plunger assemblies·

'INSTRUCTIONS'

Apply Dow Corning No 200 fluid, with not less than 100 centistrokes viscosity grade at 25· C (or other suitable nontoxic lubricant compatible with solutions to be used) to the 'O' - ring as required to allow piston to move freely throughout excursions of barrel after sterilization '

Workmanship The syringe shall be free from cracks, pits, crevices, burrs, tool marks, and dirt or polishing materials Internal flaking of plating shall be cause for rejection

Visual and dimensional examination Syringes shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements

Materials. The supplier shall be responsible for performing tests to assure that materials are in accordance with requirements

Unit Each (EA) One syringe, as specified, constitutes one unit

QUALITY ASSURANCE PROVISIONS

Responsibility for inspection Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein Except otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:30Z
Check the source to verify that this is the current version before use.

Records Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means and chemical properties) of the item

Sampling for examination Sampling for examination shall be conducted in accordance with MIL-STD-105, and as specified herein Unit of product for examination purposes shall be one syringe Sampling shall be inspection level II with an AQL of 1 0

Examination Scissors shall be examined to determine compliance with all specification requirements

Sampling for tests Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein Unit of product for test purposes shall be one syringe Sampling shall be Inspection Level S-2 with an AQL Of 1.0.

Tests Tests shall be conducted to determine compliance with specification requirements Where feasible, the same sample shall be used for the determination of two or more test characteristics

Fluid tightness test Completely assembled clean syringes shall be hydrostatically tested at 10 psig in horizontal position Water shall be applied through the syringe tip. With the plunger set at one-ounce and two ounces, check for water leakage past the plunger and for leakage at the threaded connections of the tip and guard After one minute at the one or two ounce settings, there shall be no more than 3cc of water leakage past the plunger and less than one clinging drop past the threaded tip connections

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:30Z
Check the source to verify that this is the current version before use.

Contractor certification The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers'own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the lastest revision of Federal Standard No 376, and all other requirements of this document are met

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the products is acceptable

The contracting officer has the option of accepting or rejecting the product

Regulatory requirements

Federal Food, Drug, and Cosmetic Act If the product covered by this document has been determined by the U S Food and Drug Administration to be under its jurisdiction , the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug Cosmetic Act, as amended, and regulations promulgated thereunder In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal State, and local statutes, ordinances, and regulations.

Recovered materials The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23 4 to the maximum extent practical

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:30Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc , involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and /or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-00-379-5850	SYRINGE, IRRIGATING, METAL, EAR, 2 oz

Ordering data (Intermediate lexterior package quantities and labeling and making must be specified in the contract and/or order)

MILITARY INTERESTS

Custodians

Army - MD
Navy - MS
Air Force - 03

CIVIL AGENCY COORDINATING ACTIVITIES

VA-OSS
PHS
FDA-MPQAS

PREPARING ACTIVITY

DoD-MB

Agent:

DLA-DM

PROJECT NO 6515-4846

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:30Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 Alexandria, VA

POSTAGE WILL BE PAID BY

Commander
Defense Personnel Support Center
ATTN: DPSC-RSTS
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:30Z
Check the source to verify that this is the current version before use.

TO DETACH THIS FORM, CUT ALONG THIS LINE.)

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions — Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-54208	SYRINGE, IRRIGATING (Metal, Far, 2 oz)

3a NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER (Last, First, MI — Optional)	b WORK TELEPHONE NUMBER (Include Area Code) — Optional
(Mailing Address, City, State, ZIP Code — Optional)	8 DATE OF SUBMISSION (YYMMDD)

DD FORM 1426 (PREVIOUS EDITION MAY BE USED)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:30Z
Check the source to verify that this is the current version before use.