
FED. SUP CLASS
6640



PART NUMBER	FOR FLASK SIZE MILLILITERS	WATTAGE RANGE
MS36223 -1	50	50 to 65
MS36223 -2	100	75 to 85
MS36223 -3	125	75 to 85
MS36223 -4	200	95 to 105
MS36223 -5	250	135 to 190
MS36223 -6	300	135 to 190
MS36223 -7	500	200 to 280
MS36223 -8	1000	320 to 410
MS36223 -9	2000	480 to 540
MS36223 -10	3000	480 to 575
MS36223 -11	5000	575 to 650

MATERIAL: Exterior cylindrical rigid housing shall consist of an aluminum case or a glass fiber reinforced plastic case, that is resistant to chemicals, high temperatures and fire. Internal construction materials, from the case inwards, shall consist of a coated glass cloth cover, an insulation layer of glass fabric or fiber glass, heating tapes containing resistance wire, and a liner of glass cloth.

DESIGN: Heating mantle interiors shall be of hemispherical shape. Exterior case bottoms may be recessed to provide a stable base, or fitted with feet to provide an air space beneath. Sizes up to 3000 ml shall be fitted with a thermocouple. The 5000 ml size may be fitted with a thermocouple. The glass fabric temperature limit shall be not less than 450°C (842°F).

APPROVED 22 JUN 1979 REVISED

This standard has been approved by the Department of the and is mandatory for use by that activity. All other military activities are required to employ this standard where suitable.

P.A.	EA	TITLE	MILITARY STANDARD
Other Cust	MS,99 MB	MANTLE, HEATING, ELECTRICAL, TABLE MODEL, WITHOUT BUILT-IN TEMPERATURE CONTROL	MS 36223
Procurement Specification None		SUPERSEDES:	SHEET 1 OF 2

DD FORM 1 SEP 63 672 (Coordinated) EDITION OF 1 SEP 57 MAY BE USED. PROJECT NUMBER 6640-1073

Source: https://assist.dla.mil -- Downloaded: 202640-23T17:36Z
Check the source to verify that this is the current version before use.

FED. SUP CLASS
6640

All electrical components and insulation shall meet the standards of Underwriters Laboratories, Inc., and shall operate on 110 to 120 volts, 60 Hz, AC.

NOTES: Items differing from this Military Standard by minor variations in design are considered standard under this document, provided such variations do not affect performance and are not in conflict with material requirements as specified herein.

This standard has been approved by the Department of the and is mandatory for use by that activity. All other military activities are required to employ this standard where suitable.

P.A.	TITLE	MILITARY STANDARD
Other Cust	MANTLE, HEATING, ELECTRICAL, TABLE MODEL, WITHOUT BUILT-IN TEMPERATURE CONTROL	MS 36223
Procurement Specification	SUPERSEDES:	SHEET 2 OF 2

DD FORM 1 SEP 63 672 (Coordinated) EDITION OF 1 SEP 57 MAY BE USED.

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:36Z
Check the source to verify that this is the current version before use.