FEDERAL SPECIFICATION

BASSINET AND DRESSING TABLE COMBINATION, HOSPITAL, CORROSION-RESISTING STEEL

This specification was approved by the Commissioner, Federal Supply Service, General Services Administration, for the use of all Federal agencies.

1. SCOPE

1.1 This specification covers a combination bassinet and dressing table, suitable for use in hospital pediatric departments for the individual care of newborn babies.

2. APPLICABLE SPECIFICATIONS, STANDARDS, AND DRAWING

2.1 The following specifications, standards, and drawing, of the issues in effect on date of invitation for bids, form a part of this specification:

Federal Specifications:

FF–C–77—Casters; Institutional Duty.
QQ–S–763—Steel Bars, Shapes, and Forgings—Corrosion Resisting.
QQ–S–766 — Steel, Corrosion-Resisting, Plates, Sheets, Strips, and Structural Shapes.
UU–P–271 — Paper, Wrapping, Waterproofed Kraft.
PPP–B–601—Boxes, Wood, Cleated-Plywood.
PPP–B–621—Boxes, Wood, Nailed and Lock-Corner.

Federal Standards:

Fed. Std. 102—Preservation, Packaging, and Packing Levels.
Fed. Std. 123—Marking for Domestic Shipment (Civilian Agencies).
Fed. Test Method Std. No. 151—Metals; Test Methods.

(Activities outside the Federal Government may obtain copies of Federal Specifications, Standards, and Handbooks as outlined under General Information in the Index of Federal Specifications, Standards, and Handbooks and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U. S. Government Printing Office, Washington 25, D. C.

(Single copies of this specification and other product specifications required by activities outside the Federal Government for bidding purposes are available without charge at the General Services Administration Regional Offices in Boston, New York, Atlanta, Chicago, Kansas City, Mo., Dallas, Denver, San Francisco, Los Angeles, Seattle, and Washington, D. C.

(Federal Government activities may obtain copies of Federal Specifications, Standards, and Handbooks and the Index of Federal Specifications, Standards, and Handbooks from established distribution points in their agencies.)

Military Standards:

MIL–STD–105 — Sampling Procedures and Tables for Inspection by Attributes.
MIL–STD–129 — Marking for Shipment and Storage.
MS16686—Medical Services Symbol.

Military Medical Supply Agency Drawing:

1914 Bassinet and Dressing Table Combination, Hospital, Corrosion-Resisting Steel.

(Copies of Military standards and drawings required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

481466/59/229

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:38Z
Check the source to verify that this is the current version before use.

3. REQUIREMENTS

3.1 Material.

3.1.1 *Sheet, strip, and tubing.* — Sheet, strip, and tubing used in the construction shall be corrosion-resisting steel conforming to class 1 or 2 of Federal Specification QQ–S–766, and shall conform to the chemical requirements specified therein, when tested as specified in 4.4.1.

3.1.2 *Rods and bars.*—Rods and bars used in the construction shall be corrosion-resisting steel conforming to class 302 or 304, condition A, of Federal Specification QQ–S–763, and shall conform to the chemical requirements specified therein, when tested as specified in 4.4.1.

3.2 Construction.—The bassinet shall consist of a tubular carriage frame, removable basket, and a cabinet-type dressing table. The bassinet shall conform to Military Medical Supply Agency Drawing 1914 and this specification.

3.2.1 *Carriage frame.*—The carriage frame shall be as shown on sheet 2 of the drawing. Cross braces, channels, and pull handle shall be continuously welded. The ends of the "U" shaped tubes forming the legs of the carriage frame shall be fitted with casters.

3.2.2 *Basket.*—Basket shall be formed of round-edge strip and shall be as shown on sheet 2 of drawing. Two angles shall be welded to the cross members, as shown, for the purpose of tilting the basket in the head-up or head-down position, when desired. The lengthwise and crosswise members shall form an interlaced basket weave. A corrosion-resisting steel cardholder to accommodate 3- by 5-inch identification cards shall be welded to one end of the basket.

3.2.3 *Cabinet.* — Cabinet shall be box shaped, of all welded construction, and shall have a lower storage compartment fitted with a receding-type door on the left side, and an upper drawer which shall pull out from the front of the cabinet. The cabinet sides and bottom shall be of 0.0375-inch (No. 20, U.S. Standard gage) sheet, and the top shall be of 0.109-inch (No. 12, U.S. Standard gage) sheet. A top shield of 0.078-inch (No. 14, U.S. Standard gage) sheet shall be welded to the body on three sides. The shield shall be attached to the top and soldered in such a manner as to eliminate all crevices on the top working surface. A paper-bag holder of 0.037-inch (No. 20, U.S. Standard gage) sheet shall be securely welded to the body near the bottom of the front end. The paper-bag holder shall be open on the top and bottom, except for a single strap across the bottom to support the paper-bag. Front end of cabinet shall be mounted on two legs of 1-inch-outside-diameter tubing, 0.065-inch (No. 16, Birmingham Wire gage) wall, fitted with casters. A pull handle shall be attached to the front shield. The pull may be of brass, bronze, steel, corrosion-resisting steel, or anodized aluminum. Where brass or bronze pulls are furnished, they shall be plated with chromium over nickel. Where steel pulls are furnished, they shall be plated with chromium over nickel over copper. The cabinet shall slide in and out of the carriage frame on 1⅛-inch-diameter rubber-tired ball-bearing rollers (see 6.2), riding in the longitudinal channels attached to the carriage frame. The channels shall be fitted with positive stops to prevent complete removal of the cabinet from the carriage frame. In addition, there shall be a similar channel welded to the bottom of the cabinet which shall ride over another 1⅛-inch-diameter rubber-tired ball-bearing roller (see 6.2), attached to the front lower cross brace of the carriage frame for the purpose of guiding the cabinet into the carriage frame. The cabinet shall have lift-up automatic locking rods, front and rear. The locking rods shall ride free of the carriage frame for silent movement of the cabinet with respect to the carriage frame. The locking rods shall automatically engage into the proper hole in the bar strike (see section E–E, sheet 4 of drawing) to act as a positive lock, when the cabinet is completely retracted

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:38Z
Check the source to verify that this is the current version before use.