AA-S-130B
14 February 1991
SUPERSEDING
AA-S-00130A(DSA-DM)
2 March 1966
SUPERSEDING
AA-S-130
25 October 1955

FEDERAL SPECIFICATION

# SCREENS, BEDSIDE, ALUMINUM

This specification is approved by the Commissioner, Federal Supply Service, General Services Administration, for the use of all Federal Agencies.

## 1. SCOPE AND CLASSIFICATION

1.1 Scope. This specification covers 2 styles of aluminum, folding, bedside screens furnished without curtains.

1.2 Classification. Screens shall be of the following styles as specified (see 6.1).

Style 1 - with folding legs

Style 2 - with rigid legs

## 2. APPLICABLE DOCUMENTS

2.1 The following documents, of the issues in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein.

SPECIFICATIONS

Federal

| | |
|---|---|
| FF-C-88 | Casters, Rigid and Swivel, Industrial Duty |
| PPP-B-585 | Boxes, Wood, Wirebound |
| PPP-B-601 | Boxes, Wood, Cleated-Plywood |
| PPP-B-621 | Boxes, Wood, Nailed and Lock-Corner |
| PPP-B-636 | Boxes, Shipping, Fiberboard |
| PPP-B-640 | Box, Fiberboard, Corrugated, Triple-Wall |

FSC 6530

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:39Z
Check the source to verify that this is the current version before use.

(The Code of Federal Regulations (CFR) and the Federal Register are for sale on a subscription basis by the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20402. When indicated, reprints of certain regulations may be obtained from the Federal agency responsible for issuance thereof.)

Military

| | |
|---|---|
| MIL-L-10547 | Liners, Case, and Sheet, Overwrap, Water-Vaporproof, or Waterproof:, Flexible |

STANDARDS

Federal

| | |
|---|---|
| FED-STD-123 | Marking for Shipment (Civil Agencies) |

Military

| | |
|---|---|
| MIL-STD-105 | Sampling Procedures and Tables for Inspection by Attributes |
| MIL-STD-129 | Marking for Shipment and Storage |
| MIL-STD-147 | Palletized Unit Loads |
| MIL-STD-794 | Parts and Equipment, Procedures for Packing of |

(Copies of Military Specifications and Standards required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer).

DRAWINGS

Defense Personnel Support Center (DPSC)

| | |
|---|---|
| No. 24024 | Screen, Bedside, Folding |

(Copies of this drawing may be obtained from the Defense Personnel Support Center DPSC-RST, 2800 So. 20th Street, Philadelphia, PA 19101.)



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:39Z
Check the source to verify that this is the current version before use.

(Activities Outside the Federal Government may obtain copies of Federal Specifications, Standards, and Handbooks as outlined under General Information in the Index of Federal Specifications and Standards and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20402).

(Single copies of this specification and other Federal Specifications required by activities outside the Federal Government for bidding purposes are available without charge from Business Service Centers at the General Services Administration Regional Offices in Boston, MA; New York, NY; Washington, DC; Atlanta, GA; Chicago, IL; Kansas City, MO; Forth Worth, TX; Denver, CO; San Francisco, CA; Los Angeles, CA; and Seattle, WA.)

(Federal Government activities may obtain copies of Federal standardization documents and the Index of Federal Specifications, Standards and Commercial Item Descriptions from established distribution points in their agencies.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless a specific issue is identified, the issue in effect on date of invitation for bids or request for bids or request for proposal shall apply.

**ASTM**

| | |
|---|---|
| ASTM B221 | Standard Specification for Aluminum-Alloy Extruded Bars, Rods Shapes and Tubes |
| ASTM B580 | Standard Specification for Anodic Oxide Coatings on Aluminum |
| ASTM E34 | Standard Test Methods for Chemical Analysis of Aluminum and Aluminum Base Alloys |

(Application for copies should be addressed to ASTM, 1916 Race Street, Philadelphia, PA 19103).

(Nongovernment standards and other publications are normally available from the organizations which prepare or which distribute the documents. These documents also may be available in or through libraries or other informational services.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:39Z
Check the source to verify that this is the current version before use.

## 3. REQUIREMENTS

3.1 Material. The material shall be suitable in every respect for the purpose intended and shall be as herein specified.

3.2 Construction. Construction shall conform to requirements specified herein for style 1 or 2, and as shown on Drawing No. 24024.

For style 1, assembly view, the 4-1/2 inch dimension (2 places) shall, be from center of rod to top or bottom of frame as applicable. The unit shall be a three-panel, folding bedside screen, 70 inches high and 95 inches wide. The unit shall be constructed of 1-1/4 inches square extruded aluminum tubing in accordance with ASTM B221, requirements 6063-T4. The tubing shall have a wall thickness of 0.065 inch (No. 16, Birmingham Wire gage) and a corner radius of 1/8 inch. Aluminum alloy shall meet the requirements for a chemical composition, when tested as specified in 4.4.1.

3.2.1 Center panel (Style 1). The center panel shall be made of one piece of tubing with a weld at the bottom or two pieces of tubing with a weld at the top and bottom. The center panel shall be mounted between a base consisting of formed tubular legs. The legs shall be hinged and bolted perpendicular to the panel and fitted with fixed glides. The hole for the central bolt shall be drilled after assembly of the legs to the center panel. This procedure shall insure that the bolt can readily be inserted into the tubular sections of the screen.

3.2.1.1 Continuous hinges (Style 1). Continuous hinges shall be chromium-plated steel or brass, corrosion-resisting steel, or anodized aluminum. Hinges shall have an open width of 1-1/16 inches and a metal thickness of 0.036 inch minimum. Hinges shall be held in place by sheet metal screws as shown on view EE of Drawing No. 24024 or shall be welded in place in a manner suitable for the purpose intended.

3.2.2 End panels. The end panels shall be attached to the center panel by means of double-action hinges. The end panels shall be fitted with self-adjusting glides.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:39Z
Check the source to verify that this is the current version before use.

3.2.2.1 Double-action hinges. Double-action hinges shall be brass plated steel with a clear lacquer coating, chromium plated steel or brass, or corrosion-resisting steel. Hinges for Style 1 shall have an open width of 2-3/8 inches, a length of 2 inches and a metal thickness of 0.036 inch minimum. The hinges shall be held in place by oval head or flathead countersunk screws. The screws for Style 1 shall fit into headless aluminum inserts knurled on one end (to hold inserts securely) and internally threaded, on the opposite end as shown on detail B of Drawing No. 24024.

3.2.3 Glides. Unless otherwise specified, all legs shall be fitted with conventional 1-1/4 inch diameter nickel or chromium plated glides.

3.2.3.1 Center panel glides. The four center panel glides shall be of removable type. The glides for Style 1 shall be held in place by means of expandable rubber adapters.

3.2.3.2 End panel glides. The two end panel glides shall be spring actuated and self adjusting. The tension on the spring shall be such that the glide is touching the floor at all times. For Style 1 the glides shall be held in place by means of set screws and the Style, location and method of fastening shall be as shown on detail C of Drawing No. 24024.

3.2.3.3 Casters. Casters, when specified in lieu of glides, shall be of the ball bearing, swivel type. The casters shall have 2 inch diameter wheels of hard rubber composition with soft rubber tread. The casters shall be in accordance with FF-C-88.

3.2.4 Curtain rods. The six curtain rods shall be in accordance with ASTM B221, requirements 6061-T6, anodized aluminum. The rods shall be removable and shall be threaded and fitted with knurled nuts for fastening, as shown on detail B of Drawing No. 24024 for Style 1 and as shown for Style 2.

3.2.4.1 Curtains. Curtains shall not be supplied.

3.2.5 Tolerances. All tolerances shall be kept to a minimum. Commercial tolerances on gage sizes are acceptable. Unless otherwise specified on the drawing, tolerance on fractions shall be plus or minus 1/32 inch; tolerance on 3-place decimals shall be plus or minus 0.005 inch.

3.2.6 Stability. The screen shall be so constructed as to permit ready movement and adjustment in its use. The screen shall be sturdy and sufficiently stable to prevent toppling due to an accidental bump or blow in normal use.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:39Z
Check the source to verify that this is the current version before use.

3.3 Finish. Aluminum parts and components, excepting fasteners shall be anodized in accordance with Type D, of ASTM B 580. Unless otherwise specified, hardware, except corrosion-resisting steel and brass, shall be nickel, or zinc plated.

3.4 Identification marking. Each screen shall be permanently and legibly marked with the following:

Name or Registered Trademark of the Manufacturer.
Contract or Purchase Order Number.
National Stock Number.
Manufacturer's Model Number.

3.5 Workmanship. The screen shall be free from defects which detract from its appearance or may impair its serviceability. All surfaces shall be free from sharp edges. All tubing shall be properly bent or formed to the correct radii without ripples, dents, scratches or tears.

3.5.1 Welding. The screen shall be welded as shown on Detail A of Drawing No. 24024. Welds shall be sound, well fused and free from blisters or blowholes. Welds shall develop a strength equal to or greater than the part welded. The rod shall be 5 percent silicon aluminum welding rod. All welds shall be clean and free from twists or strains. All exposed welds shall be ground smooth to eliminate pitmarks and seams and shall be polished to blend with the color of the adjacent surfaces. All welding of side panels and center panel shall be continuous.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for Inspection. Unless otherwise specified in the contract, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract, the contractor may use his own or any other facilities suitable for the perforamnce of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure that supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory, investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.



Source: https://assist.dla.mil -- Downloaded: 2026l0-26T11:39Z
Check the source to verify that this is the currentversion before use.

4.1.2 Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and table I.

TABLE I. Sampling for examination.

| | Inspection level | AQL (defects per hundred units) |
|---|---|---|
| For visual examination | | |
| Major defects | II | 1.0 |
| Total defects (Major and minor combined) | II | 4.0 |
| For dimensional examination | S-2 | 2.5 |

4.3 Examination. The screen shall be examined to determine compliance with all requirements contained in this specification.

4.3.1 Classification of defects. Defects apply to the style being furnished. The examination shall be conducted in accordance with the following classification of defects in table II.

TABLE II. Categories and Defects *

| Categories | Defects |
|---|---|
| Major | |
| 101 | Any component part missing, broken, or unusable. |
| 102 | Any part of tubing not free of tears, ripples, or dents. |
| 103 | Any component not free of cracks, rupture or tears. |
| 104 | Any exposed part not free of burrs and sharpness. |
| 105 | Finish not free of deep orange peel, seed, and foreign matter. |



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:39Z
Check the source to verify that this is the current version before use.

TABLE II. Categories and Defects* - Continued

| Categories | Defects |
|---|---|
| Minor | |
| 201 | Identification marking missing, incomplete, illegible or not properly attached. |
| 202 | Finish not free of scratches or not uniformly applied. |
| 203 | Finish not sufficiently glossy to give the desired eye appeal. |
| 204 | Unit not clean, not free of dirt, oil, or grease. |

*Examination shall not be limited to the classified possible defects listed herein.

4.4 Tests. Tests shall be conducted as necessary, to determine compliance with specification requirements.

4.4.1 Chemical composition. Tests for chemical analysis shall be conducted in accordance with ASTM E34.

4.5 Inspection of packaging. The sampling and inspection of the packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

5. PACKAGING

5.1 Preservation. Preservation shall be level A or C, as specified (see 6.1).

5.1.1 Level A.

5.1.1.1 Unit package. Each screen shall be folded in the closed position. The legs of the style 1 screen shall be placed in the closed position. The screen shall be wrapped in kraft paper or other suitable material in such a manner to prevent scratching or marring of the unit. Closure shall be as specified in the applicable box specification.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:39Z
Check the source to verify that this is the current version before use.

5.1.2 Level C.

5.1.2.1 Unit package. Preservation for the screen shall conform to the requirements of MIL-STD-794 for level C.

5.2 Packing. Packing shall be level A, B or C, as specified (see 6.1).

5.2.1 Level A. Each screen, prepared as specified in 5.1.2, shall be packed in an exterior container designed for a type 3 load and constructed in accordance with PPP-B-636, class weather-resistant or PPP-B-640, class 2. Closure, strapping and waterproofing shall be as specified in the applicable box specification.

5.2.1.1 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall be as specified in the liner specification. Case liner shall not be required for fiberboard boxes. Each fiberboard box shall be waterproofed as specified in PPP-B-636.

5.2.2 Level B. Each screen, prepared as specified in 5.1.2, shall be packed in an exterior container designed for a type 3 load and constructed in accordance with PPP-B-636, class domestic or PPP-B-640, class 1. Closure shall be as specified in the applicable box specification.

5.2.3 Level C. Each screen shall be packed in accordance with the requirements of MIL-STD-794 for level C.

5.2.4 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which will insure safe delivery to destination.

5.2.5 Unitized loads. Unitized loads, commensurate with the level of packing specified in the contract or order, shall be used whenever total quantities for shipment to one destination exceed 250 pounds or 20 cubic feet. Loads shall be unitized on pallets conforming to MIL-STD-147. Quantities less than 20 cubic feet need not be unitized. Pallet loads including the pallet shall not exceed 43 inches in length, 52 inches in width, and 54 inches in height. Pallet shall conform to type IV or type V.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:39Z
Check the source to verify that this is the current version before use.

5.2.5.1 Level A and B. The screen packed as specified in 5.2.1 and 5.2.2 shall be unitized on pallets as specified in 5.2.5. MIL-STD-147.

5.2.5.2 Level C. The screen packed as specified in 5.2.3, shall be unitized as specified in MIL-STD-794.

5.3 Marking.

5.3.1 Civil agencies. Marking shall be as specified in FED-STD-123.

5.3.2 Military activities.

5.3.2.1 Levels A, B and C. Each unit pack, intermediate package, exterior container and unitized load shall be marked as specified in MIL-STD-129.

5.4 General.

5.4.1 Exterior container. Exterior container (see 5.2.1, 5.2.2, and 5.2.3) shall be of minimum tare and cube consistent with the protection required and shall contain equal quantities of identical stock numbered items to the greatest extent practicable.

5.4.2 Packaging inspection. the inspection of these packaging requirements shall be in accordance with 4.5.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

a. Title, number, and date of this specification.
b. Style required (see 1.2).
c. Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.2 This specification is intended to cover the following item appearing in the Federal Supply Catalog:

| National Stock No. (NSN) | Item Identification |
|---|---|
| 6530-00-707-7975 | Screen, Bedside, Aluminum |



Source: https://assist.dla.mil -- Downloaded: 202610-26T11:39Z
Check the source to verify that this is the currentversion before use.

6.3 This specification does not cover all types, classes, grades, or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the types which are normally procured to meet requirements.

MILITARY CUSTODIANS:

Army - MD
Navy - MS
Air Force - 03

CIVIL AGENCY COORDINATING ACTIVITIES:

GSA-FSS
VA-DMS
COMMERCE-NBS

PREPARING ACTIVITY:
DOD-MB

AGENT:
DLA-DM

Project No. 6530-1771

Location: FEDSP/SCREEN/H13-14



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:39Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:39Z
Check the source to verify that this is the current version before use.

# STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

| I. RECOMMEND A CHANGE: | 1. DOCUMENT NUMBER<br>AA-S-130B | 2. DOCUMENT DATE (YYMMDD)<br>14 February 1991 |
|---|---|---|

3. DOCUMENT TITLE
SCREENS, BEDSIDE, ALUMINUM

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

| a. NAME (Last, First, Middle Initial) | b. ORGANIZATION |
|---|---|
| c. ADDRESS (Include Zip Code) | d. TELEPHONE (Include Area Code)<br>(1) Commercial<br>(2) AUTOVON<br>(If applicable) |
| | 7. DATE SUBMITTED (YYMMDD) |

8. PREPARING ACTIVITY

| a. NAME<br>DEFENSE PERSONNEL SUPPORT CENTER<br>ATTN: DPSC-MSTE | b. TELEPHONE (Include Area Code)<br>(1) Commercial: AC 215-737-2870<br>(2) AUTOVON: AV 444-2870 |
|---|---|
| c. ADDRESS (Include Zip Code)<br>2800 SO. 20TH STREET<br>PHILADELPHIA, PA 19101-8419 | IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT:<br>Defense Quality and Standardization Office<br>5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466<br>Telephone (703) 756-2340 AUTOVON 289-2340 |

DD Form 1426, OCT 89 — Previous editions are obsolete

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:39Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:39Z
Check the source to verify that this is the current version before use.