MIL-A-36926A (DSA-DM)
21 May 1974
SUPERSEDING
MIL-A-36926 (DSA-DM)
27 November 1972

MILITARY SPECIFICATION

ANALYZING KIT, MEAT FAT CONTENT, 115 volt, 60 Hz

This limited coordination Military specification has been prepared by the Defense Personnel Support Center based upon currently available technical information, but it has not been approved for promulgation as a coordinated revision of MIL-A-36926 (DSA-DM). It is subject to modification. However, pending its promulgation as a coordinated Military specification, it may be used in procurement.

1. SCOPE

1.1 This specification covers a meat fat analyzing kit used to determine the percentage of fat in ground beef in a simple, fast and accurate manner.

2. APPLICABLE DOCUMENTS

2.1 Specifications and standards. The following documents of the issue in effect on date of invitation for bids or request for proposals form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

DD-G-541	Glass (Laboratory).
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-636	Box, Shipping, Fiberboard.

Military

MIL-P-116	Preservation, Methods of.
MIL-B-3854	Beef, Ground, Frozen.
MIL-M-7298	Manual, Technical: Commercial Equipment.

FSC 6695

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:35Z
Check the source to verify that this is the current version before use.

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

(Copies of specifications and standards required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposals shall apply.

AMERICAN NATIONAL STANDARDS INSTITUTE (ANSI)

ANSI C73-11	Caps and Receptacles, 125 Volts, 15 Amperes, 2 Pole, 3 Wire, Grounding Type, General Purpose.

(Application for copies should be addressed to the American National Standards Institute, 10 East 40th Street, New York, NY 10018.)

AMERICAN SOCIETY FOR TESTING AND MATERIALS (ASTM)

ASTM C148	Tentative Methods of Polariscopic Examination of Glass Containers.

(Application for copies should be addressed to the American Society for Testing and Materials, 1916 Race Street, Philadelphia, PA 19103.)

UNDERWRITERS' LABORATORIES, INC. (UL)

UL 499	Standards for Safety-- Electric Heating Appliances.

(Application for copies should be addressed to Underwriters' Laboratories, Inc., Walt Whitman Road, Melville, L.I., NY 11749; 207 East Ohio Street, Chicago, IL 60611; and 1655 Scott Boulevard, Santa Clara, CA 95050.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:35Z
Check the source to verify that this is the current version before use.

3. REQUIREMENTS

3.1 Material.

3.1.1 Glass. Glass shall conform to DD-G-541, Type I, Class A or Class B or Type II as specified.

3.1.2 Metal. Unless otherwise specified, metal shall be corrosion-resisting or shall be painted or plated to resist corrosion.

3.1.3 Rubber. Rubber shall be neoprene.

3.2 Fire and casualty hazards.

3.2.1 The contractor shall submit proof, prior to delivery, that the items to be delivered under the contract conform to the applicable requirements of UL 499. Proof shall consist of one of the following:

a. UL label or UL listing of the product, provided the contractor agrees to have Underwriters' Laboratories, Inc. forward a copy of the UL report directly to the contracting officer, upon the latter's request.

b. Laboratory report from a qualified independent laboratory. If an independent laboratory is used, it must be adequately equipped and competent to determine compliance with UL standards. The qualification of such laboratory shall be determined by the Government which reserves the right to require a facilities report and to conduct a facilities survey. Any laboratory report submitted shall include actual test methods, data taken, and results of tests. Detailed findings of examinations necessary to determine compliance shall also be included. The test report shall also have a section that describes the test sample in such a way that the report can be used to insure identity of production items with the test sample. This section shall be supplemented with construction information such as materials used, spacings between electrical components, types of bushings, photographs, wiring diagrams, sketches and parts list, all appropriately identified.

3.2.2 Compliance with above paragraph does not excuse compliance with all other requirements of the contract and specifications.

3.3 Design and construction. Overall dimensions of the case, when closed, shall be 6 inches in width, 11 inches in length, and 17 inches in height. Specified overall dimensions shall be within a tolerance of ± 1 inch. Case shall be of double-wall construction, painted carbon steel outer wall and anodized aluminum

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:35Z
Check the source to verify that this is the current version before use.

liner or inner wall. Case shall open into halves or compartments. The compartments shall be joined together along the 17-inch side by means of a metal piano type hinge, which is long enough to extend to within 1 inch of each end of the case.

* 3.3.1 Handle. The exterior shall have a metal handle firmly affixed to the side opposite the hinge. Handle shall be centrally located on the 17-inch side.

* 3.3.2 Catches. There shall be two metal catches capable of holding the case firmly and securely closed. The catches shall be equidistant from the handle. One catch shall be located to the left of the handle, and the other catch shall be located to the right of the handle.

* 3.3.3 Vent. A perforated vent shall be located on the top side of the left compartment, directly above the heating element. Vent shall be of such size and construction that it prevents accidental contact with the heating element.

3.3.4 Rubber feet. The case shall have four rubber feet on the hinged side and six rubber feet on the side opposite the perforated vent. Position of the feet shall be as follows: hinged side, one in each of the four corners; opposite the vent side, for each compartment, one in each of the two corners proximal to the hinge and one at the opposite end centrally located between the two corners. Rubber feet shall be securely bolted to the case. For purposes of this specification, the side which has six feet shall be defined as the bottom, and the closed case shall be used for positioning the feet.

* 3.3.5 Finish. The exterior surface of the case shall be finished with paint which does not peel due to excessive heat.

3.3.6 Interior. Interior walls and brackets shall be fabricated of unpainted metal. Interior walls or liner shall be fabricated of anodized aluminum.

3.3.6.1 Left compartment. The interior compartment shall securely house a timer and heating element in the top corners. These shall be securely mounted in a manner to prevent damage and to be readily replaceable. Two brackets properly spaced to allow winding of the electrical cord shall be secured on the side wall below the timer. On the side wall below the heating element there shall be a bracket for holding a funnel and positioning a test tube. On the bottom interior wall there shall be a rubber cushion for support of a test tube. In use, positioning from top to bottom shall be heater, sample screen, funnel and test tube. Sample screen shall be supported by the top rim of the funnel, and the test tube shall be held in position by the bracket and cushion. Stem of the funnel shall not extend into the test tube to a distance that interferes with testing. A spring hook arrangement shall be secured to the bracket. Spring hook shall prevent movement of the

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:35Z
Check the source to verify that this is the current version before use.

funnel while in transit. Rubber cushion shall be securely bolted. To the left and parallel to the test tube shall be a metal post, secured firmly at the bottom. On the post shall be a sliding hairline pointer and graduated scale.

3.3.6.2 Right compartment. Brackets shall be provided to securely hold two tubes, a funnel, two sample screens and two brushes. In addition, the interior wall shall be permanently and legibly imprinted with the procedure for analysis and cleaning. The limitations of the kit, and means of shaping the beef into doughnut form for analysis, shall also be included.

3.4 Accessories. Analyzing kit shall be furnished with the following accessories in the quantity specified:

Accessory	Quantity
Funnel	2
Test tube	3
Brush	2
Sample screen	2

* 3.4.1 Funnels. Funnels shall be fabricated of Type I borosilicate glass (see 3.1.1). Shall be capable of being secured to and removed from the funnel bracket in the left and right interior compartments. The top rim of the funnels shall furnish support for the sample screens.

* 3.4.2 Test tubes. Test tubes shall be fabricated of Type I or Type II glass (see 3.1.1). Two test tubes shall be capable of being secured to and easily removed from the test tube brackets in the right compartment. In operation, a third test tube shall rest beneath the funnel and upon a rubber cushion provided in the interior left compartment. However, this test tube shall be packaged separately within the case to prevent breakage in transit.

3.4.3 Brushes. Brushes shall be suitable for cleaning the sample screens, funnels, and test tubes. Shall be capable of being secured to and removed from the brush brackets in the right compartment. Brushes shall have a ring at the top for suspension from one bracket and shall be held by the lower bracket. One brush shall be of suitable diameter to clean the test tube; the other brush shall be of suitable diameter to clean the stem of the funnel and the perforated sample disks.

3.4.4 Sample screens. Sample screens shall be perforated disks fabricated of corrosion-resisting metal. Dimensions of the screen shall be such that the screen shall rest over the outer diameter of the funnel. Screen shall be constructed so that the fat drains completely and only into the funnel (see 3.4.1). Screen shall be capable of being secured and easily removed from the screen bracket.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:35Z
Check the source to verify that this is the current version before use.

3.5 Pointer and graduated scale. Sliding pointer and graduated scale shall move up and down on the metal post. Shall be secured to the post by means of an adjusting screw. Scale shall be graduated from 0 to 40. Each integer above 10 shall be marked with a graduation line and number. In addition, scale shall be marked to indicate "Percent Fat--Ground Beef". The pointer which correlates the amount of fat in the test tube with the scale shall slide easily over the graduated scale, and shall be sufficiently tight to retain its position. Hairline pointer shall be capable of being set at zero without obstruction due to the adjusting screw.

3.6 Heating element. Heating element shall be capable of furnishing sufficient heat to render fat from ground beef samples. Shall operate on 115 volts, 60 Hz, and provide a heating power of 500 watts.

3.7 Timer and bell. Shall be a variable mechanical timer, capable of operating for 15 minutes without resetting. Dial of timer shall be marked and numbered at 0, 5, 10 and 15. The timer shall have an accuracy of $\pm$ 30 seconds at 15 minutes when tested as specified in 4.4.1. The timer shall be equipped with a bell which shall sound when the timer returns to zero. The timer shall function to energize the circuit for the heating element. When the timer returns to the zero setting, the current to the heating element shall be cut off. The bell shall be audible at 50 feet.

3.8 Electrical cord. The electrical cord shall be a 3-wire cord with a plug at the end. The cord shall be not less than 6 feet in length and shall conform to ANSI C73-11.

3.9 Performance.

3.9.1 Accuracy. The average of results shall not vary by more than one unit when tested as specified in 4.4.2.

3.10 Strain. Funnels and test tubes shall contain no ring strain greater than a Temper No. 1 when tested as specified in 4.4.3. Any longitudinal strain shall be faint and highly diffused.

3.11 Identification marking. Exterior case shall have permanently affixed to it a metal nameplate, upon which shall be inscribed the following:

Name of manufacturer
Model number
Serial number
115 volts AC - 4.5 amps - 60 cycles
National Stock Number
Contract or purchase order number

Markings secured by adhesives shall not be acceptable.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:35Z
Check the source to verify that this is the current version before use.

3.12 Service data. The contractor shall furnish with each unit two copies each of a technical manual which contain complete instructions for installation, operation and maintenance and lists of component parts. The manual shall comply with the requirements of MIL-M-7298. As soon as practicable after award of the contract or purchase order, the contractor shall furnish to the contracting officer, via the cognizant Government inspector, two preliminary copies of his proposed technical manual, for review and comment by the Directorate of Medical Materiel, Defense Personnel Support Center. The preliminary manuals shall be submitted sufficiently early to permit adequate review by the Directorate of Medical Materiel, Defense Personnel Support Center (based on a maximum period of 30 days for such review) and to allow incorporation by the contractor of any required revisions and corrections in the final manual without delaying delivery under the contract or purchase order. In addition, the contractor shall furnish 15 copies of the approved technical manual to the contracting officer for distribution. Should the supplier have furnished acceptable manuals for the identical item within two years preceding the date of contract and propose to furnish manuals identical with those previously accepted, the requirements for review of drafts and submission of 15 additional copies may be waived upon request to the contracting officer. Waiver of submission of 15 additional copies shall be at the option of the Government.

* 3.13 Dimensions. Unless otherwise specified, dimensions of components and accessories shall be such that they are of a size suitable for performing the functions specified herein and those specified in the contractor's service data and instructions.

* 3.14 Weight. Kit with specified components and accessories shall weigh no more than 25 pounds.

* 3.15 Workmanship. Workmanship shall be first class throughout. Kit shall be free from defects listed in Table III and other defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:35Z
Check the source to verify that this is the current version before use.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificate of quality. Certificate of quality, supplied by the manufacturer of the glass, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the Item Identification, the name of the material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and Table I.

Table I For examination

	Inspection level	AQL (percent defective)
For visual examination		
Major defects	II	1.0
Minor defects	II	2.5
For dimensional examination	S-2	2.5

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and Table II. Acceptance number shall be zero.

Table II For tests

Component (unit of product)	Characteristic	Requirement	Test Procedure	Inspection level
Timer and bell	Timer accuracy	3.7	4.4.1	S-1
Kit	Accuracy	3.9.1	4.4.2	S-1
Glassware	Strain	3.10	4.4.3	S-2

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:35Z
Check the source to verify that this is the current version before use.

4.3 Examination. The kit shall be examined to determine compliance with all requirements contained in this specification. Examination shall be in accordance with, but shall not be limited, the defects listed in Table III.

Table III Classification of defects

Category	Defect
Major	
101	Component missing.
102	Metal components not free from burrs or sharp edges.
103	Painted or plated surfaces not free from blisters, chips and cracking.
104	Painted surfaces not free from blisters, chips, cracking, peeling or discoloration.
105	Interior brackets do not hold glass components securely.
106	Timer not secured to case.
107	Heating element not secured to case.
108	Handle or catches not secured to case.
109	Interior brackets not secure.
110	Timer not audible at 50 feet.
111	Heating element does not turn off when bell sounds.
112	Glass components not free from sharp edges, chipping, cracks or protrusions.
113	Graduations on scale not present, complete, correct and legible.
Minor	
201	Glass components not free from grease or dirt.
202	Identification marking and procedures listed on interior wall not present, complete, correct and legible.
203	Case, open or closed, does not rest on flat surface without wobbling.
204	Identification plate does not lie flat on the surface of the case.

* 4.3.1 Dimensional examination. The kit shall be examined for defects in dimensions and weights. Any dimension or weight not within the tolerance specified herein shall be classified as a defect.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

* 4.4.1 Timer. The timer furnished with the kit shall be tested for accuracy, using a stopwatch or an electric timer as a standard.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:35Z
Check the source to verify that this is the current version before use.

* 4.4.2 Accuracy. Samples shall be taken from a single lot of ground beef. Ten portions of the sample shall be tested for fat in accordance with MIL-B-3854; the results shall be recorded and averaged. In a room where the temperature is above 65° Fahrenheit, ten portions shall be tested in accordance with procedures delineated on the wall of the right interior compartment; the results shall be recorded and averaged. Average results of the two methods shall be compared.

4.4.3 Strain. Strain shall be tested in accordance with ASTM Designation C148, Method A or B.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One analyzer kit, as specified, constitutes one unit of issue.

5.1.2 Level A. Components comprising each kit shall be positioned in their respective compartments within the case and the case secured. The case shall be placed in a cellophane bag and further packaged in a box of appropriate size conforming to PPP-B-636, type CF, class domestic, 275 pound test. Box shall contain liner and top and bottom pads. All work shall be accomplished in a manner to prevent damage to the unit.

5.1.3 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing.

5.2.1 Level A. Each unit shall be overboxed in an exterior container designed for a type 1 load and constructed in accordance with PPP-B-601, overseas type, or PPP-B-636, class weather-resistant. Closure and strapping shall be as specified in the appendix of the applicable box specification.

* 5.2.2 Level B. No further overpacking shall be required.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

* 5.3 Marking. Interior and exterior containers shall be marked in accordance with MIL-STD-129.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:35Z
Check the source to verify that this is the current version before use.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

a. Title, number and date of this specification.
b. National Stock Number (NSN).
c. Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.2 This specification does not cover all types, classes, grades or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the type normally procured to meet military requirements.

6.3 This specification is applicable to the following item listed in the Federal Supply Catalog:

National Stock No.	Item Identification
6695-00-926-4502	ANALYZING KIT, MEAT FAT CONTENT, 115 volts, 60 Hz, 500 watts

6.4 The margins of this specification are marked with an asterisk to indicate where changes (additions, modifications, corrections, deletions) from the previous issue were made. This was done as a convenience only and the Government assumes no liability whatsoever for any inaccuracies in these notations. Bidders and contractors are cautioned to evaluate the requirements of this document based on the entire content irrespective of the marginal notations and relationship to the last previous issue.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6695-0039

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:35Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

MIL-A-36926A (DSA-DM)

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

a. GIVE PARAGRAPH NUMBER AND WORDING

b. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

☐ YES ☐ NO (If "Yes", in what way?)

4. REMARKS

SUBMITTED BY (Printed or typed name and address - Optional)

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426

REPLACES EDITION OF 1 JAN 68 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:35Z
Check the source to verify that this is the current version before use.