MIL-A-36950
12 February 1973

MILITARY SPECIFICATION

ASPIRATOR UNIT, SUCTION-PRESSURE APPARATUS

This specification has been approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 This specification covers a sinus vacuum cleaner with nasal tip and two cannulas suitable for eustachian and post nasal procedures.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposals, form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

GG-I-526	Instruments, Dental and Surgical; General Specifications.
PPP-B-566	Boxes, Folding, Paperboard.
PPP-B-585	Box, Wood, Wirebound.
PPP-B-601	Box, Wood, Cleated-plywood.
PPP-B-621	Box, Wood, Nailed and Lock-corner.
PPP-B-636	Box, Fiberboard.
PPP-B-676	Box, Set-up.

Military

MIL-L-10547	Liners, Case and Sheet, Overwrap; Water-Vaporproof or Waterproof.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:38Z
Check the source to verify that this is the current version before use.

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

DRAWING

Defense Personnel Support Center (DPSC)
Drawing No. 23277 Aspirator Unit, Suction-Pressure Apparatus

(Copies of specifications, standards, and drawing required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposals shall apply:

AMERICAN SOCIETY FOR TESTING AND MATERIALS (ASTM)

ASTM E-36 Brasses, Chemical Analysis of.

(Application for copies should be addressed to American Society for Testing and Materials, 1903 Race Street, Philadelphia, Pa. 19103.)

3. REQUIREMENTS

3.1 Material. The material shall be as indicated herein and of highest quality and suitable in all respects for the purpose intended. All components of the aspirator unit shall be capable of being autoclaved.

3.1.1 Metal parts. All metal parts, except the gasket and body delivery tube, shall be brass conforming to Alloy Nos. 260, 270 or 360 of Class 8 of GG-I-526. The parts shall be nickel plated or chromium over nickel plated in accordance with Class 8 requirements of GG-I-526.

3.1.2 Bottle. The bottle shall be made of glass and the metal ring shall be cemented to it. The bottle and ring assembly shall be capable of being autoclaved.

3.1.3 Gasket. The gasket for the body-bottle joint shall be made of lead.

3.1.4 Body delivery tube. The body delivery tube shall be austenitic stainless steel.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:38Z
Check the source to verify that this is the current version before use.

3.2 Style, design, and dimensions. Shall be a sinus vacuum cleaner complete with a nasal tip and two cannulas suitable for eustachian and post nasal procedures. The aspirator unit shall be for use with Operating Apparatus, Ear, Nose and Throat, 6515-935-4112. The aspirator unit shall consist of a bottle with a metal ring having a female thread, body with a male thread and short delivery tube, nasal tip, a large and small cannula and finger valve. The nasal tip and the cannulas shall be interchangeable in order to fit one port of the body. The finger valve shall fit the other port of the body. The aspirator unit shall be in accordance with Defense Personnel Support Center Drawing No. 23277. The replacement bottle with attached metal ring, for the aspirator unit, is obtainable as Federal Stock No. 6515-597-9948.

3.3 Corrosion-resistance. The unit shall comply with salt spray resistance requirements in accordance with GG-I-526.

3.4 Workmanship. Workmanship shall be first class throughout. The aspirator unit shall be free from defects which detract from its appearance or may impair its serviceability.

4. QUALITY ASSURANCE PROVISIONS.

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:38Z
Check the source to verify that this is the current version before use.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the components of the aspirator unit, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the FSN, the item identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as indicated in Table I.

Table I. Sampling for Examination

Type of defect	Inspection level	AQL* (percent defective)
Major	S-3	1.0
Minor	S-2	1.5

* If acceptance number is not exceeded, lot shall be accepted. If rejection number is equaled or exceeded, inspection at Level II shall be performed to determine final rejection or acceptance.

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as indicated in Table II below:

Table II. Sampling for Tests

Characteristic	Requirement	Test	Insp. level	AQL
Composition	3.1.1	4.4.1	S-1	1.0
Autoclaving	3.1	4.4.2	S-2	1.0
Corrosion resistance	3.3	4.4.3	S-1	1.0

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:38Z
Check the source to verify that this is the current version before use.

4.3 The aspirator unit shall be examined to determine compliance with all the requirements contained in this specification. Examination shall be conducted as specified in Tables I and III. Examination, however, shall not be restricted to defects listed in Table III. Any aspirator units in the sample containing one or more defects shall be rejected and if the number of defective aspirator units in any sample exceeds the acceptance number for that sample, the lot represented by that sample shall be rejected.

Table III. Classification of Defects

Categories	Defects
Major	
101	Components do not assemble smoothly and without difficulty.
102	Components do not remain attached during normal handling.
103	Dimensions related to interchangeability are not in agreement with applicable drawing.
104	Dimensions of nasal tip and OD and length of cannulas are not in agreement with applicable drawing.
Minor	
201	Dimensions not related to fit and assembly or function are at variance with applicable drawing.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Chemical composition. The chemical analysis of the brasses shall be accomplished in accordance with ASTM Standard Method E 36, Chemical Analysis of Brasses.

4.4.2 Autoclaving. All components of the aspiration unit shall withstand exposure to steam at a temperature of 250° ± 3.6°F (121° ± 2°C) for 20 minutes without showing any deterioration. There shall be no blistering of plating or other deterioration of any component. The bond of the metal ring to the bottle shall not be damaged.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:38Z
Check the source to verify that this is the current version before use.

4.4.3 Corrosion resistance. All components of the aspirator unit shall be exposed to the salt spray test as specified in GG-I-526. The presence of some corrosion on the lead gasket shall not be cause for rejection.

5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be level A or C as specified (see 6.1).

5.1.1 Unit of issue. One aspirator-atomizer unit (complete) constitutes one unit of issue. One medicinal atomizer bottle constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in Table IV shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table IV. Packaging Quantities

Item	Unit package	Intermediate package	Exterior container
Aspirator Unit (Complete)	1 unit	Not required	25 units
Bottle	1 unit	12 units	144 units

5.1.2.1 Packing variation permitted. If the required number of units in the entire shipment is less than the number of units specified to be overpacked in an exterior container, such units may be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Unit package.

5.1.3.1.1 Aspirator unit. All components comprising the complete unit shall be thoroughly cleaned, wrapped in neutral tissue or other suitable anticorrosive material and packaged in a box of appropriate size constructed in accordance with PPP-B-676 except that commercial colors are acceptable. The box shall be designed with suitable interior

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:38Z
Check the source to verify that this is the current version before use.

fittings or shall contain sufficient cushioning material to insure against shifting of or damage to the contents. Closure shall be adequate to prevent spilling of contents under normal handling.

5.1.3.1.2 Bottle. Each bottle shall be thoroughly cleaned wrapped in neutral tissue or other suitable anti-corrosive material and packaged in a box of appropriate size constructed in accordance with PPP-B-566 or PPP-B-676 except that commercial colors are acceptable. Closure shall be adequate to prevent spilling of contents under normal handling.

5.1.3.2 Intermediate package. Intermediate package shall be a box of appropriate size constructed in accordance with PPP-B-676 except that commercial colors are acceptable or PPP-B-636, type CF, class domestic. Closure shall be adequate to prevent spilling of contents under normal handling.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.1).

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant. Fiberboard boxes shall conform to the special requirements specified in Table II of PPP-B-636.

5.2.1.2 Case liner. Each level A wood box shall be lined with a water-proof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes.

5.2.1.3 Closure. Closure of wood boxes shall be in accordance with appendix of applicable box specification. Closure of each fiberboard box shall be as specified in the appendix, and waterproofing shall conform with 30.4 of PPP-B-636.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:38Z
Check the source to verify that this is the current version before use.

5.2.1.4 Strapping. Strapping, when required, shall be in accordance with appendix of applicable box specification.

NOTE: Strapping shall not be required for shipments forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure of wood boxes shall be as specified in the appendix of the applicable box specification. Closure of fiberboard boxes shall conform to method II of PPP-B-636. Fiberboard boxes shall conform to the special requirements specified in Table I of PPP-B-636.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Interior packages. Each unit package and each intermediate package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Ordering data. Purchasers should exercise any desired options offered herein. and the procurement documents should specify the following:

(a) Title, number, and date of this specification.
(b) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:38Z
Check the source to verify that this is the current version before use.

6.3 This specification is intended to cover the following item appearing in the Federal Supply Catalog:

Federal Stock No.	Item Identification
6515-104-9583	ASPIRATOR UNIT, SUCTION-PRESSURE APPARATUS
6515-597-9948	BOTTLE, Medicinal, Atomizer, Glass

It does not include all types or sizes as indicated by the title nor all types which are commercially available.

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6515-1703

Review activities:

Army - MD
Navy - MS
Air Force - 03

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:38Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers, vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

MIL-A-36950

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426

REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:38Z
Check the source to verify that this is the current version before use.