MIL-A-36955
AMENDMENT-1
9 May 1974

MILITARY SPECIFICATION

ANOSCOPE, FANSLER, 6 cm

This amendment forms a part of Military Specification MIL-A-36955, dated 26 July 1973, and is approved for use by all Departments and Agencies of the Department of Defense.

Page 1:

1.1 Scope. Delete "practological" and insert "proctological".

Page 2:

3.1 General requirements. Delete entirely and substitute:

"3.1 General requirements. The anoscope shall be in accordance with the applicable requirements of GG-I-526, and as specified herein."

3.3 Hardness. Delete entirely and substitute:

"3.3 Hardness. Rockwell hardness of the anoscope shall be B75 to B85."

3.4 Style, design and dimensions. In last line, delete "obdurator" and insert "obturator".

Add following paragraph:

"3.5.1 Passivation. All components of the anoscope shall be properly passivated."

Page 5:

4.2.2 Sampling for tests. Delete Table II and notes under Table II entirely and substitute:

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:38Z
Check the source to verify that this is the current version before use.

Page 5: (Cont'd)

"Table II. Sampling for Tests

Component (unit of product)	Characteristic	Require-ment	Test procedure	Inspection level	AQL (percent defective)
1-Anoscope	Hardness	3.3 and GG-I-526	GG-I-526	S-2	1.5
1-Anoscope	Dimensions	3.4	-------	S-2	2.5
1-Anoscope	Boil test	GG-I-526	GG-I-526	S-2	1.5

."

Add the following paragraph:

"4.2.3 Raw materials. A sufficient quantity of each lot of each metal used in the construction of the anoscope shall be selected for chemical analysis."

Page 7:

5.1.3.5 Intermediate package. Delete entirely.

Page 8:

5.3.1 Interior packages. Line 1, delete "and each intermediate package."

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6515-1853

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:38Z
Check the source to verify that this is the current version before use.