MIL-B-36929 (DSA-DM)
INTERIM AMENDMENT-1 (DSA-DM)
28 March 1973

MILITARY SPECIFICATION

BOTTLE, URINE SPECIMEN, SHIPPING

This Interim Amendment is issued for use by the Department of Defense, Defense Personnel Support Center with limited coordination Military Specification MIL-B-36929 (DSA-DM) dated 7 December 1972.

Page 2

3.1.2 Bottle. After first sentence, insert: "The bottle shall be cylindrical with a threaded finish on top and shall have a rounded shoulder below the bottom level of the neck." Third line below end of Table I, delete "below the threaded portion of the opening" and substitute "90 ml. ± 15 ml."

Page 5

4.4.1 Capacity fill. Line 2, delete "90 ml. of water." and substitute "water up to a point that shall not extend above the bottom level of the neck or below the bottom level of the shoulder. The volume of water used to fill the bottle shall be measured and recorded."

Custodians:
Army - MD
Navy - MS
Air Force 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6640-0788

FSC 6640

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:35Z
Check the source to verify that this is the current version before use.