MILITARY SPECIFICATION

BOTTLE, URINE SPECIMEN, SHIPPING

This limited coordination Military specification has been prepared by the Defense Personnel Support Center based upon currently available technical information, but it has not been approved for promulgation. It is subject to modification. However, pending its promulgation as a coordinated Military specification, it may be used in procurement.

1. SCOPE

1.1 This specification covers one type of compartmented shipping box (set) containing twelve (12) 3-ounce (oz.) (90 milliliter (ml.)) capacity plastic bottles with lined screw caps, suitable for mailing urine specimens to laboratories.

2. APPLICABLE DOCUMENTS

2.1 Specifications and standards. The following documents of the issue in effect on date of invitation for bids or request for proposals form a part of this specification to the extent specified herein:

SPECIFICATION

Federal

PPP-B-636	Boxes, Fiberboard.

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

(Copies of specifications and standards required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:35Z
Check the source to verify that this is the current version before use.

3. REQUIREMENTS

3.1 Components.

3.1.1 Shipping box. The shipping box and partitions shall be fabricated and assembled in accordance with the applicable subparagraphs of 5.1. The shipping box and partitions shall be free of tears, cuts, holes, punctures and deformation.

3.1.2 Bottle. Shall be a wide-mouth bottle of one-piece construction fabricated of white, translucent or opaque high density polyethylene. The open end of the bottle shall have the following characteristics to accommodate a screwcap closure (see 3.1.3):

a. A continuous type exterior thread.

b. A flat top surface at the opening that is level and free of ridges.

The bottle shall have a 3-oz. (90 ml.) capacity and shall conform to the requirements of Table I for dimensions.

Table I Dimensions (inches)

Inner diameter of opening	1-3/16 to 1-5/16
Overall height (without cover)	2-7/8 to 3-1/8
Wall thickness	1/64 min.
Overall width	2 max.

The bottle shall have a symmetrical shape and shall remain in an upright position when placed on a flat and level surface. When tested in accordance with 4.4.1, the fill of the bottle shall be below the threaded portion of the opening. The exterior surfaces of the bottle shall be smooth, except that impressions from commercial markings and seams from mold marks shall be acceptable but free of flash. The bottle shall be free of cracks, chips, breaks, deformation, sharp edges and burrs.

3.1.3 Screw cap. Shall be either a one-piece metal cap or a one-piece molded plastic cap of rigid construction having inner edges sufficiently threaded with a continuous type thread for closure on the bottle (see 3.1.2). A liner shall be secured to the inside of the cap in such a manner that it does not fall out of the cap, or is not off-center when in the closure position on the bottle. The liner shall be supplied with a facing that is impervious to water and urine. The liner shall be of sufficient thickness and resiliency to form an effective seal with the flat top surface of the bottle (see 3.1.2). When tested in accordance with 4.4.2, the cap shall not crack, split or break; there shall be no evidence of leakage. The cap shall be free of deformation, cracks, sharp edges, burrs, protrusions or flash. Manufacturer's normal commercial finish for the plastic cap shall be acceptable. The metal cap shall have a suitable finish to resist corrosion.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:35Z
Check the source to verify that this is the current version before use.

3.2 Workmanship. Workmanship shall be first class throughout. The bottles, caps and shipping boxes shall be free from defects which detract from their appearance or may impair their serviceability. All components shall be supplied clean, dry and free of dirt or foreign matter.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted on each lot of material in accordance with MIL-STD-105 and Table II. As applicable, the units of product for inspection purposes shall be the individual bottle, cap and shipping box.

Table II For examination

	Inspection level	AQL (percent defective)
For visual examination		
Major defects	II	2.5
Minor defects	II	4.0
For dimensional examination	S-3	2.5

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:35Z
Check the source to verify that this is the current version before use.

4.2.2 For tests. Sampling for tests shall be conducted on each lot of material in accordance with Table III and MIL-STD-105. The unit of product for the leakage test shall be the individual bottle with cap.

Table III For tests

Component	Characteristic	Requirement	Test Procedure	Inspection level	AQL
Bottle	Capacity fill	3.1.2	4.4.1	*	**
Bottle and cap	Leakage	3.1.3	4.4.2	S-1	**

* One bottle per mold cavity.
** Acceptance number shall be zero.

4.3 Examination. All components shall be examined to determine compliance with requirements contained in this specification.

4.3.1 Classification of defects. Examination shall be conducted in accordance with Table IV. Examination shall not be restricted to the classified possible defects listed in the table.

Table IV Classification of defects

Categories (Bottle or cap)	Defects
Major	
101	Bottle or cap not free of cracks, chips or breaks.
102	Bottle or cap not free of deformation.
103	Bottle or cap not free of sharp edges or burrs.
104	Top of bottle not flat or not free of discernible ridges.
105	Bottle not clean, dry or ready for use.
106	Bottle not free of foreign matter. (Note: Application of cap to bottle during capping process may result in deposit of some small plastic or fiber shavings in the bottle. Presence of a few of these shavings shall not be considered as a defect.)
107	Liner not aligned or secured within screw cap.
108	Bottle not stable when placed on a flat and level surface.
109	Thread on cap or bottle not free of stripping, distortion, or discontinuity.
110	Cap not capable of being secured to bottle.
111	Metal caps not free of corrosion.
Minor	
201	Bottle or cap not free of scratches, blemishes or pits.
202	Cap not clean.
203	Bottle or cap not free of imbedded foreign material.
204	Bottle or cap not free of flash.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:35Z
Check the source to verify that this is the current version before use.

Classification of defects (cont'd)

(Bottle or cap)	
Major	
301	Partitions not complete or properly assembled in box.
302	Box not free of deformation.
303	Box not free of separation.
304	Box not free of tears, cuts, holes, splits or punctures.
305	Box not capable of being properly closed.
306	Box or partitions not free of moisture.
Minor	
401	Partition not free of deformation.
402	Box or partition not free of dirt or foreign matter.
403	Marking on box not complete, legible or present.

4.3.2 Dimensional examination. The bottles shall be examined for defects in dimensions. Any dimension not within the specified limits shall be classified as a defect.

4.4. Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Capacity fill. The test shall be conducted at a room temperature of 25° Centigrade (°C.) ± 5°. A bottle shall be filled with 90 ml. of water.

4.4.2 Leakage. The bottles shall be filled with water to the threaded portion, and closures shall be applied handtight. The bottles shall be suspended in an inverted position for two hours at room temperature, and for four hours at 120° Fahrenheit (°F.) ± 5°.

5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be Level A.

5.1.1 Unit of issue. One package, comprising 10 individual boxes containing 12 capped bottles, as specified, constitutes one unit of issue.

5.1.2 Level A.

5.1.2.1 Individual box. Twelve (12) bottles shall be packaged in a box of appropriate size conforming to PPP-B-636, type CF, class domestic, style CSSC, having a minimum bursting strength of 275 pounds. Box design shall include full height chipboard partitions so arranged that each capped bottle is contained in an individual cell. Top flaps shall not be glued, taped or stapled. Bottom flaps shall be closed in accordance with method I or II of PPP-B-636. If staples are used for closure, a bottom pad shall be furnished to prevent contact of the staples with the individual bottles.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:35Z
Check the source to verify that this is the current version before use.

5.1.2.2 Unit package. Ten (10) individual boxes shall be packed in a box of appropriate size conforming to PPP-B-636, type CF, class domestic, having a minimum bursting strength of 200 pounds. Closure of box shall be in accordance with method I or II of PPP-B-636.

5.2 Packing. No further overpacking shall be required.

5.3 Marking.

5.3.1 Individual box. Individual box shall be marked on two sides with the word "UP" with an arrow toward the top of the container. No other markings shall be shown.

5.3.2 Unit package. Each unit package shall be marked in accordance with MIL-STD-129, as specified for the exterior container.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

a. Title, number and date of this specification.
b. Federal Stock Number (FSN).
c. Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.2 This specification is applicable to the following item listed in the Federal Supply Catalog:

Federal Stock Number	Item Identification
6640-165-5778	BOTTLE, URINE SPECIMEN, SHIPPING, 120s

6.3 This specification does not cover all types, classes, grades or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the types which are normally procured to meet military requirements.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force -03

Preparing activity:
Defense Supply Agency - DM

Project No. 6640-0781

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:35Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/ vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE
MIL-B-36929 (DSA-DM)

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?
   A. GIVE PARAGRAPH NUMBER AND WORDING.

   B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

   ☐ YES ☐ NO *(If "Yes", in what way?)*

4 REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426 REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:35Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, Pa. 19101

Defense Supply Agency

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304

U.S. MAIL

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 20260-21T09:35Z
Check the source to verify that this is the current version before use.