MILITARY SPECIFICATION

DRAPES, SURGICAL, DISPOSABLE, Nonwoven Fabric

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 This specification covers sterile, disposable drapes for use in surgical operating rooms.

1.2 Classification. The drapes shall be of the following sizes and classes as specified (see 6.2).

Class	Size
A - split sheet	87 x 120
B - half sheet	44 x 72
C - large sheet	87 x 104

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect, on date of invitation for bids or request for proposals, form a part of this specification, to the extent specified herein:

SPECIFICATIONS

Federal

PPP-B-636	Box, Fiberboard.

Military

MIL-S-36586	Sterilization Test Strip Set, Bacterial Spore.
MIL-F-36901	Fabric, Nonwoven, Disposable (Components for Surgical Packs).

FSC 6530

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:34Z
Check the source to verify that this is the current version before use.

STANDARDS

Military

MIL-STD-105 Sampling Procedures and Tables for Inspection by Attributes.

MIL-STD-129 Marking for Shipment and Storage.

(Copies of specifications and standards required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer).

2.2 Other publications. The following document forms a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposals shall apply.

U. S. PHARMACOPEIAL CONVENTION PUBLICATION:
Pharmacopeia of the United States.

(Application for copies should be addressed to the Mack Publishing Company, Easton, Pa.)

3. REQUIREMENTS

3.1 Nonwoven fabric, dimensions, quantity and assembly of component. Components in the immediate package shall be functionally folded and sequence packed for use. Components shall conform to the requirements specified in table I.

TABLE I

Type Nonwoven Fabric - MIL-F-36901	Component	No. Required	Minimum Dimension (inches) Before Sterilization
Class A - split sheet			
I - IA - IB - IC	Outerwrap	1	Adequate to maintain sterility
III - IIIA - IIIB	Drape sheet with 44 inch long tail	1	120 x 87

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:34Z
Check the source to verify that this is the current version before use.

TABLE I (Cont'd.)

Type Nonwoven Fabric - MIL-F-36901	Component	No. Required	Minimum Dimensions (inches) Before Sterilization
Class B - half sheet			
I - IA - IB - IC	Outerwrap	1	Adequate to maintain sterility
III - IIIA - IIIB	Drape sheet	1	72 x 44
Class C - large sheet			
I - IA - IB - IC	Outerwrap	1	Adequate to maintain sterility
III - IIIA - IIIB	Drape sheet	1	104 x 87

3.2 Sterility. The drape pack in its immediate container shall be sterile when tested with a biological indicator as specified in 4.3.

3.2.1 Certificate of sterility. A certificate of sterility, stating that samples from each sterilizer lot have been tested and found to be sterile, shall be submitted to the procuring activity. The certificate of sterility shall include:

a. Item Identification.
b. Federal stock number.
c. Contract or purchase order number.
d. Sterility lot or control number.
e. Date of sterilization.
f. Quantity of product in sterilizer lot.
g. Type of sterilization (gas, radiation, heat, steam, etc.).
h. Form of biological indicator (i.e., inoculated paper spore strip or carrier, inoculated product, or inoculated simulated product).
i. Test organism(s) and population density used.
j. Test medium, incubation temperature, incubation time, and dates of tests.
k. Number of test units.
l. Statement that samples comply with sterility tests.
m. Signature of head of laboratory or a designated qualified alternate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:34Z
Check the source to verify that this is the current version before use.

When inspection is made at destination, the certificate of sterility shall be submitted by the contractor to the Directorate of Medical Materiel, ATTN: DPSC-ATQ, Defense Personnel Support Center, 2800 South 20th Street, Philadelphia, Pa., 19101, and a copy forwarded with each shipment to the consignee. When inspection is made at source, certificate of sterility shall be furnished to the cognizant Government Quality Assurance Representative for submission to the procuring activity.

3.2.2 Sterility lot. A sterility lot is that single quantity of product subjected to the same manufacturing operation and simultaneously sterilized in the same sterilizer chamber. Each unit package shall be clearly marked with a sterility lot (control) number and date of sterilization as specified in 5.3.1. Each sterility lot shall be traceable to the pertinent manufacturing lot(s).

3.3 Folding. All components including the outerwrap shall be folded for optimum aseptic techniques.

3.4 Instructions. The immediate wrapper and all components must be marked to indicate opening, unfolding methods, and instructions for use. The ink shall be sterilizable.

3.5 Workmanship. Workmanship shall be first class throughout. The drape packs shall be free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available for review to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:34Z
Check the source to verify that this is the current version before use.

4.1.2 Inspection. Inspection, as used in this specification is defined as both examination (such as visual or auditory) investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the components listed in 3.1, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the FSN, the item identification, the name of the component or material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements. Where certificates of quality are submitted, the Government reserves the right to check test such items to determine the validity of the certification.

4.2 Visual examination of the pack. Defects found during this examination shall be classified in accordance with table III. The sample unit for this examination shall be one pack.

4.2.1 Sampling for examination. Sampling for examination shall be performed in accordance with MIL-STD-105. The inspection levels and acceptable quality levels shall be as shown in table II. The sample unit for examination shall be one pack.

TABLE II

	Inspection level	AQL (percent defective)
Major Defects	S-3	1.0
Minor Defects	S-3	2.5

4.2.2 Classification of defects (pack). The pack shall be visually examined for the defects listed in table III.

TABLE III. Classification of defects

Major	
101	Any component missing or other than specified.
102	Components not functionally folded.
103	Printed instructions not free of bleeding into material.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:34Z
Check the source to verify that this is the current version before use.

TABLE III. Classification of defects (Cont'd.)

Major		
	104	Printed instructions omitted, incorrect, illegible, misplaced or not on proper components.
	105	Any component not free of holes and tears.
	106	Any component not clean, not free from foreign material.
Minor		
	201	Outer wrap torn or ripped.
	202	Components not proper length and width.
	203	Unit marking not complete, illegible.

4.3 Sterility test. The drape pack shall be tested for sterility in accordance with the method specified in the USP for purified cotton, gauze, surgical dressings and related material using inoculated product, inoculated simulated product, or paper spore strip (carrier). Product containing a biological indicator shall be marked or labeled so that it clearly differentiates from product intended for distribution. The incubation medium and minimum time in each case shall be soybean casein digest for 7 days for inoculated product, inoculated simulated product, or paper spore strip (carrier). The incubation temperature shall be as indicated in paragraph 4.3.1. The minimum number of test units, regardless of lot size, shall be 10, except for inoculated carrier which shall be 20 for Type I Sterilization and 30 for Type II Sterilization including inoculated carrier and product test units.

4.3.1 Inoculated product or simulated product. For each type of sterilization, the organism used shall be as listed below:

Type of Sterilization	Organism	Incubation Temperature
Steam	B. stearothermophilus	55°C - 65°C
Dry Heat	B. stearothermophilus, or	55°C - 65°C
	B. subtilis var. niger (or globigii)	30° - 35°C
Gas	B. subtilis var. niger (or globigii)	30°C - 35°C
Radiation	B. pumilus	30° - 35°C

The number of viable cells shall be sufficient to provide assurance that the product has been sterilized.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:34Z
Check the source to verify that this is the current version before use.

4.3.2 Inoculated carrier. Inoculated carriers (paper spore strips) if used shall be in accordance with MIL-S-36586. Other paper carriers shall be inoculated to be of equivalent resistance. The method incubation and number of test units shall be as indicated in 4.3 Sterility test; and 4.3.1. Inoculated product or simulated product. The method of sterilization will control the organism in the spore strip to be tested for as shown in 4.3.1. NOTE: When an inoculated carrier is chosen, product samples must also be tested as required by the USP.

4.3.3 Alternate method. Suppliers conducting sterility tests or designating lots in a manner other than that indicated herein shall submit proof that the alternate method provides equal assurance concerning the sterility of the product. However, all sterility tests shall be conducted using biological indicators. Requests for approval of any alternate method of testing for sterility shall be made to the Contracting Officer, and approval obtained before employment of the alternate method on a government contract.

4.3.4 Product testing. Testing for sterility of the product after receipt by the government shall be without a biological indicator and in accordance with the USP.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One package, containing ten drapes, as specified, constitutes one unit of issue.

5.1.2 Level A.

5.1.2.1 Individual package. Each folded drape and overwrap shall be assembled as specified in 3.1 and enclosed in a heat sealed, sterilizable, tamperproof, white 50 lb. (24 x 36 - 500) kraft barrier paper bag. The bag shall be provided with a tear string or tear tape to facilitate opening. The bag shall be capable of maintaining sterility of contents unless damaged or opened. In addition, each kraft bag shall be further inserted into a clear polyethylene bag having a minimum thickness of 0.003 inches. Closure of the polyethylene bag shall be adequate to prevent spilling of contents under normal handling.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:34Z
Check the source to verify that this is the current version before use.

5.1.2.2 Unit package. Each unit (10 drapes) shall be packaged in a box of appropriate size constructed in accordance with PPP-B-636, type CF, class weather-resistant. Closure shall be as specified in the appendix and waterproofing shall conform to 30.4 of PPP-B-636.

5.1.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.2 Packing.

5.2.1 Levels A and C. Units packaged as specified in 5.1.2 and 5.1.3 shall require no further packing.

5.3 Marking.

5.3.1 Individual package.

5.3.1.1 Paper bag. In addition to any commercial markings, each paper bag shall be marked in accordance with the requirements of the Federal Food, Drug and Cosmetic Act. In addition marking shall include the name of contractor and the legend: "STERILE" and "STERILITY GUARANTEED UNLESS PACKAGE IS DAMAGED OR OPENED".

5.3.2.2 Plastic bag. Each polyethylene bag shall be marked with the legend: "CAUTION: PLASTIC BAG SHOULD BE REMOVED PRIOR TO USE IN HAZARDOUS ATMOSPHERE". As an alternate, the caution legend may be shown on the strip of tape used to close the polyethylene bag.

5.3.2 Unit package. Each unit package shall be marked as required for the exterior container specified in MIL-STD-129. Marking shall include the sterility lot (control) number and the method of sterilization (G-Gas, R-Radiation, S-Steam, C-Cold, D-Dry Heat sterilization) shall be shown following the level packaging and date of packaging. Example: "A-10/72-G". Final sterilization shall be on or after indicated packaging date.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:34Z
Check the source to verify that this is the current version before use.

6. NOTES

6.1 Intended use. The drapes covered by this specification are for use in surgical operating rooms of military medical treatment facilities.

6.2 Ordering data. Procurement documents should specify the following:

(a) Title, number and date of this specification.
(b) Class and size required (see 1.2).
(c) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.3 This specification does not include all types, classes, grades, or sizes, of the commodity indicated by the title of this specification, or which are commercially available, but is intended to cover the types which are normally procured to meet military requirements.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6530-1266

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:34Z
Check the source to verify that this is the current version before use.