MILITARY SPECIFICATION

FORCEPS, HEMOSTATIC, STRAIGHT, RANKIN, 6-1/4 INCH

This amendment forms a part of Military Specification MIL-F-36937, dated 25 April 1973 and is approved for use by all Departments and Agencies of the Department of Defense.

Page 2:

3.3 Hardness. Line 2, delete "C55" and substitute "C45".

3.4 Style, design and dimensions. Last sentence, delete period (.) and add: "with 3x3 ratchets."

5.1.3.5 Intermediate package. Delete entirely and substitute:

"5.1.3.5 Intermediate package. Intermediate package shall be a box of appropriate size and design constructed in accordance with PPP-B-566 or PPP-B-676, except that commercial colors will be acceptable; or PPP-B-636, type CF, class domestic. Closure shall be adequate to prevent accidental opening under normal handling."

5.2 Packing. Delete entirely and substitute:

"5.2 Packing.

"5.2.1 Level A.

"5.2.1.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant, grade V3c. Closure and strapping shall be as specified in the appendix of the applicable box specification. Special requirements of applicable table shall apply for PPP-B-636 box.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:36Z
Check the source to verify that this is the current version before use.

"5.2.1.2 Case liner. Each Level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes. Fiberboard box shall be waterproof in accordance with paragraph 30.4 of PPP-B-636.

"5.2.2 Level B.

"5.2.2.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure shall be as specified in the appendix of the applicable box specification. Fiberboard box shall be closed by method II of PPP-B-636. Special requirements of applicable table shall apply for PPP-B-636 box."

Custodians:

Army - MD
Navy - MS
Air Force - 03

Review activities:

Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6515-1865

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:36Z
Check the source to verify that this is the current version before use.