MIL-F-36940A
20 October 1975
SUPERSEDING
MIL-F-36940
30 March 1973

MILITARY SPECIFICATION

FORCEPS, HEMOSTATIC, PEAN, 9 INCH

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 This specification covers two types of hemostatic forceps of the Pean design, suitable for clamping large vessels of the pelvis and the broad ligament during surgical procedures.

1.2 Classification.

1.2.1 Types. The forceps shall be of two types, as specified (see 6.3).

Type I Curved

Type II Straight

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposals, form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

GG-I-526	Instruments, Dental and Surgical; General Specifications.
PPP-B-566	Boxes, Folding, Paperboard.
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Boxes, Shipping, Fiberboard.
PPP-B-676	Boxes, Set-up.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:37Z
Check the source to verify that this is the current version before use.

Military

MIL-P-116	Preservation, Method of.
MIL-L-10547	Liners, Case, and Sheet Overwrap; Water-Vaporproof or Waterproof, Flexible.

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

(Copies of specifications and standards, required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 General Requirements. The forceps shall be in accordance with the applicable requirements of GG-I-526 and as specified herein.

3.2 Material. The forceps shall be fabricated from Class 4, Type 410, 420, or 420F material.

* 3.3 Hardness. After proper stress relieving, the forceps shall have a Rockwell hardness of C40 to C47 with the exception of the screw pin.

3.4 Style, design and dimensions. The forceps shall be of the Pean design, suitable for clamping large vessels of the pelvis and the broad ligament during surgical procedures. Shall be in accordance with Figure 1.

3.4.1 Locks. Shall be box type lock.

* 3.5 Finish. Finish shall be mirror or satin, non-reflective. Locks and ratchets shall be smooth and shall be polished but need not be buffed to a mirror finish.

3.6 Identification marking. In addition to GG-I-526 requirements, forceps of foreign manufacture shall also be marked "stainless."

3.7 Jaw and ratchet operation. When tested as specified in 4.4.1, the jaw tips shall close first and the jaws shall progressively tighten with each ratchet position. Each ratchet position shall hold firmly and release instantly.

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:37Z
Check the source to verify that this is the current version before use.

FIGURE 1.

FORCEPS, HEMOSTATIC, PEAN 9 INCH

Source: https://assist.dla.mil -- Downloaded: 202640-21T09:37Z
Check the source to verify that this is the current version before use.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection shall be in accordance with GG-I-526.

4.1.2 Records. Records of examination and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.3 Certificates of quality. Certificates of quality supplied by the manufacturer of the corrosion resisting steel, may be furnished in lieu of actual performance of such testing by the contractor provided lot identity has been maintained and can be demonstrated to the Government. The certificates shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the Item Identification, the name of the material, the lot number, the lot size, the sample size, the date of testing, the test method, the individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as indicated in Table I. The unit of product for examination purposes shall be one hemostatic forceps.

Table I. Sampling Criteria for Examination

Type of defect	Inspection level	AQL (Percent defective)
Major	II	1.0
Minor	II	2.5

Source: https://assist.dla.mil -- Downloaded: 202640-21T09:37Z
Check the source to verify that this is the current version before use.

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, and as indicated in Table II. Unit of product for test purposes shall be one forceps.

Table II. Sampling Criteria for Tests

Component	Characteristic	Requirement	Test procedure	Inspection level	AQL
1-Forceps	Composition analysis	3.2 and GG-I-526	GG-I-526	**	*
	Hardness	3.3 and GG-I-526	GG-I-526	S-2	1.0
	Dimensions	3.4	---------	S-2	4.0
	Boil test	GG-I-526	GG-I-526	S-2	1.0
	Copper sulfate	GG-I-526	GG-I-526	S-2	1.0
	Distortion test	GG-I-526	GG-I-526	S-2	1.0
	Jaw and ratchet operation test	3.7	4.4.1	S-2	1.0

* Acceptance number shall be zero for all sample sizes.
** Sufficient amount from each lot of metal used in the construction of the forceps shall be selected for chemical analysis.

4.2.3 Sample unit usage. Where feasible, the sample unit shall be used for the determination of two or more test characteristics.

4.3 Examination.

4.3.1 Examination of forceps. The forceps shall be examined to determine compliance with all the requirements contained in this specification. Examination shall be conducted as specified in Table I and III. Examination however, shall not be restricted to defects listed in Table III. Any forceps in the sample containing one or more defects shall be rejected, and if the number of defective items in any sample exceeds the acceptance number for test sample, the lot represented by the sample shall be rejected.

Table III Classification of Defects

Categories	Defects
Critical	None defined.
Major	
101	Style and design not as specified.
102	Forceps not free from cracks, pits, dents, nicks, sharp edges, burrs, rough areas, grind marks, overlaps, crevices, scratches or blemishes.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:37Z
Check the source to verify that this is the current version before use.

Table III. Classification of Defects (cont'd)

Categories	Defects
103	Serrations not as specified.
104	Ratchet teeth not uniform or function poorly.
105	Finish not as specified.
106	Forceps not free from excessive play in box lock area.
107	Jaws do not mesh properly, or are not properly aligned.
Minor	
201	Identification marking not present, correct, complete or legible.

* 4.3.2 Examination for preparation for delivery. An examination shall be made to determine whether the packaging, packing, and marking comply with the requirements of Section 5. Defects shall be scored as specified in Table IV. Sampling shall be in accordance with MIL-STD-105. The sample unit shall be one container fully prepared for delivery. The lot shall be the number of containers offered for inspection at one time. The inspection level shall be S-2 and the acceptable quality level shall be 4.0 defects per hundred units.

Table IV. Examination for Preparation for Delivery

Examination	Defect
Containers	Not as specified.
Quantity	Quantity in interior packages and exterior containers not as specified.
Packaging and packing materials	Any component missing or damaged.
Workmanship	Inadequate application of components, such as incomplete closure of container flaps, tape improperly applied, or containers distorted.
Markings	Omitted, incorrect, illegible, improper location, sequence, or method of application.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:37Z
Check the source to verify that this is the current version before use.

4.4.1 Jaw and operation test. The forceps shall be positioned to the first ratchet position, but not locked. Forceps with jaw tips not touching shall be rejected. Forceps that do not progressively tighten with each ratchet advance shall be rejected.

5. PREPARATION FOR DELIVERY.

5.1 Packaging. Packaging shall be level A or C as specified (see 6.2).

5.1.1 Level A.

5.1.1.1 Unit package. Each forceps shall be thoroughly cleaned with the applicable method specified in MIL-P-116 and wrapped in neutral tissue or other suitable anticorrosive material. The forceps shall be further packaged in accordance with MIL-P-116, method IC-1 except that a preservative shall not be applied.

* 5.1.1.2 Intermediate package. Six unit packages, containing forceps of one type only, shall be packed in a snug fitting box constructed in accordance with PPP-B-566, PPP-B-636, class domestic or PPP-B-676. Closure of the box shall be adequate to prevent accidental opening.

5.1.2 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be Level A, B or C as specified (see 6.2).

5.2.1 Level A. Twenty four intermediate packages (144 units) shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant. Closure and strapping shall be as specified in the appendix of the applicable box specification.

* 5.2.1.1 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes. Each fiberboard box shall be waterproofed as specified in the appendix of PPP-B-636.

5.2.2 Level B. Twenty four intermediate packages (144 units) shall be packaged in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure shall be as specified in the appendix of the applicable box specification. Closure of fiberboard boxes shall conform to method II as specified in PPP-B-636.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:37Z
Check the source to verify that this is the current version before use.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Interior packages. Each unit package and each intermediate package shall be marked in accordance with MIL-STD-129. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Intended use. The hemostatic forceps are to be used for clamping large vessels of the pelvis and broad ligament during surgical procedures.

6.2 Ordering data. Procurement documents should specify the following:

(a) Number, title and date of this specification.
(b) Selection of applicable level of packaging and packing required (see 5.1 and 5.2).
(c) National Stock Number.

6.3 National Stock Number (NSN) coverage. This specification covers the following items listed in the Federal Supply Catalog:

National Stock Number	Item Identification	Type
6515-00-334-9500	FORCEPS, HEMOSTATIC, Curved, Pean, 9 inch	I
6515-00-334-9900	FORCEPS, HEMOSTATIC, Straight, Pean, 9 inch	II

It does not include all types or sizes as indicated by the title nor all types which are commercially available.

* 6.4 The margins of this specification are marked with an asterisk to indicate where changes (additions, modifications, corrections, deletions) from the previous issue were made. This was done as a convenience only and the Government assumes no liability whatsoever for any inaccuracies in these notations. Bidders and contractors are cautioned to evaluate the requirements of this document based on the entire content irrespective of the marginal notations and relationship to the previous issue.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:37Z
Check the source to verify that this is the current version before use.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6515-1958
CS

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:37Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:37Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers' vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE
MIL-F-36940A Forceps, Hemostatic, Pean, 9 Inch

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A
☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?
   A. GIVE PARAGRAPH NUMBER AND WORDING.

   B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?
   ☐ YES ☐ NO *(If "Yes", in what way?)*

4 REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426 REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:37Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, PA 19101

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304



HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:37Z
Check the source to verify that this is the current version before use.