MILITARY SPECIFICATION

INTRAVENOUS INJECTION KIT, PEDIATRIC, DISPOSABLE

This specification is approved for use by all the Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers a sterile, disposable intravenous injection kit for use in intravenous administration of fluid into a small vein.

1.2 Classification. Intravenous injection kits shall be of the following sizes as specified (see 6.2):

Size 2	21 gage needle
Size 3	19 gage needle

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposals, form a part of the specification to the extent specified herein:

SPECIFICATIONS

Federal

PPP-B-566	Boxes, Folding, Paperboard.
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.
PPP-B-676	Box, Setup.

Military

MIL-P-116	Preservation, Methods of.
MIL-L-10547	Liners, Case, and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.
MIL-S-36586	Sterilization Strip Test, Bacterial Spore.
MIL-I-36307	Intravenous Injection Set, Pediatric, Disposable Type.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:38Z
Check the source to verify that this is the current version before use.

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

(Copies of specifications, and standards, required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposals shall apply:

THE UNITED STATES PHARMACOPEIAL CONVENTION, INC.

The Pharmacopeia of the United States of America (U.S.P.)

(Application for copies shall be addressed to the Mack Publishing Company, Easton, Pennsylvania 18042.)

U. S. DEPARTMENT OF HEALTH, EDUCATION, AND WELFARE:
FOOD AND DRUG ADMINISTRATION

Federal Food, Drug, and Cosmetic Act and General Regulations for its Enforcement

(Application for copies should be addressed to Food and Drug Administration, U.S. Department of Health, Education, and Welfare, Washington, D.C. 20203.)

AMERICAN PHARMACEUTICAL ASSOCIATION PUBLICATION

National Formulary

(Application for copies should be addressed to the Mack Publishing Co., Easton, Pennsylvania 19082.)

3. REQUIREMENTS

3.1 Material. All materials as specified herein shall be of the best quality and suitable in all respects for the purpose intended.

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:38Z
Check the source to verify that this is the current version before use.

3.2 Style, design and construction. Shall be a disposable venipuncture kit consisting of a prepackaged rigid plastic tray, nominally 4-3/4 inches long, 2-5/8 inches wide and 7/8 inch deep. Shall contain the following sterile components:

Component	Size 2 Quantity	Size 3
Intravenous injection set, 19 gage needle		1
Intravenous injection set, 21 gage needle	1	
Tourniquet	1	1
Alcohol prep packet	1	1
Gauze pads	2	2
Roll of surgical tape	1	1

3.2.1 Intravenous injection set, 21 gage needle (Size 2). The material, style, design, dimensions and construction shall be in accordance with MIL-I-36307 except that packaging requirements of that specification shall not apply.

3.2.2 Intravenous injection set, 19 gage needle (Size 3). The material, style, design, dimensions and construction shall be in accordance with MIL-I-36307 except that packaging requirements of that specification shall not apply.

3.2.3 Tourniquet. Shall consist of molded gum latex rubber, 3/4 inch ± 1/32 inch wide by 18 inches ± 1/4 inch long and 0.033 inch ± 0.005 inch thick. The tourniquet shall have a textured surface on both sides and shall be held securely in place in the package by a U-shaped, high density, polyethylene, natural color clip.

3.2.4 Alcohol prep packet. Shall be a non-woven cotton pad, 1-1/8 inches ± 1/8 inch wide by 1-7/8 inches ± 1/8 inch long, saturated with approximately 1.1 ml of 70% Isopropyl Alcohol, N.F. Shall be contained in a sealed foil packet. The alcohol prep packet shall be labelled to identify it as "Alcohol Prep" containing "Isopropyl Alcohol 70% FOR EXTERNAL USE ONLY."

3.2.5 Gauze pad. The gauze pad shall be fabricated from materials specified in strict accordance with the requirements specified in Defense Medical Purchase Description covering Federal Stock Number 6510-559-3221, Sponge, Surgical Gauze, 2 by 2 Inches. The issue in effect during the invitation for bids, or request for proposals, shall apply concerning the above Defense Medical Purchase Description.

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:38Z
Check the source to verify that this is the currentversion before use.

3.2.6 Tape. Shall consist of a porous, translucent, surgical tape, 3/4 inch ± 1/32 inch wide by 24 inches ± 1 inch long, wound on a 1 inch minimum paper core.

3.2.7 Plastic tray. Dimensions of the tray shall be as specified in 3.2.

3.3 Sterility. All components of the intravenous injection kit shall be sterile when tested with a biological indicator as specified in 4.4.

3.3.1 Certificate of sterility. A certificate of sterility, stating that samples from each sterilizer lot have been tested and found to be sterile, shall be submitted to the procuring activity. The certificate of sterility shall include:

a. Item identification.
b. Federal stock number.
c. Contract or purchase order number.
d. Sterility lot or control number.
e. Date of sterilization.
f. Quantity of product in sterilizer lot.
g. Type of sterilization (gas, radiation, heat, steam, etc.).
h. Form of biological indicator (i.e., inoculated paper spore strip or carrier, inoculated product, or inoculated simulated product).
i. Test organism(s) and population density used.
j. Test medium, incubation temperature, incubation time, and dates of tests.
k. Number of test units.
l. Statement that samples comply with sterility tests.
m. Signature of head of laboratory or a designated qualified alternate.

When inspection is made at destination, the certificate of sterility shall be submitted by the contractor to the Directorate of Medical Materiel, ATTN: DPSC-ATQ, Defense Personnel Support Center, 2800 South 20th Street, Philadelphia, Pa., 19101, and a copy forwarded with each shipment to the consignee. When inspection is made at source, certificate of sterility shall be furnished to the cognizant Government Quality Assurance Representative for submission to the procuring activity.

3.3.2 Sterility lot. A sterility lot is that single quantity of product subjected to the same manufacturing operation and simultaneously sterilized in the same sterilizer chamber. Each unit package shall be clearly marked with a sterility lot (control) number and date of sterilization as specified in 5.3.1. Each sterility lot shall be traceable to the pertinent manufacturing lot(s).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:38Z
Check the source to verify that this is the current version before use.

3.4 Safety (toxicity). All equipment shall be safe. When tested as specified in 4.4.2, the kits shall comply with requirements for safety (toxicity) of the U.S.P.

3.5 Freedom from pyrogens. The fluid path and end connectors shall be pyrogen-free when tested as specified in 4.4.3.

3.6 Certificates of freedom from pyrogen safety and biological tests. A certificate listing all lot or control numbers shall state that representative samples from each lot or control have been tested and found to be free from pyrogens, nontoxic (safe), and have passed the biological tests. This certificate shall include the specific tests performed and the test results. For destination inspection, certificates shall be submitted by the contractor directly to Defense Personnel Support Center, Directorate of Medical Materiel, Technical Operations Division, 2800 South 20th Street, Philadelphia, Pa. 19101, Attention: Quality Assurance Branch, and one copy of each shall be forwarded with each shipment to the consignee. For source inspection, certificates shall be furnished to the procuring activity via the government inspector.

3.7 Manufacturer's record. The manufacturer shall place a code or lot number on the paper label, or on the printed carton, which can be verified against the manufacturer's records of all tests made on nontoxic, sterile, pyrogen free and biological tests performed on the kits. These records shall be made available for inspection by any duly authorized Government representative. No shipment shall be made until all tests have been completed, verified, and recorded in the permanent records of the manufacturer.

3.8 Workmanship. Workmanship shall be first class throughout. Kits shall be free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:38Z
Check the source to verify that this is the current version before use.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2 Certificates of quality. Certificates of quality, supplied by the manufacturer of the component or material, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the FSN, the item identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as indicated in Table I.

Table I. Sampling for examination

Defect category	Inspection level	AQL (percent defective)
Major A	S-2	1.0
Major B	S-2	1.0
Minor	S-2	1.5

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as indicated in Table II.

Table II. Sampling Criteria for tests

Component	Characteristic	Requirement	Test procedure	Inspection level	AQL (percent defective)
Kit	Sterility	3.3	4.4.1	**	*
	Safety	3.4	4.4.2	**	*
	Freedom from pyrogens	3.5	4.4.3	**	*

* Acceptance number shall be zero for all sample sizes.
** Sampling for tests shall be on component lots.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:38Z
Check the source to verify that this is the current version before use.

4.2.3 Sample unit usage. Where possible, the same sample used in the construction of the kit shall be used for the determination of two or more test characteristics.

4.3 Examination. The kit shall be examined to determine compliance with all requirements contained in this specification. Examination shall be conducted in accordance with MIL-STD-105 and the classification of defects specified in Table III. Examination shall not be restricted to the classified possible defects listed in Table III. Any intravenous injection kit in the sample containing one or more defects shall be rejected and, if the number of defective items in any sample exceeds the acceptance number for that sample, the lot represented by that sample shall be rejected.

Table III. Classification of defects

Categories	Defects
Critical	None defined.
Major A	
101	Kit incomplete, one or more components missing.
Major B	
151	Metal or plastic parts not smoothly finished; not free of rough edges, slivers, nicks, pits, blisters, scratches, cracks, bubbles, roughness, or corrosion, as applicable.
152	Style and design of components not as specified.
Minor	
201	Packaging, packing and marking not as specified.
202	Poor workmanship (not resulting in a major defect).

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Sterility test. The intravenous injection kit shall be tested for sterility in accordance with the method specified in the USP for all components of the kit using inoculated product, inoculated simulated product, or paper spore strip (carrier). Product containing a biological indicator shall be marked or labeled so that it clearly differentiates from product intended for distribution. The incubation medium and minimum time in each case shall be soybean casein digest for 7 days for inoculated product, inoculated simulated product, or paper spore strip (carrier). The incubation temperature shall be as indicated in paragraph 4.4.1.1. The minimum number of test units, regardless of lot size, shall be 10, except for inoculated carrier which shall be 20 for Type I Sterilization and 30 for Type II Sterilization including inoculated carrier and product test units. The inspection level shall be as indicated in Table II, Sampling Criteria for Tests.

4.4.1.1 Inoculated product or simulated product for each type of sterilization, the organism used shall be as listed below:

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:38Z
Check the source to verify that this is the current version before use.

Type of Sterilization	Organism	Incubation Temperature
Steam	B. stearothermophilus	55° - 65°C
Dry Heat	B. stearothermophilus	55° - 65°C
	B. subtilis var, niger (or globigii)	30° - 35°C
Gas	B. subtilis var. niger (or globigii)	30° - 35°C
Radiation	B. pumilus	30° - 35°C

The number of viable cells shall be sufficient to provide assurance that the product has been sterilized.

4.4.1.2 Inoculated carrier. Inoculated carrier (paper spore strips) if used shall be in accordance with MIL-S-36586. Other paper carriers shall be inoculated to be of equivalent resistance. The method of incubation and number of test units shall be as indicated in 4.4.1 Sterility test and 4.4.1.1 Inoculated product or simulated product. The method of sterilization will control the organism in the spore strip to be tested for as shown in 4.4.1.1. NOTE: When an inoculated carrier is chosen, product samples must also be tested as required by the USP.

4.4.1.3 Alternate method. Suppliers conducting sterility tests or designating lots in a manner other than that indicated herein shall submit proof that the alternate method provides equal assurance concerning the sterility of the product. However, all sterility tests shall be conducted using biological indicators. Requests for approval of any alternate method of testing for sterility shall be made to the Contracting Officer, and approval obtained before employment of the alternate method on a government contract.

4.4.1.4 Product testing. Testing for sterility of the product after receipt by the Government shall be without a biological indicator and in accordance with the U.S.P.

4.4.2 Safety. The kits shall be tested for safety in accordance with the U.S.P.

4.4.3 Freedom from pyrogens. Samples of the fluid path and end connectors shall be tested for freedom from pyrogens, in accordance with the U.S.P.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One package containing ten kits, as specified, constitutes one unit of issue.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:38Z
Check the source to verify that this is the current version before use.

5.1.2 Packaging quantities. The number of units of issue indicated in Table IV shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table IV. Packaging Quantities

Unit package	Intermediate package	Exterior container
1 unit	Not applicable	6 units

5.1.2.1 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units may be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Kit. Components comprising each kit shall be packaged in a plastic tray as specified in 3.2. The tray shall be provided with a sealed tamperproof, peel-open type plastic cover. The sealed tray shall be capable of maintaining sterility of contents unless damaged or opened.

5.1.3.2 Unit package. Ten kits shall be packaged in a box of appropriate size constructed in accordance with PPP-B-566 or PPP-B-676 except that commercial colors are acceptable. Closure shall be adequate to prevent spilling of contents under normal handling.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant. Fiberboard boxes shall conform to the special requirements of Table II in PPP-B-636.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:38Z
Check the source to verify that this is the current version before use.

5.2.1.2 Case liner. Each level A wood box shall be lined with waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes.

5.2.1.3 Closure. Closure of wood boxes shall be in accordance with appendix of applicable box specification. Closure of each fiberboard box shall be as specified in PPP-B-636. In addition, waterproofing shall be in accordance with 30.4.

5.2.1.4 Strapping. Strapping, when required, shall be in accordance with appendix of applicable box specification.

NOTE: Strapping shall not be required for shipments forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure shall be as specified in the appendix of the applicable box specification. Fiberboard box shall be closed by method II of PPP-B-636. Fiberboard boxes shall conform to special requirements of Table I PPP-B-636.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Kit. In addition to any commercial markings, each kit (tray) shall be marked in accordance with the requirements of the Federal Food, Drug, and Cosmetic Act. Marking shall also include the lot (control) number, contents, name and address of the contractor or registered trademark, the legends that the kit is disposable, shall not be re-used or re-sterilized and that sterility is guaranteed unless kit is damaged or opened or similar statements.

5.3.2 Unit packages. Each unit package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Lot (control) number, contract or purchase order number and name and address of manufacturer shall be shown. The method of sterilization (G-Gas, R-Radiation, C-Cold, S-Steam, D-Dry Heat Sterilization) shall be shown following the level of packaging and the date of packaging. Example: "A-3/73-G". Final sterilization shall be on or after indicated packaging date.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:38Z
Check the source to verify that this is the current version before use.

5.3.3 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129. Lot (control) number shall be shown. The method of sterilization as specified in 5.3.2 shall also be shown.

6. NOTES

6.1 Intended use. The intravenous injection kits are intended for use in intravenous administration of fluid into a small vein.

6.2 Ordering data. Purchasers should exercise any desired options offered herein, and the procurement documents should specify the following:

(a) Title, number, and date of this specification.
(b) Size (see 1.2).
(c) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.3 This specification is intended to cover the following items appearing in the Federal Supply Catalog:

Federal Stock No.	Item Identification
6515-083-6529	INTRAVENOUS INJECTION KIT, 19 Gage Needle, Pediatric, Disposable 10s.
6515-083-6533	INTRAVENOUS INJECTION KIT, 21 Gage Needle, Pediatric, Disposable, 10s.

It does not include all types or sizes as indicated by the title nor all types which are commercially available.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6515-1704

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:38Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:38Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

[ ] YES [ ] NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426 REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

S/N 0102-014-1802

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:38Z
Check the source to verify that this is the currentversion before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:38Z
Check the source to verify that this is the current version before use.