A-A-0053059 (DM)
17 August 1987

INTERIM COMMERCIAL ITEM DESCRIPTION

KITS, THREAD-FORMING MOUNTED PINS, DENTAL

The General Services Administration has authorized the use of this Interim Commercial Item Description which was developed by DLA, Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA 19101, based upon currently available technical information.

This interim commercial item description covers complete dental thread forming pin kits (consisting of mounted pins, drills, and drivers), replacement kits (consisting of replacement mounted pins), and a replacement driver. The kits shall be suitable for use in the thread-forming pin technique. The kits shall be of the following classes, styles, and sizes as specified below. The driver shall be class 3 as specified below.

Class 1 - Complete Kit (pins, drill and driver)

Style A - Single pin
Size 0.015, 0.019 0.024, 0.030 inch diameter

Style B - Single pin, shoulder-stop
Size 0.024, 0.030 inch diameter

Style C - Double pin
Size 0.024, 0.030 inch diameter

Style D - Double pin, shoulder-stop
Size 0.024, 0.030 inch diameter

Class 2 - Replacement Kit (pins)

Style A - Single Pin
Size 0.015, 0.019, 0.024, 0.030 inch diameter

Style B - Single Pin, shoulder-stop
Size 0.024, 0.030 inch diameter

Style C - Double Pin
Size 0.024, 0.030 inch diameter

Style D - Double Pin, shoulder-stop
Size 0.24, 0.030 inch diameter

Class 3 - Driver, self-threading pin

\MSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:14Z
Check the source to verify that this is the current version before use.

Salient Characteristics:

The complete kits, replacement kits and replacement driver shall be Whaledent International, or equal, for design and performance. The applicable Whaledent Item part numbers for the applicable National Stock Number are as listed herein (see notes).

Class 1 - Complete Kit - The complete kit shall consist of 20 single, style A or B, pins, or 20 double (40 single), style C or D, pins, one drill (style A and B) or two drills (style C and D), and one hand driver, as specified for the applicable NSN.

Class 2 - Replacement Kit - The replacement kit shall consist of 50 single, Style A or B pins, or 50 double (100 single), style C or D, pins, as specified for the applicable NSN.

Class 3 - Driver - The driver shall be a replacement hand driver.

Style A - Style A pins shall be Whaledent "TMS Link" system Self-Threading pins, or equal. Shall be a single pin mounted in a latch type contra-angle disposable mandrel such that the pin may be installed to full depth and then shear from the mandrel. The diameter and length of pin, drill, and hole size shall be as specified in Table I and II for the applicable class, style and size pin.

Style B - Style B pins shall be Whaledent "Link Plus" self-threading retention pins, or equal. Shall be a single pin mounted in a latch-type contra-angle disposable mandrel such that the pin may be installed to full depth and then shear from the mandrel. Shall have a shoulder-stop, 0.010 inch long, for positively controlling depth of insertion. The diameter and length of pin, drill, and hole size shall be as specified in Table I and II for the applicable class, style and size pin.

Style C - Style C pins shall be Whaledent "TMS Links Series" Self-Threading pins, or equal. Shall be two joined pins separated by a shear groove with one end mounted to a latch type contra-angle disposable mandrel such that one pin may be installed to full depth, shear, and then permit the remaining pin to be similarly installed. The diameter and length of pin, drill, and hole size shall be as specified in Table I and II for the applicable class, style and size pin.

Style D - Style D pins shall be Whaledent "Link Plus" self-threading retention pins, or equal. Shall be two joined pins separated by a shear groove with one end mounted to a latch type contra-angle disposable mandrel such that one pin may be installed to full depth, shear, and then permit the remaining pin to be similarly installed. Shall feature a shoulder-stop for positively controlling the depth of insertion. The diameter and length of pin, drill, and hole size shall be as specified in Table I and II for the applicable class, style and size pin.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:14Z
Check the source to verify that this is the current version before use.

All pins shall have the following characteristics:

The pins shall be self-threading, self-shearing dental pins mounted in a latch type contra-angle mandrel and shall be suitable for the retention of dental restorative materials. The pins shall be constructed of AISI 302 austenitic corrosion-resistant steel, flash plated with elemental gold so as to be capable of amalgamation bonding and insure biocompatibility. The pins shall be flexibly coupled (linked) to a disposable plastic mandrel so as to allow the pin to self align within the predrilled hole upon insertion. The plastic mandrel shall be fabricated of DuPont "Delrin" or equal. The mandrel shall be color coded such that the color identifies the size of pin. The colors shall be as specified in Table II. The length of the mandrels shall be 0.625 ± .020 inch. The pins shall be capable of advancing to full depth and locking within the smooth walled dentin drilled hole. The threads shall be right handed and pitched so as to advance the pin 2 millimeters (mm) in no more than 9 revolutions. All mounted pins shall be suitable for use with class 3 driver. The mandrels shall be latch type and designed such that the mandrel shall snap into the driver and be gently retained by the driver. The mandrel shall fit and interface with standard latch type handpieces.

Class 3 driver shall be suitable for the insertion of self-threading mounted dental pins. The driver shall be Whaledent "TMS Link and Link Plus Series" hand driver, or equal. The color of the driver shall be white and shall be fabricated of E. DuPont "Celcon", or equal. The driver shall be capable of easily engaging the latch end of the pins mandrel so as to permit manual insertion of the pins. The mounted pin shall be positively and securely retained by the driver. The driver shall be capable of being easily and quickly disengaged. The driver shall be 0.500 ± .015 inch in length, 0.187 ± .010 inch in diameter, and shall permit access to posterior teeth. The driver shall have a knurled, scored, grooved or serrated handle to facilitate finger grip.

The drills shall be as specified in Table II. The drills shall be latch angle type and meet the applicable requirements of MIL-D-36786B for the applicable size as specified in Table II. Drill shanks shall be color coded to match the applicable size pins specified herein.

Tolerances. Unless otherwise specified, tolerances on dimensions shall be within plus or minus ten percent of specified value.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:14Z
Check the source to verify that this is the current version before use.

TABLE I

Class	Style	Size (Diameter of Pin, inch)	Length of Pin (Single pin, inch)	Whaledent Kit Number (or Equal)	Components of Kit
1	A	0.015	0.083	L-501	20 Single Mounted Pins, 1 Drill, 1 Driver
1	A	0.019	0.120	L-511	20 Single Mounted Pins, 1 Drill, 1 Driver
1	A	0.024	0.200	L-521	20 Single Mounted Pins, 1 Drill, 1 Driver
1	A	0.030	0.200	L-541	20 Single Mounted Pins, 1 Drill, 1 Driver
1	B	0.024	0.200	L-721	20 Single Mounted Pins, 1 Drill, 1 Driver
1	B	0.030	0.200	L-741	20 Single Mounted Pins, 1 Drill, 1 Driver
1	C	0.024	0.177	L-531	40 Mounted (20 Double) Pins, 2 Drills, 1 Driver
1	C	0.030	0.177	L-551	40 Mounted (20 Double) Pins, 2 Drills, 1 Driver
1	D	0.024	0.197	L-731	40 Mounted (20 Double) Pins, 2 Drills, 1 Driver
1	D	0.030	0.197	L-751	40 Mounted (20 Double) Pins, 2 Drills, 1 Driver

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:14Z
Check the source to verify that this is the current version before use.

TABLE 1 Continued.

Class	Style	Size (Diameter of Pin, inch)	Length of Pin (Single pin, inch)	Whaledent Kit Number (or Equal)	Components of Kit
2	A	0.015	0.083	L-502	50 Single Mounted Pins
2	A	0.019	0.120	L-512	50 Single Mounted Pins
2	A	0.024	0.200	L-522	50 Single Mounted Pins
2	A	0.030	0.200	L-542	50 Single Mounted Pins
2	B	0.024	0.200	L-722	50 Single Mounted Pins
2	B	0.030	0.200	L-742	50 Single Mounted Pins
2	C	0.024	0.177	L-532	100 Mounted (50 Double) Pins
2	C	0.030	0.177	L-552	100 Mounted (50 Double) Pins
2	D	0.024	0.197	L-732	100 Mounted (50 Double) Pins
2	D	0.030	0.197	L-752	100 Mounted (50 Double) Pins

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:14Z
Check the source to verify that this is the current version before use.

TABLE II

PIN AND DRILL DIMENSIONS

Whaledent Kit Number (or Equal)	Whaledent Pin Designation and mandrel color	Whaledent Drill Number (Ref only)*	Drill Hole Diameter (Inch)*	Drill Hole Length (Inch)*
L-501	Minuta (Pink)	K-89	0.0135	0.051
L-511	Minikin (Red)	K-91	0.017	0.059
L-521	Minim (Silver)	K-93	0.021	0.157
L-541	Regular (Gold)	K-97	0.027	0.197
L-721	Minim (Silver)	K-92	0.021	0.079
L-741	Regular (Gold)	K-96	0.027	0.079
L-531	Minim (Silver)	K-92	0.021	0.079
L-551	Regular (Gold)	K-96	0.027	0.079
L-731	Minim (Silver)	K-92	0.021	0.079
L-751	Regular (Gold)	K-96	0.027	0.079
L-502	Minuta (Pink)	K-89	0.0135	0.051
L-512	Minikin (Red)	K-91	0.017	0.059
L-522	Minim (Silver)	K-93	0.021	0.157

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:14Z
Check the source to verify that this is the current version before use.

TABLE II Continued.

PIN AND DRILL DIMENSIONS

Whaledent Kit Number (or Equal)	Whaledent Pin Designation and mandrel color	Whaledent Drill Number (Ref only)*	Drill Hole Diameter (Inch)*	Drill Hole Length (Inch)*
L-542	Regular (Gold)	K-97	0.027	0.197
L-722	Minim (Silver)	K-92	0.021	0.079
L-742	Regular (Gold)	K-96	0.027	0.079
L-532	Minim (Silver)	K-92	0.021	0.079
L-552	Regular (Gold)	K-96	0.027	0.079
L-732	Minim (Silver)	K-92	0.021	0.079
L-752	Regular (Gold)	K-96	0.021	0.079

*Drill size shall be in accordance with DPSC Drawing Numbers 22160 and 23318 (Class AL).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:14Z
Check the source to verify that this is the current version before use.

Workmanship. The Dental Thread-Forming Pin Kit shall be free from defects which detract from their appearance or may impair their serviceability, proper functioning, or intended use.

Unit. For Class 1, Style A and Class 1, Style B- Each (EA). One kit containing 20 pins of one class and style only, one drill and one driver, as specified, constitutes one unit.

For Class 1, Style C and Class 1, Style D- Each (EA). One kit containing 40 pins of one class and style only, two drills and one driver, as specified, constitutes one unit.

For Class 2, Style A and Class 2, Style B- Package (PG). One package containing 50 pins of one class and style only, as specified, constitutes one unit.

For Class 2, Style C and Class 2, Style D- Package (PG). One package containing 100 pins of one class and style only, as specified, constitutes one unit.

For Class 3- each (EA). One driver, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:14Z
Check the source to verify that this is the current version before use.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection purposes shall be one kit. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-1, AQL 1.0.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric Products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:14Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing, shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

N S N	Item Identification	Whaledent Part No.	Specification Designation
6520-01-240-5801	DRIVER, SELF-THREADING PIN, DENTAL, 0.75 inch (19.0 mm) Long, for Latch Type Mandrels	L-590	Class 3
6520-01-240-5829	PIN, THREAD-FORMING, DENTAL, Mounted, Self-Threading, Self-Shearing, Shoulder Stop, 0.024 inch (0.6 mm) Diameter by 0.197 inch (5.0 mm) Long, 50s.	L-722	Class 2, Style B
6520-01-240-8723	PIN, THREAD-FORMING, DENTAL, Mounted, Self-Threading, Self-Shearing, Shoulder Stop, 0.030 inch (0.77 mm) Diameter by 0.197 inch (5.0 mm) Long, 50s	L-742	Class 2, Style B
6520-01-240-8724	PIN, THREAD-FORMING, DENTAL, Mounted, Self-Threading, Self-Shearing, 0.015 inch (0.38 mm) Diameter by 0.085 inch (2.16 mm) Long, 50s	L-502	Class 2, Style A
6520-01-240-8725	PIN, THREAD-FORMING, DENTAL, Mounted, Self-Threading, Self-Shearing, 0.019 inch (0.5 mm) Diameter by 0.0120 inch (3.0 mm) Long, 50s.	L-512	Class 2, Style A

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:14Z
Check the source to verify that this is the current version before use.

NSN	Item Identification	Whaledent Part No.	Specification Designation
6520-01-240-8726	PIN, THREAD-FORMING, DENTAL, Mounted, Self-Threading, Self-Shearing, 0.024 inch (0.6 mm) Diameter, 0.200 inch (5.0 mm) Long, 50s.	L-522	Class 2, Style A
6520-01-241-2396	PIN KIT, THREAD-FORMING, DENTAL, 40 Mounted Pins, Self-Threading, Self-Shearing, 0.030 inch (0.76 mm) Diameter by 0.197 inch (5.0 mm) Long, 2 in 1, Shoulder-Stop, 1 Driver, 1 Drill.	L-751	Class 1, Style D
6520-01-241-2403	PIN KIT, THREAD-FORMING, DENTAL, 20 Mounted Pins, Self-Threading, Self-Shearing, 0.015 inch (0.38 mm) Diameter by 0.085 inch (2.16 mm) Long, 1 Driver, 1 Drill.	L-501	Class 1, Style A
6520-01-241-2427	PIN KIT, THREAD-FORMING, DENTAL, 40 Mounted Pins, Self-Threading, Self-Shearing, 0.024 inch (0.58 mm) Diameter by 0.197 inch (5.0 mm) Long, 2 in 1, Shoulder Stop, 1 Driver, 1 Drill.	L-731	Class 1, Style D
6520-01-241-2428	PIN KIT, THREAD-FORMING, DENTAL, 20 Mounted Pins, Self-Threading, Self-Shearing, 0.030 inch (0.77 mm) Diameter by 0.197 inch (5.0 mm) Long, Shoulder Stop, 1 Driver, 1 Drill	L-741	Class 1, Style B

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:14Z
Check the source to verify that this is the current version before use.

N S N	Item Identification	Whaledent Part No.	Specification Designation
6520-01-241-2429	PIN KIT, THREAD-FORMING, DENTAL, 20 Mounted Pins, Self-Threading, Self-Shearing, 0.24 inch (0.59 mm) Diameter by 0.200 inch (5.0 mm) Long, 1 Driver, 1 Drill.	L-521	Class 1, Style A
6520-01-241-4984	PIN, THREAD-FORMING, DENTAL, Mounted, Self-Threading, Self-Shearing, Shoulder Stop, 2 in 1, 0.024 inch (0.6 mm) Diameter by 0.197 inch (5.0 mm) Long, 100s.	L-732	Class 2, Style D
6520-01-241-7017	PIN KIT, THREAD-FORMING, DENTAL, 20 Mounted Pins, Self-Threading, Self-Shearing, 0.024 inch (0.58 mm) Diameter by 0.197 inch (5.0 mm) Long, Shoulder Stop, 1 Driver, 1 Drill.	L-721	Class 1, Style B
6520-01-243-3685	PIN KIT, THREAD-FORMING. DENTAL, 40 Mounted Pins, Self-Threading, Self-Shearing, 0.030 inch (0.76 mm) Diameter by 0.177 inch (4.49 mm) Long, 1 Driver, 1 Drill.	L-551	Class 2, Style C
6520-01-244-2779	PIN, THREAD-FORMING, DENTAL, Mounted, Self-Threading, Self-Shearing, Shoulder Stop, 2 in 1, 0.030 inch (0.77 mm) Diameter by 0.197 inch, (5.0 mm) 100s	L-752	Class 2, Style D
6520-01-249-2160	PIN, THREAD-FORMING, DENTAL, Mounted, Self-Threading, Self-Shearing, 2 in 1, 0.024 inch (0.6 mm) Diameter by 0.177 inch (4.5 mm) Long, 100s.	L-532	Class 2, Style C

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:14Z
Check the source to verify that this is the current version before use.

NSN	Item Identification	Whaledent Part No.	Specification Designation
6520-01-249-2161	PIN, THREAD-FORMING, DENTAL, Mounted, Self-Threading, Self-Shearing, 0.030 inch (0.76 mm) Diameter by 0.200 inch (5.0 mm) Long, 50s.	L-542	Class 2, Style A
6520-01-249-2162	PIN, THREAD-FORMING, DENTAL, Mounted, Self-Threading, Self-Shearing, 0.030 inch (0.76 mm) Diameter by 0.177 inch (4.5 mm) Long, 2 in 1, 100s.	L-552	Class 2, Style C
6520-01-249-6380	PIN KIT, THREAD-FORMING, DENTAL, 20 Mounted Pins, Self-Threading, Self-Shearing, 0.019 inch (0.48 mm) Diameter by 0.120 inch (3.0 mm) Long, 1 Driver 1 Drill.	L-511	Class 1, Style A
6520-01-249-9783	PIN KIT, THREAD-FORMING, DENTAL, 20 Mounted Pins, Self-Threading, Self-Shearing, 0.030 inch (0.76 mm) Diameter by 0.200 inch (5.0 mm) Long, 1 Driver, 1 Drill.	L-541	Class 1, Style B
6520-01-156-3155	PIN KIT, THREAD-FORMING, DENTAL, 40 Mounted Pins, Self-Threading, Self-Shearing, 0.024 inch (0.61 mm) Diameter by 0.177 inch (4.49 mm) Long, 2 in 1, 1 Driver, 1 Drill.	L-531	Class 1, Style C

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:14Z
Check the source to verify that this is the current version before use.

A-A-0053059 (DM)

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

AGENT:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

61013

Project No. 6520-2023

Location: NCR Disk #4

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:14Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in ubmitting comments and suggestions for improvements All users of military standardization documents are invited to provide suggestions This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and ed In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was igid, restrictive loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems Enter in block 6 any remarks not related to a specific paragraph of the document If block 7 is filled out an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-ATTS
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:14Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-0053059(DM)	KITS, THREAD-FORMING MOUNTED PINS, DENTAL

3a NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b ADDRESS *(Street City State ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)*

5 PROBLEM AREAS

a Paragraph Number and Wording

b Recommended Wording

c Reason/Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER *(Last First, MI)* – Optional	b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c MAILING ADDRESS *(Street City State, ZIP Code)* – Optional	8 DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426

PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:14Z
Check the source to verify that this is the current version before use.