A-A-53796
27 December 1988

COMMERCIAL ITEM DESCRIPTION

SYRINGE, HYPODERMIC Luer-Lock, Disposable

The General Services Administration has authorized the use of this commercial item description in lieu of Military Specification MIL-S-36157E for the given syringes.

This commercial item description covers sterile, disposable, Luer-Lock, hypodermic syringe suitable for a single injection.

Salient characteristics:

General. Syringes shall be of the following sizes:

Size	Syringe Volume (ml)	Barrel Length (inch)	Main Graduations (ml)	Sub Graduations (ml)	Number of syringes per unit package
1	3	2-1/2-3-1/4	0.5	0.1	100
2	20	4-1/4-4-1/2	5	1	40
3	30	4-1/2-4-7/8	5	1	25

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:24Z
Check the source to verify that this is the current version before use.

Material. Syringe barrel and plunger shall be rigid polypropylene which meets the USP requirements for Class III plastics. The interior of the barrel shall be lightly coated with an inert silicone lubricant. The plunger tip shall be inert natural rubber or neoprene which meets the USP requirements for Elastomeric Closures for Injections.

Style and design. Syringe barrel tip shall be in accordance with ANSI/HIMA MD70.1. Syringe barrel shall be flanged to provide finger grip and anti-roll feature. Inside of the barrel shall have a smooth, uniform surface, free from pits, grind marks, bubbles, flash, or other imperfections. Plunger shall glide through the barrel while maintaining fluid-tight fit. The plunger tip and barrel shall both be beveled and form a flush fit at zero volume. Barrel shall be transparent and the plunger tip shall have two visible rings touching the sides of the barrel.

The outside surface of the barrel shall have each 0.5 ml marked with a main graduation marking and with the numerical volume. Each 0.1 ml shall be subgraduated. Main graduation markings shall be at least 33 percent longer than subgraduation markings. Barrel shall also have the name or registered trademark of the contractor. Actual fluid volume shall be determined by position of the distal ring of the plunger tip.

Accuracy test. Syringe shall deliver $\pm$ 5 percent of the capacity indicated at each graduation mark when tested against a calibrated burette.

Leakage test. The water pressure test system shall have a water pressure gage and a rubber plug. The needle shall be secured to the Luer tip of the syringe by applying it with a 5 pound axial force and while maintaining this compressive stress twisting the needle hub 45 degrees relative to the Luer tip. This method of attachment does not apply to syringes with permanently affixed needles or syringes with Luer locks. The needle of the hypodermic shall be inserted into the plug and the water pressure brought up to 45 psig. The plunger shall then be withdrawn to the 50 percent point of the graduated capacity where it shall be allowed to remain for 10 seconds. The plunger shall then be withdrawn to the 75 percent point and held there for another 10 seconds. The plunger shall then be withdrawn to the 100 percent for a final 10 seconds.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:24Z
Check the source to verify that this is the current version before use.

During each 10-second period there shall not be enough water to pass the connection between the needle hub and the Luer tip of the syringe or needle-syringe junction and between the plunger tip and barrel of the syringe to produce a falling drop. After each successive test, any water that may have passed shall be wiped with a dry cloth and the next test applied. The needle of the hypodermic shall be withdrawn from the rubber plug and the water ejected at the rate of 1/2 cc/seconds until the plunger reaches the zero mark. The needle shall not leak or separate at its junction with the syringe.

Sterility, pyrogenicity, toxicity. The syringe shall be nonpyrogenic and nontoxic in accordance with the U.S.P. The syringe and needle shall be sterile in accordance with U.S.P. or A.A.M.I. guidelines.

Instructions. Each unit package shall be provided with instructions describing the proper techniques for dispensing the sterile syringes and needles, as applicable, from the containers and preparing them for use. In addition, information as to methods for destroying syringes and needles, as applicable, to prevent their re-use shall also be shown.

Workmanship. The syringe shall be free from defects which detract from its appearance or impair its servicability.

Unit. Package (PG). One package containing twenty-five, forty, or one hundred sterile syringes of one type and size, with instructions, as specified, constitutes one unit. The individual package shall be a sealed, peel-open type container capable of maintaining sterility of contents unless package is damaged or opened.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:24Z
Check the source to verify that this is the currentversion before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be inspection level S-1 and AQL 1.0 in accordance with MIL-STD-105. The unit of product for tests purposes shall be one syringe.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:24Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION	Size
6515-00-462-7348	SYRINGE, HYPODERMIC, Disposable, 3 ml, Luer-Lock, 100s	1
6515-01-273-1803	SYRINGE, HYPODERMIC, Disposable, 20 ml, Luer-Lock, 40s	2
6515-01-165-6742	SYRINGE, HYPODERMIC, Disposable, 30 ml, Luer-Lock, 25s	3

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:24Z
Check the source to verify that this is the currentversion before use.

Ordering data (Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4411

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:24Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-53796	SYRINGE, HYPODERMIC Luer-Lock, Disposable

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)*: _____

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://acsel.nla.mil – Downloaded: 2026-02-28 08:42
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4966 Alexandria, VA

POSTAGE WILL BE PAID BY

Commander
Defense Personnel Support Center
ATTN: DPSC-RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:24Z
Check the source to verify that this is the currentversion before use.