A-A-51207
12 December 1985

COMMERCIAL ITEM DESCRIPTION

WEDGE AND WEDGE SET, DENTAL (WOOD)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers three classes of wooden wedges suitable for interproximal wedging of teeth.

Class 1,2,3

Salient characteristics:

Material. The wedges shall be constructed of sycamore wood, which shall resist wetting, splintering, and breaking.

Description. The dental wedges shall be suitable for interproximal wedging of teeth and shall provide for a contouring contact in dental restorative procedures. The wedges shall be of such design as to provide support for holding bands tightly against the tooth. The wedges shall be anatomically curved with concave sides, rounded tip, and a square (non-operative) end to provide ease of placement. The design of the wedges shall conform to Figure 1. Wedges shall have a smooth finish. The wedges shall conform to the requirements specified in Table I for the class specified. The average height of the wedges at blunt end shall be 1.9 to 2.1 millimeters.


TABLE I.


Class	Assortment and Dimensions				Quantity	Total Quantity
	Color	Size	Width (mm), Blunt End (Average, ± 0.1)	Length (mm) (± 0.3)		
1	Green	13 T	1.7	13	200	400
	Yellow	13	2.0	13	200	
2	Orange	13 UT	1.4	13	100	
	Green	13 T	1.7	13	100	400
	Yellow	13	2.0	13	100	
	Blue	15	2.0	15	100	
3	Green	13 T	1.7	13	400	400


AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.


Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:52Z
Check the source to verify that this is the current version before use.







VIEW B-B



VIEW A-A

FIG. 1 WEDGE, DENTAL, WOOD

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:52Z
Check the source to verify that this is the current version before use.

Workmanship. Wedges shall be free of splinters, cracks, and other defects that would injure soft gum tissue.

Unit.

Class 1 and Class 2 - Set (SE). One wedge set, as specified, constitutes one unit.

Class 3 - Package (PG). One package containing 400 wedges, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government reserves reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection purposes shall be one wedge set or package. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-1, AQL 1.0.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:52Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and Local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material lin accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

		Designation
6520-01-053-5601	WEDGE SET, DENTAL, WOOD, SET of 400	Class 1
6520-01-057-1266	WEDGE SET, DENTAL, WOOD, SET of 400	Class 2
6520-01-132-7948	WEDGE, DENTAL, WOOD, 400s	Class 3

MILITARY INTEREST:
Custodians:
Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:
DoD-MB

Agent:
DLA-DM

CIVILIAN AGENCY COORDINATING ACTIVITY:
VA-OSS

05919

Project No. 6520-1857

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:52Z
Check the source to verify that this is the current version before use.