A-A-50902A
29 July 1988
SUPERSEDING
A-A-50902
21 March 1985

COMMERCIAL ITEM DESCRIPTION

HANDLE, ELECTRODE, ELECTROSURGERY

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a sterile, disposable, electrode handle with integral cord suitable for use with electrosurgical units.

Salient characteristics:

The handle shall be a sterile, disposable electrode handle. The electrosurgical handle shall be fabricated of a suitable plastic material and shall be not less than 6 inches in length. The design and construction of the handle shall be such as to render it slip-resistant and permit easy gripping and handling during surgical procedures. The handle shall be fitted with an internal chuck to receive and hold all standard 3/32 inch (shank diameter) electrosurgical electrodes. The cord shall have an approximate length of 10 feet and shall be fabricated of multistrand copper wire at least 24 AWG covered with a suitable insulation. It shall be neither so stiff that it impedes manipulation, nor so flexible that it tangles easily. The cable insulation shall prevent excess current flow to medical personnel or the patient.

The handle shall have a suitable metal fitting to provide a connection with an adapter (not included) fitting into the active outlet of the electrosurgical apparatus. Each handle and cord assembly shall be packaged in a peel-back type container capable of maintaining sterility of contents until manually opened.

Workmanship. The handles shall be free from defects which detract from their appearance or impair their serviceability. The handles shall be free from dust, dirt, oil, and other foreign matter.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

Unit. Package (PG). One package containing twelve handles, as specified, constitutes one unit. Each assembly shall be packaged in a peel-back type container capable of maintaining sterility of contents until manually opened.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of I. The unit of product for examination purposes shall be one electrosurgery handle as specified in this description.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

Defects.

Metal fitting not adequately secured to cord.

Handle not free of cracks, ruptures, and deformed sections.

Cord insulation not free of tears, cracks, or breaks.

Ends of cord not free of exposed wire at points of entrance into handle or fitting.

Handle interior not free of obstructions to chuck component.

Unit not free of dirt, grease and foreign matter.

Cord not flexible as required.

Container not properly sealed to prevent contamination of inner contents.

First article inspection. First article inspection shall be performed on the package of twelve electrosurgery handles when a first article sample is required. This inspection shall include the examination and tests.

Sampling for tests. Sampling for tests shall be conducted as follows:

a. First article. Where a first article sample is required, all tests shall be performed on that unit.

b. Acceptance testing. When a first article sample has not been required, one unit shall be subjected to the performance test specified herein. If a test article sample has successfully passed all tests, performance test shall not be required at acceptance.

c. Dimensional test. Sampling for dimensional test shall be conducted in accordance with MIL-STD-105 with an AQl of 1.5 (percent defective) and an inspection level of S-1. The unit of product for test purposes shall be one electrosurgery handle.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

Tests.

1. Dielectric strength test. A potential of 6000 volts peak to peak at 60 Hz shall be applied between the handle and cord conductor and a metal plate or other suitable electrode. The handle and at least one foot of cord shall be in intimate contact with the metal plate. There shall be no evidence of breakdown with the potential applied for one minute.

2. Strain relief test. The handle shall be secured in place and a force gage attached to the cord. A straight pull shall be applied to the cord parallel to the long axis of the handle until a force of 9 pounds has been reached. A metal fitting on the end of the cord shall be secured and a force gage attached to the cord. A straight pull shall be applied to the cord until a force of 9 pounds is reached. Disconnection of the cord from either the handle or the fitting shall constitute failure of the test.

3. Electrode insertion test. The handle shall be grasped in one hand and a blade electrode shall be inserted into the chuck with the other hand. A straight force shall be used without twisting. Failure of the electrode to seat in and be securely held by the chuck shall constitute failure of this test. The necessity to use greater than normal hand force to seat the electrode shall constitute failure of the test. The depth of insertion of the shaft into the handle shall not exceed 1-1/4 inch.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6515-00-493-3768

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

A-A-50902A

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4212

PGC 45954

Location:
CID/HANDLE/ELECTRODE/H16-32

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-50902A	HANDLE, ELECTRODE, ELECTROSURGERY

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)*:

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4966 WASHINGTON D. C.

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC- RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:48Z
Check the source to verify that this is the current version before use.