A-A-0053210(DM)
30 September 1987

INTERIM COMMERCIAL ITEM DESCRIPTION

RESIN, ACRYLIC, DENTAL
Inlay, Pattern and RESIN ACRYLIC LIQUID, DENTAL

The General Services Administration has authorized the use of this Interim Commercial Item Description which was developed by DLA, Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA 19101-8419, based upon currently available technical information.

This interim commercial item description covers a self-curing acrylic dental inlay pattern resin used in fabricating patterns for casting and performing soldering-related procedures.

Type I	-	Combination package
Type II	-	Standard package
Type III	-	Replacement liquid

The resin packages and/or liquid shall be Duralay Inlay Pattern Resin, combination kit red Part No. 2243, standard package red Part No. 2244 or liquid Part No. 2249C (as applicable) as manufactured by Reliance Dental Manufacturing Company, or equal, with the following salient characteristics.

Salient characteristics:

The resin shall be a red self-curing acrylic (autopolymerizing acrylate) dental inlay pattern resin suitable for use in fabricating patterns for the casting of inlays, crowns and bridges, post and cores, and also suitable for use in performing soldering-related procedures. The resin shall be suitable for use in either the direct or indirect application with either the brush method or packed into or around the preparation in pre-mixed paste form.

Type I - Combination package. The combination package shall consist of two ounces minimum red polymer powder, two ounces minimum monomer liquid, 3.5 milliliters (ml) minimum lubricant, one dropper, three red sable brushes, two "Dappen" type dishes and instructions for use.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

Type II - Standard package. The standard package shall consist of two ounces minimum red polymer powder, two ounces minimum monomer liquid, 3.5 ml minimum of lubricant, a dropper and instructions for use.

Type III - Replacement liquid. The replacement liquid shall consist of a minimum of eight ounces clear, self-curing, acrylic resin liquid monomer.

Type II and Type III shall be compatible with Type I and each other.

Powder. The powder for Type I and Type II shall be a red polymer to contrast the soft tissue and teeth in the mouth. The powder shall be very fine with a minimum of 70% of the powder passing through a #200 mesh screen.

Liquid. The liquid for Type I, Type II and Type III shall be a clear, self-curing acrylic resin monomer. The liquid shall be suitable as a replacement liquid for Type I and Type II.

Lubricant. The lubricant for Type I and Type II shall be petrolatum to facilitate removal of the resin pattern from the tooth cavity. The lubricant shall be red to contrast the tooth structure and not discolor the resin pattern.

Dropper. The dropper of Type I and Type II shall be a three inch glass vial medicine dropper.

Brushes. The brushes of Type I shall have pure red sable tips. The tips shall be securely mounted in seamless metal ferrules and attached to a plastic or wooden, 5-1/2 inches minimum length, handle. The brushes shall be of the following sizes: One brush, size 00 and two brushes, size 3/0 (000). One of the size 3/0 brushes shall have its tip angled at 90^0 to enhance accessibility to the preparation.

Dish. The "dappen" dish of Type I shall be in accordance with the salient characteristics of Commercial Item Description A-A-51072.

Instructions. Instructions for Type I and Type II shall be complete and accurate.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:17Z
Check the source to verify that this is the current version before use.

Resin.

Physical characteristics. The resin shall set within six minutes, maximum. The dimensional stability shall be within + .01%. The resin shall burn out completely resulting in a clean smooth mold, with no residue in 40 minutes at 800^0 Fahrenheit (427^0 Celsius).

Performance characteristics. The resin shall produce a pattern superior to that of inlay wax. The resin shall closely adhere to the cavity walls of the tooth with less shrinkage, reproduce better detail and produce a more accurate impression superior to the corresponding inlay wax pattern. The resin shall have sufficient mechanical strength greater than inlay wax to allow for repeated trials. The set resin shall be sufficiently hard to prevent the loss of shape or distortion, after repeated drawings from the tooth cavity. The resin, when hardened, shall show no appreciable flaking or crumbling when trimmed or carved. The resin shall be compatible with inlay waxes such that these waxes may be added and will adhere to the set resin without adversely effecting either the inlay wax or resin.

Color. Any color additive shall be approved and certified by the Food and Drug Administration under the Federal Food, Drug and Cosmetic Act.

Irritation. Unless otherwise specified, prior to delivery the manufacturer shall furnish a statement directly to the Defense Personnel Support Center, DPSC-RSTS-2, Directorate of Medical Materiel, Technical Services Branch, 2800 South 20th Street, Philadelphia, PA 19101-8419 stating that the resin does not normally irritate oral tissues and does not contain poisonous ingredients in sufficient concentration to be harmful to human beings when used as directed.

Delivery schedule. Not more than three months shall have elapsed from the date of manufacture of the resin to the date of delivery to the Government.

Workmanship. The resin shall be free from lumps and foreign matter and any defect which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing all component Type I or Type II resin acrylic dental, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:17Z
Check the source to verify that this is the currentversion before use.

Unit. Bottle (BT). For Type II. One bottle containing 8 Oz. (240 ml) of resin acrylic liquid, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection purposes shall be one package of resin or 1 bottle of replacement liquid. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-1, AQL 1.0.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:17Z
Check the source to verify that this is the currentversion before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractor/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document.

Stock Number	Item Identification	Type
6520-01-131-3438	ACRYLIC RESIN LIQUID, DENTAL, 8 Ounces.	III
6520-01-144-1517	RESIN, ACRYLIC, DENTAL, Inlay, Pattern, Red.	II
6520-01-151-1853	RESIN, ACRYLIC, DENTAL, Inlay, Pattern, Red, Combination Kit.	I

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:17Z
Check the source to verify that this is the currentversion before use.

A-A-0053210(DM)

MILITARY INTERESTS:

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD - MB

Agent:

DLA - DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PROJECT NO. 6520-2056

59216

Location: NCR - Disk 2

☆U.S. GOVERNMENT PRINTING OFFICE: 1987 - 505-033-75635

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.