A-A-53227A
17 June 1988
SUPERSEDING
A-A-0053227(DM)
25 September 1987

COMMERCIAL ITEM DESCRIPTION

WRAPPER, STERILIZATION

(Heavy Duty)

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers requirements for a blue, single use, wrapper made of nonwoven fabric.

Salient characteristics:

The wrapper shall be capable of maintaining sterility for 30 days without a dust cover in a clean, temperature controlled, humidity controlled environment. It shall be suitable for wrapping heavy or bulky items. The wrapper shall be supplied in four different sizes.

Size 1 - 24 in. x 24 in., quantity of 500's
Size 2 - 36 in. x 36 in., quantity of 250's
Size 3 - 54 in. x 54 in., quantity of 100's
Size 4 - 54 in. x 72 in., quantity of 100's

Each sterilization wrapper shall be clean, single use, fluid repellent, tear/abrasion resistant, memory free, and lint free with bacteria barrier properties for wrapping heavy or bulky items when processing items for sterilization.

The sterilization wrapper shall be fabricated of an adhesive bonded wood pulp fiber that has been chemically treated to produce a bacterial barrier and fluid repellency material. Each wrapper shall be of uniform construction.

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

The wrapper shall also be freely permeable to steam and ethylene oxide penetration. Efficient concentration of steam shall be assured to guarantee sterilization of the contents when used as a double wrapper on packs 12 x 12 x 20 inches (30.5 x 30.5 x 50.8 cm). The wrapper shall have not more than 3 percent weight gain during steam sterilization of the whole pack, 12 x 12 x 20 inches (30.5 x 30.5 x 50.8 cm), for drying efficacy. Each shall insure adequate air removal for drying efficacy.

The wrapper shall meet electrical conductivity standards NFPA 99, and flame resistance standards NFPA 702, class III. The wrapper shall be nonallergenic as evidenced by no primary irritation or sensitization as evaluated by the U.S. Public Health Service, Dermatosis Section.

The wrapper shall meet Association of Operating Room Nurses (AORN) Recommended Practices for In-Hospital Packaging Materials, published Feb. 1983, (or latest revision).

Pack density shall be no larger than 7.2 pounds/cubic feet. Shall be strong enough to hold a 17 pound (8.5 Kg) instrument set.

The wrapper shall be in accordance with the following test characteristics:

Strength:

Grap Tensile (lb)	30 x 12
Elm Tear (gms)	251 x 157
Burst (PSI)	39

Barrier:

Mason Jar (min)	0
HSH (cm)	16
Water Imp (gms) (1-ply)	27.0
(2-ply)	12.5
Water spray	70
Air Perm (2-ply)	31.5
BFE percent (1-ply)	59.6
(2-ply)	89.7
Alcohol repellency	9*

Aesthetics:

HOM (gms)	253 x 68
Drape Resistance (in\lbs.)	1.3 x 1.8
Wet Colorfastness	Good

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

Weight:

Ounces per square yd(min) 2.0

Miscellaneous:

Tape Adhesion	Good
Static Decay (sec)	<0.5
Flammability (sec)	>3.5

For Single Use Only
*Based on 1 (poor) to 10 (excellent) rating systems.

Workmanship. The sterilization wrapper shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG).

For size 1, one package containing 500 sterilization wrappers as specified, constitutes one unit.

For size 2, one package containing 250 sterilization wrappers as specified, constitutes one unit.

For size 3 and size 4, one package containing 100 sterilization wrappers of one size only, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of pnysical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 with an AQL of 1.0 and an inspection level of II. The unit of product for examination purposes shall be one sterilization wrapper as specified in this description.

Defects.

Wrapper not made of an adhesive bonded wood pulp.

Wrapper not free of tears, cuts, holes or perforations.

Weight of wrapper not correct.

Wrapper not free of grease, dirt, or foreign material.

Sampling for tests. Sampling for tests including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 1.5 (percent defective) and inspection level of S-1. The unit of product for tests purposes shall be one sterilization wrapper.

Tests

Tests shall be conducted to determine compliance with table III. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

a. Surface Resistivity NFPA 99
b. Flame Resistivity NFPA 702

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

Size 1 - 6530-01-254-5826
Size 2 - 6530-01-156-3399
Size 3 - 6530-01-156-3398
Size 4 - 6530-01-169-3153

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6530-1998

Location: CID/WRAP/H18-23

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS. In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problems, areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-[redacted] RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-53227A	WRAPPER, STERILIZATION (Heavy Duty)

3a. NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)*

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6 REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.