A-A-53229A
13 October 1989
SUPERSEDING
A-A-0053229 (DM)
6 October 1987

COMMERCIAL ITEM DESCRIPTION

INTRODUCER SET, CATHETER, PERCUTANEOUS

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a catheter introducer set suitable for cardiovascular procedures.

Salient characteristics:

General. The introducer set shall be capable of percutaneous introduction of 7.5 or 8.0 French size closed end catheters, balloon catheters, and pacing leads into the cardiovascular system.

Components. The set shall consist of the following components:

Component	Length (cm)
Introducer Sheath	10 to 12
Vessel dilator	15 to 20
Side port extension	15 to 22
3 mm "J" guide wire (0.032 to 0.038 inch dia.)	40 to 50

Material. The sheath, dilator, and side port extension shall be fabricated of materials which meet the requirements of the USP for Class IV plastic. The guide wire shall be fabricated of Class 3 corrosion-resistant steel in accordance with ASTM Designation F899.

FSC 6515

AMSC/NA

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:17Z
Check the source to verify that this is the current version before use.

Style and design. The sheath and dilator may be separate components or unit component. After insertion of the sheath and dilator, guide wires shall not be required for catheter exchanges. The side port extension shall connect to the sheath by means of a hemostatic valve. The hemostatic valve shall prevent back-flow of blood and prevent air embolus formation when used properly. The side port extension shall be transparent or translucent and be capable of permitting continuous flushing, periodic aspiration, pressure monitoring, and serve as an infusion site.

Sterility, toxicity and pyrogenicity. The introducer set shall be sterile, non-toxic, and non-pyrogenic in accordance with the USP. Sterility assurance level shall be 1×10^{-6}.

Workmanship. The sheath and dilator shall have smooth mating surfaces. The introducer shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing ten individually packaged sterile introducer sets, as specified, constitutes one unit. The individual package shall be a sealed, peel-open type container capable of maintaining sterility of contents unless the package is damaged or opened.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:17Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, inspection level S-1 and acceptance number zero. The unit of product for test purposes shall be one introducer set.

Tests. The introducer set shall be subjected to the tests specified by the USP for Class IV plastics and transfusion and infusion assemblies.

Certificate of biocompatibility. A certificate of compliance, stipulating offered items meet the requirements specified herein, shall be acceptable in lieu of actual testing, provided that the contractor has supplied acceptable items to the Government within the past three years, and that there has been no changes to material formulation and method of processing. The certificate shall include the current contract number and applicable NSN, the date of actual testing, the test method and results, the name and title of the person evaluating tests and a statement to the effect that material formulation and processing have not changed since the testing was performed.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:17Z
Check the source to verify that this is the current version before use.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subscontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-253-8293	INTRODUCER,CATHETER, PERCUTANEOUS, RADIOPAQUE, Disposable 8 Fr, 10s, Sterile

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4555

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:17Z
Check the source to verify that this is the current version before use.