A-A-53234
29 September 1987

COMMERCIAL ITEM DESCRIPTION

PLIERS, DENTAL (STEIGLITZ, 5-3/16 inches)

The General Services Administration has authorized the use of this Commercial Item Description.

This commercial item description covers a dental pliers of steiglitz design suitable for use in the placement or recovery of silver points during endodontic procedures.

Salient characteristics:

General. The pliers shall be in accordance with the applicable requirements of GG-I-526 and as specified herein.

Material. The pliers shall be fabricated from class 4, type 410 or 420 martensitic stainless steel. As an alternate, the screw or pin shall be fabricated of class 4, type 416 martensitic stainless steel. The composition for the pliers, and screw or pin shall be as specified in ASTM F899.

Hardness. After proper stress relieving, the pliers shall have a Rockwell hardness of C40 to C45. The joint screw or pin shall be of adequate hardness for purposes intended when tested as specified herein.

Distortion. The pliers shall meet the distortion requirements specified in GG-I-526 when tested as specified herein.

Style, design and dimensions. The pliers shall be of Steiglitz design, suitable for use in the placement and recovery of silver points during endodontic procedures. The pliers shall be the finger-ring, ratchet-locking type. The pliers shall be capable of gripping, clamping, and holding silver points having a diameter (non-operative end) from 0.40 to 1.31 millimeters and shall permit manual transfer of the pliers while holding a silver point. The inside surface of the jaws shall have transverse serrations with respect to the long axis of the jaws and a central concavity to facilitate gripping of the silver point. The style, design and dimensions of the pliers shall conform to Defense Personnel Support Center (DPSC) Drawing 23212.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

Locks. The joint of the pliers shall be a box lock or lap joint type, conforming to the requirements of GG-I-526. The joint shall not bind, and shall not exhibit play or looseness in any position of the joints from full open to full closed position. The edges of the joints shall be properly beveled or chamfered to prevent pinching of tissues. The joint of the pliers shall be properly centered and uniform, in cross-section when tested as specified herein.

Finish. The external surfaces of the pliers shall have a bright, satin or matte finish, except for the inside portions of the jaws and inside surface of the finger ring shall have a smooth polished surface. The matte finish shall exhibit a uniform nondirectional surface texture equivalent to that produced by lightly blasting the prepared surface with fine abrasive under pressure. The outside surfaces of the joint shall present a smooth surface when the pliers are in the closed position. The outside surface of the pin or screw shall be flush with the outside contour of the pliers in the open or closed position.

Corrosion resistance. The pliers shall meet the resistance to copper sulfate and resistance to boil test requirements of GG-I-526 when tested as specified herein.

Identification marking. Marking shall be in accordance with applicable requirements of GG-I-526. Marking of instruments of foreign manufacture shall also be marked with the word "Stainless". All marking shall be legible and permanent.

Workmanship. The pliers shall be free from defects which detract from their appearance or may impair their serviceability, proper functioning, or intended use.

Unit. Each (EA). One pliers, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection purposes shall be one pair of pliers. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-1, AQL 1.0.

Tests.

Hardness. Hardness shall be determined in accordance with American Society for Testing and Materials (ASTM) Designation E18 (Rockwell) or E92 (Vickers or Diamond Pyramid).

Distortion. The pliers shall be subjected to the distortion test as specified in GG-I-526, except that the thickness of the wood or strap shall be 3/32 to 1/8 inch.

Lock stiffness. The pliers shall be subjected to the lock stiffness test as specified in GG-I-526.

Corrosion resistance. The pliers shall be subjected to the copper sulfate and the boil tests as specified in GG-I-526.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

Metric Products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerance specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing, shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

National Stock Number	Item Identification
6520-00-224-0397	PLIERS, DENTAL, STEIGLITZ, 5-3/16 inch (13.17 cm).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

47534

PREPARING ACTIVITY:

DoD-MB

AGENT:

DLA-DM

Project No. 6520-2015

Location: NCR Disk #4

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements All users of military standardization documents are invited to provide suggestions This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive loose, ambiguous, or was incompatible and give proposed wording changes which would alleviate the problems Enter in block 6 any remarks not related to a specific paragraph of the document If block 7 is filled out an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN DPSC-ATTS
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

TO DETACH THIS FORM, CUT ALONG THIS LINE.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions - Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-53234	PLIERS, DENTAL (STEIGLITZ, 5-3/16 inches)

3a NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b ADDRESS *(Street City State ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)*

5 PROBLEM AREAS

a Paragraph Number and Wording

b Recommended Wording

c Reason/Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER *(Last First MI)* - Optional	b WORK TELEPHONE NUMBER *(Include Area Code)* - Optional
c MAILING ADDRESS *(Street City State ZIP Code)* - Optional	8 DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.