A-A-0053245(DM)
13 October 1987

INTERIM COMMERCIAL ITEM DESCRIPTION

LUMBAR PUNCTURE KIT

The General Services Administration has authorized the use of this interim commercial item description which was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA, 19101, based upon currently available technical information.

This interim commercial item description covers a lumbar puncture kit for infants.

Salient characteristics.

General. The lumbar puncture kit shall be sterile, disposable and suitable for use in the withdrawal of spinal fluid for diagnostic procedures.

The kit shall consist of the following items as a minimum:

One 22 gauge by 1-1/2 inch spinal needle.

One syringe 2-1/2 ml with 25 gauge by 5/8 inch needle, with the syringe marked in 1/10 ml increments.

One 2 ml vial of 1% Lidocaine HCl, USP.

One fenestrated drape of an approximate 14 inch square.

Three prep swabs. Each swab shall be one inch square minimum, with attached handle.

Three gauze sponges approximately 2 inches by 2 inches.

One nonwoven towel.

Three specimen vials with caps. Each vial shall be numbered and have minimum volume capacity of 5 mls.

One removable prep cup with minimum 25 ml volume.

One three-way stopcock.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

One manometer gauged to 35 cm of spinal fluid in 1/2 cm increments.

One polyvinylchloride extention tube of approximate 6 inch length supplied with rigid end connectors.

One puncture-site adhesive bandage.

Material. The lidocaine glass ampul shall be Type I or Type III, prescored, meeting the requirements of the Pharmacopeia of the United States (USP). The ampul shall be made of molded glass or drawn glass tubing, with walls of uniform thickness with drawn, rolled-in, or stuck-on necks. The ampul shall be supplied ready for opening and shall break cleanly.

Plastics shall meet the requirements of the USP for Class II plastics.

Needles shall be fabricated from austenitic stainless steel with composition in accordance with ASTM F899. All needles shall be furnished with protectors.

The kit shall be packaged in a container capable of maintaining the sterility of the tray and contents until manually opened. The tray shall be ready for use and have all new components in sequence lined packaging.

Expiration dating. The kit shall be furnished with no less than 30 months to the expiration date at the time of delivery to the Government.

Sterility and pyrogenicity. Components shall be sterile and free from pyrogens. The kit shall have a minimum sterility assurance level of $1X10^{-6}$.

Workmanship. The kit shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing ten individually packaged sterile lumbar puncture kits, as specified, constitutes one unit. The individual package shall be a sealed, peel open type container capable of maintaining sterility of contents unless the package is damaged or opened.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination. The lumbar kit shall be examined to determine compliance with specification requirements. Sampling shall be in accordance with MIL-STD-105, sampling level II, AQL 2.5.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified below. The unit of product for test purposes shall be one lumbar puncture kit. Sampling shall be at an inspection level of S-2 with an AQL of 1.0.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:17Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-181-3155	LUMBAR PUNCTURE KIT, Infant, With Manometer, Sterile, Disposable, 10's

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PGC S0134 Project No. 6515-4216

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.