A-A-53246
8 October 1987

# COMMERCIAL ITEM DESCRIPTION

## FORCEPS, HYSTERECTOMY (Heaney-Ballantine)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a hysterectomy forceps of the Heaney-Ballantine design, used in gynecological procedures.

Salient characteristics.

General. Both jaws of the hysterectomy forceps shall have parallel longitudinal grooves and a single lateral tooth must have been dropped approximately 1/4 inch from the distal tip. The definitions of the terms and components used in this specification are in accordance with the applicable requirements of ASTM F1026.

Material. The forceps shall be Class 4 corrosion-resistant steel of composition in accordance with American Society for Testing and Materials (ASTM) Standard F899.

Hardness. The Rockwell hardness of the forceps shall be C40 to C47.

Style, design and dimensions. The forceps shall be in accordance with Defense Personnel Support Center (DPSC) Drawing 24478.

Passivation. The forceps shall be properly passivated.

Finish. The finish of the forceps shall be mirror, satin or matte. The matte finish shall exhibit a uniform nondirectional surface texture equivalent to that produced by lightly blasting the prepared surface with fine abrasives under pressure.

Corrosion-resistance. The forceps shall shown no signs of corrosion or plating of copper when tested for corrosion-resistance as specified in Quality Assurance Provisions.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

Specific performance. Box lock performance, jaw and ratchet operation, and elasticity shall be as specified in ASTM F1026.

Identification marking. In addition to the requirements specified in ASTM F1026, the forceps shall also be marked "stainless".

Tests. Corrosion-resistance, hardness and performance shall be tested as specified under Quality Assurance Provisions.

Workmanship. The forceps shall be processed in such a manner as to be uniform and free from defects which affect life, utility, or appearance.

PRODUCT QUALITY LEVEL.

Visual and dimensional examination. Shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements.

Materials. The supplier shall be responsible for assuring that materials are in accordance with requirements specified.

Unit. Each (EA). One hysterectomy forceps, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination. The forceps shall be examined to determine compliance with all requirements contained in this specification.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one forceps.

TABLE I. Sampling For Inspection.

| Type of defect | Inspection level | AQL (percent defective) |
|---|---|---|
| Major | II | 1.0 |
| Minor | II | 4.0 |

TABLE II. Classification Of Defects

| Categories | Defects |
|---|---|
| Critical | None defined. |
| Major | |
| 101 | Forceps not free from cracks, pits, burs, sharp edges, scale, rough areas or corrosion. |
| 102 | Finish of the forceps not specified. |
| 103 | Style and design not as specified. |
| Minor | |
| 201 | Identification markings not present, not complete, not correct, or not legible. |

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified below. The unit of product for test purposes shall be one forceps.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

TABLE III. Sampling For Tests.

| Characteristic | Requirement | Test Procedures | Inspt. level | AQL |
|---|---|---|---|---|
| Composition | ASTM F899 | | * | 0.0 |
| Hardness test | ASTM E18<br>E92<br>E140 | ASTM E18<br>E92<br>E140 | S-2 | 1.0 |
| Dimensions | Rqrmt's | | S-1 | 2.5 |
| Copper Sulfate test | GG-I-526 | GG-I-526 | S-2 | 1.0 |
| Boil test | GG-I-526 | GG-I-526 | S-2 | 1.0 |
| Performance | Rqrmt's | ASTM F1026 | S-2 | 1.0 |

Tests.

Corrosion-resistance. The forceps shall be tested as specified in Federal Specification GG-I-526.

Hardness. Hardness shall be in accordance with ASTM Standard E18 or E92 whichever is applicable. Conversion of hardness values shall be in accordance with ASTM E140.

Performance. Performance shall be tested in accordance with applicable requirements of ASTM F1026.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

| NSN | ITEM IDENTIFICATION |
|---|---|
| 6515-01-139-8187 | FORCEPS, HYSTERECTOMY, Straight, Heaney-Ballantine, 8 Inch (20.3 cm) |
| 6515-01-164-8886 | FORCEPS, HYSTERECTOMY, Curved, Heaney-Ballantine, 8-1/4 Inch, (20.96 cm) |

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

A-A-53246

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-3498



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

**INSTRUCTIONS** In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

*NOTE* This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

*(Fold along this line)*

*(Fold along this line)*

DEFENSE LOGISTICS AGENCY

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

**BUSINESS REPLY MAIL**
FIRST CLASS PERMIT NO. 4966 WASHINGTON D.C.

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-RSTS
2800 South 20th Street
Philadelphia, Pennsylvania 19101-6419

NO POSTAGE
NECESSARY
IF MAILED
IN THE
UNITED STATES

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

# STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

*(See Instructions - Reverse Side)*

| 1. DOCUMENT NUMBER | 2. DOCUMENT TITLE |
|---|---|
| A-A-53246 | FORCEPS, HYSTERECTOMY (Heaney-Ballantine) |

| 3a. NAME OF SUBMITTING ORGANIZATION | 4. TYPE OF ORGANIZATION (Mark one) |
|---|---|
| | ☐ VENDOR |
| | ☐ USER |
| b. ADDRESS (Street, City, State, ZIP Code) | ☐ MANUFACTURER |
| | ☐ OTHER (Specify) |

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

| 7a. NAME OF SUBMITTER (Last, First, MI) - Optional | b. WORK TELEPHONE NUMBER (Include Area Code) - Optional |
|---|---|
| c. MAILING ADDRESS (Street, City, State, ZIP Code) - Optional | 8. DATE OF SUBMISSION (YYMMDD) |

DD FORM 82 MAR 1426

PREVIOUS EDITION IS OBSOLETE

(FOLD ON THIS LINE)

(FOLD ON THIS LINE)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.