A-A-53253
26 October 1987

COMMERCIAL ITEM DESCRIPTION

RONGEUR (Straight, Lempert)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers straight rongeur of the Lempert design for otolaryngological surgical procedures.

Salient characteristics:

General. Shall be single action rongeur of box lock construction, and shall have cupped jaws. The slender, light rongeur shall have a double spring. The definitions of terms and components used in this specification are in accordance with ASTM F921. Performance shall be in accordance with applicable requirements of ASTM F1026.

Material. The material of the rongeur shall be Class 4 martensitic stainless steel in accordance with ASTM F899.

Hardness. The Rockwell hardness of the rongeur shall be C-48 to C-55.

Style and design. The rongeur shall be in accordance with Defense Personnel Support Center (DPSC) Drawing No. 22268.

Finish. The finish of the rongeur shall be mirror, satin, or matte. The matte finish shall exhibit a uniform, non-directional surface texture equivalent to that by lightly blasting the prepared surface with fine abrasives under pressure.

Corrosion-resistance. The rongeur shall show no signs of corrosion or plating of copper when tested for corrosion resistance as specified in Quality Assurance Provisions.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

Other performance. When closed, the cupped portion of the jaws shall mate completely starting at the tips. The action of the rongeur shall be easy, without drag or binding. Under no circumstances should looseness or free play of either shank permit misalignment of the jaws in closing.

Identification markings. In addition to the requirements specified in ASTM F1026, the rongeur shall also be marked "stainless".

Tests. Corrosion resistance, hardness, and performance shall be tested as specified under Quality Assurance Provisions.

Workmanship. The rongeur shall be free from defects which detract from its appearance or impair its serviceability.

Product Quality Level.

Visual and dimensional examination. Shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements.

Materials. The supplier shall be responsible for assuring that materials are in accordance with requirements specified.

Unit. Each (EA). One rongeur, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, and as indicated in Table I. Unit of product for examination purposes shall be one rongeur.

TABLE I. Sampling For Inspection.

Type of defect	Inspection level	AQL (percent defective)
Major	II	1.0
Minor	II	4.0

TABLE II. Classification Of Defects.

Categories	Defects
Critical	None defined.
Major	
101	Rongeur not free from cracks, pits, burs, sharp edges, scale, rough areas or corrosion.
102	Finish of the rongeur not as specified.
103	Style and design not as specified.
Minor	
201	Identification markings not present, not complete, not correct, or not legible.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified below. The unit of product for test purposes shall be one rongeur.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

TABLE III. Sampling For Tests.

Characteristic	Requirement	Test procedures	Inspt. level	AQL
Composition Analysis	ASTM F899	Rqrmt's	*	0.0
Hardness test	ASTM E18 E92 E140	ASTM E18 E92 E140	S-2	1.0
Dimensions	Rqrmt's	--------	S-1	2.5
Copper Sulfate test	GG-I-526	GG-I-526	S-2	1.0
Boil test	GG-I-526	GG-I-526	S-2	1.0
Performance	Rqrmt's	--------	S-2	1.0

* Sufficient amount from each lot of each metal used in the construction of the instrument shall be selected for composition analysis.

Tests.

Corrosion resistance. The rongeur shall be tested as specified in Federal Specification GG-I-526.

Composition analysis. The rongeur shall be subjected to analysis in accordance with methods 111 or 112 of Federal Test Method Standard No. 151.

Hardness. Hardness of the rongeur shall be tested in accordance with ASTM E18 or E92 whichever is applicable. Conversion of hardness values shall be in accordance with ASTM E140.

Performance. Performance of the rongeur shall be tested in accordance with applicable requirements of ASTM F1026.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number(s) are covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-113-9210	RONGEUR, STRAIGHT, LEMPERT, 7-1/4 Inch (18.4 cm)
6515-01-139-8291	RONGEUR, STRAIGHT, LEMPERT, SLENDER, LIGHT, 6-1/2 Inch (16.5 cm)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4118

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2 DOCUMENT TITLE

3a. NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b ADDRESS *(Street City State ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6 REMARKS

7a. NAME OF SUBMITTER *(Last First MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street City State ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426

PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements All users of military standardization documents are invited to provide suggestions This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems Enter in block 6 any remarks not related to a specific paragraph of the document If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-ATTS
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.