A-A-53263
1 March 1988

COMMERCIAL ITEM DESCRIPTION

ADHESIVE LIQUID, SURGICAL (AEROSOL)

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a Sterile, Liquid Surgical Adhesive.

Salient Characteristics:

Shall be a sterile liquid surgical adhesive in aerosol spray form suitable for use in securing surgical dressings and surgical film to the skin. The adhesive shall adhere to glass and human skin and shall not contain any foreign matter. When sprayed onto glass and onto human skin the adhesive shall dry to the point of adherence within 90 seconds and shall form an adhesive film.

After spraying onto glass the adhesive film shall be examined for foreign matter. When applied to human skin the adhesive film shall be adhesive, flexible and peelable and shall not cause dermatological irritation. The adhesive film shall be capable of being removed with alcohol and with acetone.

The adhesive shall be of the following formulation (nominal percentages by weight):

a. Hydroxy-Vinyl Chloride Acetate Copolymer, 5.9 percent.
b. Sebacic Acid Copolymer, 2.9 percent.
c. Modified Maleic Rosin Ester, 2.2 percfent.
d. Glycolate Plasticizer, 3.7 percent.
e. All of the above in an Ethyl Acetate- Acetone solvent with propane gaseous propellant.

Neither the can nor the valve shall leak when the valve is in the on position and when it is in the off position. When released, the valve shall shut completely off.

Workmanship. The cans of adhesive shall be free from defects which detract from their appearance or impair their serviceability.

AMSC/NA

FSC 6510

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

Unit. Can (CN). One can containing 5 fl oz, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item. Testing and examination shall be conducted to determine compliance with all specification requirements. Test methods shall be suitable, accurate and reproducible.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for examination shall be one can of adhesive. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5 maximum percent defective.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

Metric Products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6510-01-126-4327	ADHESIVE LIQUID, SURGICAL, 5 fl oz

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

Ordering data. Expiration dating required, intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PGC 44412

Project No. 6510-0666

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.