A-A-53267A
20 April 1990
SUPERSEDING
A-A-0053267 (DM)
6 November 1987

COMMERCIAL ITEM DESCRIPTION

ANALYZER, CENTRIFUGAL, HEMATOLOGY

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers the requirements for centrifugal hematology analyzer for use in a laboratory environment.

Salient characteristics.

The centrifugal hematology analyzer shall be a seven-parameter hematology screening device which yields quantitative values from centrifugal blood tubes for the following parameters:

Hemoglobin
Hematocrit
Platelet count
White blood cell count
Granulocyte count (as a percentage of total white count)
Granulocyte count (total count)
Lymphocyte-monocyte (as a percentage of total white count)
Lymphocyte-monocyte (total count)

Platelet count, white blood cell count and counts of the granulocyte and lymphocyte/monocyte white cell subpopulation shall be derived from electro-optical measurements of the packed cell volumes.

Shall be suitable for use with venous and capillary hematocrit tubes having the following salient characteristics: The length of the tube shall be 2.953±0.10 inches and have an inner diameter of 0.066±.00035 inch.

FSC 6630

AMSC N/A
DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

Shall include the following:

a. Microprocessor controlled electro-optical instrument for measuring packed cell volumes and converting to quantitative cell counts. Components of the instrument include a 7.5 monocular microscope, rotating tube carriage assembly, optical encoder (reader), UV illuminator lamp, various lenses and filters to cause stained white cells to fluoresce.

b. A table model hematocrit centrifuge with a 12 place head, head cover and a fixed 5 minute timer. Shall operate at a nominal speed of 12,000 RPM and a nominal relative centrifugal force (RCF) of 14,387 gravities.

c. A semi-automatic pipettor for filling the microhematocrit tubes with venous blood. The plunger of the pipettor shall have two positions: Fully retracted and fully depressed. Shall have a non-adjustable fill volume at 111.1 microliters. Shall include spare O-rings and grease.

d. A work station to hold a tray with the capacity to hold ten blood tubes, slotted receptacle for the storage of the pipettor, various sized blood tube wells and notched front rack for the storage of centrifuged blood tubes.

e. A spare parts kit which includes the following: venous and capillary check tube; spare lamps; 1 fuse; pair of forceps for handling loose floats; 5/64 inch hex key for service and maintenance; and a counter top test guide.

Each check tube shall consist of a 3 inch glass capillary tube containing alternating segments of plastic. Both ends shall be internally sealed. One end shall be capped with a standard tube closure. When inserted in the reader the white plastic shall fluoresce a pale green. The check tubes serve to verify the ability of the reader to measure band lengths and derive values.

The analyzer shall operate from either a 120 volt, 60 Hz or 240 volt, 50 Hz supply lines, as specified in the contract or purchase order (See ordering data).

Workmanship. The analyzer shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One analyzer, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for tests. Sampling for tests shall be conducted in accordance with commercial practice. The unit of product for tests purposes shall be one analyzer.

Tests. Tests shall be conducted to determine compliance with specification requirements.

Performance.

a. Centrifuge. The centrifuge shall be tested to demonstrate its ability to operate properly.

Centrifugation Time. The centrifuge shall be connected to a suitable electrical source. The centrifuge shall be operated by turning the switch to the "on" position, and the time of the programmed centrifugation shall be determined using a suitable timer.

Rotor Speed. The centrifuge shall be connected to a suitable electrical source. The centrifuge shall be operated and the rotor speed shall be measured by use of a strobe light or other suitable means.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:17Z
Check the source to verify that this is the currentversion before use.

b. Optical reader. The optical reader shall be tested to demonstrate its ability to operate properly.

Operational Test. The optical reader shall be plugged into a suitable electrical source. The power source switch shall be turned on to the self check sequence. The ability of the reader to measure bands lengths and derive values, and satisfactory operation of all lamps and displays shall be monitored.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/ contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:17Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6630-01-234-6794 ANALYZER, CENTRIFUGAL HEMATOLOGY

Ordering data. (Applicable voltage required intermediate/exterior package quantities and labeling and marking must be specified in tne contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6630-0503

Location: CID/ANALYZER/S35/12

Source: https://assist.dla.mil -- Downloaded: 2026t0-26T09:17Z
Check the source to verify that this is the currentversion before use.