MIL-B-0037189A(DLA-DM)
11 April 1978
USED IN LIEU OF
MIL-B-37189
29 April 1975

MILITARY SPECIFICATION

BAGS, URINE COLLECTION

This limited coordination military specification has been prepared by the Defense Personnel Support Center based upon currently available technical information but it has not been approved for promulgation as a coordinated revision of MIL-B-37189. It is subject to modification. However, pending its promulgation as a coordinated specification, it may be used in procurement.

1 SCOPE

1.1 Scope. This specification covers two types and three styles of disposable, plastic, closed system, sterile urine collection bags suitable for use in the collection and measurement of urinary output.

1.2 Classification. The urine bags shall be of the following types and styles, as specified (see 6.2):

1.2.1 Types.

Type I - Bag, Urine Collection.
Type II - Bag, Urine Collection (with urometer).

1.2.2 Styles.

Style A - Translucent with white opaque back.
Style B - Translucent.
Style C - Translucent with white opaque back and vented adapter.

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-ATT, 2800 South 20th Street, Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

FSC 6530

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:09Z
Check the source to verify that this is the current version before use.

2. APPLICABLE DOCUMENTS

2.1 Issues of documents. The following documents of the issue in effect on date of invitation for bids or request for proposals, form a part of this specification to the extent specified herein.

SPECIFICATIONS

FEDERAL

PPP-B-636 - Boxes, Shipping, Fiberboard.

MILITARY

MIL-L-10547 - Liners, Case, and Sheet, Overwrap, Water-Vaporproof or Waterproof, Flexible.
MIL-S-36586 - Sterilization Test Strip Set, Bacterial Spore.

STANDARDS

MILITARY

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 - Marking for Shipment and Storage.

(Copies of specifications, standards, drawings, and publications required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposal shall apply.

AMERICAN SOCIETY FOR TESTING AND MATERIALS (ASTM)

D882 Method of Test for Tensile Properties of Thin Plastic Sheeting

(Application for copies should be addressed to the American Society for Testing and Materials, 1916 Race Street, Philadelphia, PA 19103.)

U. S. PHARMACOPEIAL CONVENTION, INC.

The United States Pharmacopeia (USP).

(Application for copies should be addressed to the Mack Publishing Company, Easton, PA 18042.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:09Z
Check the source to verify that this is the current version before use.

3. REQUIREMENTS

3.1 First article. When specified in the procurement document, the contractor shall furnish sample unit(s) for first article inspection and approval (see 4.1.2.1 and 6.2).

3.2 Material. The material shall be as specified herein, of the highest quality throughout, and suitable in all respects for the purpose intended.

3.2.1 Polyvinyl chloride (PVC). The PVC sheets shall have a minimum tensile strength of 2000 p.s.i. and a minimum elongation of 200 percent when tested as specified in 4.5.2. Where white pigment has been added to the PVC, an average elongation of 175 percent shall be acceptable.

3.2.2 Wire. The wire hangers, if applicable, shall be 0.120 inch, (No. 11 Birmingham Wire gage) cold drawn steel, zinc, cadmium or tin plated.


3.2.3 Cord. The cord used with the wire hangers shall be nylon, cotton, or wax impregnated cotton, tightly woven, 20 to 36 inches long and 3/32 inch minimum diameter. The cord shall be furnished with a suitable corrosion-resistant hook, capable of supporting the weight of a bag filled with 2000 cc of water.

3.2.4 Strap hanger. Plastic strap hangers, if used in lieu of wire or cord, shall be fabricated of PVC material suitably notched for attachment to the bag. The strap shall be capable of supporting the weight of a bag filled with 2000 cc of water.

3.3 Construction.

3.3.1 Type I. Type I sterile urine collection bag shall consist of a urine collecting bag, an inlet tube with a safety chamber, a bottom drain tube with a shut-off mechanism, and a hanger for suspension from the bed, chair or patient. The closed system shall be the same or similar in design with minor variations as Style A, B or C as shown in Figure 1 for Style A, Figure 2 for Style B or Figure 3 for Style C. The collection bag shall be packaged in a polyethylene bag or polyethylene/polymylar pouch or other suitable overwrap which shall maintain sterility of the bag until manually opened. There shall be no water inside the overwrapper or the collection bag when tested as specified in 4.5.3. The collection bag shall be sufficiently transparent or translucent to observe the flow and contents. The bag shall be capable of suspension by the catheter adapter, and no joint or connection shall either separate, or begin to separate, when tested as specified in 4.5.4.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:09Z
Check the source to verify that this is the current version before use.



BAG - PVC SHEETS 0.0065 MINIMUM THICKNESS

TOLERANCES

UP TO 1 INCH ± 1/16
1 UP TO 6 INCH ± 1/8
6 UP TO 12 INCH ± 1/4
12 OR GREATER ± 1/2

ALL DIMENSIONS ARE IN INCHES.

FIGURE 1 - Style A - Bag, urine collection.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:09Z
Check the source to verify that this is the current version before use.





FIGURE 2 - Style B - Bag, urine collection.

Source: https://assist.dla.mil -- Downloaded: 202610-21T10:09Z
Check the source to verify that this is the current version before use.



NOTES:

1. TOLERANCES: Up to 1" ± 1/16, 1" up to 6" ± 1/8, 6" up to 12" ± 1/4, 12" or greater ± 1/2" unless otherwise specified.
2. Heat seals 1/16" minimum width.
3. All dimensions are in inches.
4. Taper on catheter adapter shall be from 0.250 inches min. I.D. to 0.465 inches max. O.D.

FIGURE 3 - Style C - Bag, urine collection.

Source: https://assist.dla.mil -- Downloaded: 202610-21T10:09Z
Check the source to verify that this is the current version before use.

3.3.2 Type II. Type II sterile urine collection bags shall consist of a main drainage urine container, an integral or attached urometer, input and drain tube assemblies, and an adjustable hanger to permit suspension from a bed, chair or patient. The closed system shall be the same as or similar to Figure 4. The design shall be such that urine automatically overflows from the urometer into the main bag when the capacity of the urometer is exceeded. Design shall also permit the elective emptying of the urometer contents into the main bag. Closed system shall be maintained during overflow or emptying of urometer contents into main drainage bag. The collection bag shall be packaged in a suitable overwrap which shall maintain sterility of the bag until manually opened. There shall be no water inside the overwrapper or the collection bag when tested as specified in 4.5.3. The collection bag shall be sufficiently transparent or translucent to observe the flow and contents. The bag shall be capable of suspension by the catheter adapter, and no joint or connection shall either separate, or begin to separate, when tested as specified in 4.5.4.

3.3.3 Bag. The urine collection bag shall be constructed of two sheets of PVC, heat sealed on all edges and vented at the top or in the catheter adapter. The bag shall not leak water when tested as specified in 4.5.5. The bag shall not burst at a pressure less than 150 mm Hg when tested as specified in 4.5.8. The bag shall have a capacity of at least 2000 cubic centimeters (cc). It shall be fitted with a horizontal air vent at the top, kept open with a rigid tube in the channel. As an alternate, a circular filter vent secured over a port in the top portion of the bag or an air vent incorporated in the catheter shall be acceptable.

3.3.3.1 Graduations and markings. Graduations and markings shall be clear and indelible when tested as specified in 4.5.6. They shall not be attached or secured with adhesives. The numbers and letters shall be a minimum of 1/8 inch in height.

3.3.3.1.1 Graduations. Type I bag graduations shall be in 100 cc steps beginning at 100 cc and ending at not less than 2000 cc but not more than 2500 cc. At least every other graduation, beginning at 200 cc, shall be identified. Additional graduations shall be permissible in a lower corner. Type II main drainage container bags shall be graduated the same as Type I bags. Urometer graduations of Type II bags shall be in 10 cc steps beginning at 10 cc and ending at not less than 200 cc. From 10 cc to 80 cc the following graduation markings shall at least be identified: 10, 20, 30, 40, 50, 60, 70 and 80 cc; from 80 cc to 200 cc, the following graduation markings shall at least be identified: 80, 100, 120, 140, 160, 180 and 200 cc.

3.3.3.1.2 Marking. Other marking shall be as shown on Figures 1, 2, 3 or 4. The manufacturer's name or registered trademark shall be printed on the bag.

3.3.3.2 Volumetric accuracy. Type I bags shall have a volumetric accuracy of $\pm$ 10 percent when tested as specified in 4.4.7. Type II bags shall have a volumetric accuracy of $\pm$ 10 percent for the main drainage urine container and an accuracy of $\pm$ 5 percent for the urometer when tested as specified in 4.5.7 and 4.5.7.1.

Source: https://assist.dla.mil -- Downloaded: 202610-21T10:09Z
Check the source to verify that this is the currentversion before use.




FIGURE 4 - Type II - Bag, urine collection.

Source: https://assist.dla.mil -- Downloaded: 202610-21T10:09Z
Check the source to verify that this is the current version before use.

3.3.4 Urometer (Type II bags). Shall be fabricated from clear plastic material to permit the viewing of flow and contents. Urometer shall be either integral or adequately attached to the main drainage urine container.

3.3.5 Safety chamber. The safety chamber shall be of rigid plastic and shall be permanently attached to the inlet tube and collection bag. As an alternate, the vent may be located in the catheter adapter. The design shall be such as to prevent the retrograde migration of bacteria. Type II safety (drip) chamber shall be attached to the inlet tube and urometer.

3.3.6 Inlet tube. The inlet tube shall have a minimum wall thickness of 0.040 inch. The PVC inlet tube shall be permanently attached to the catheter adapter and the safety chamber or inlet port. The catheter adapter shall be fitted with a tip protector. Tubing shall be sufficiently clear to view flow and contents.

3.3.7 Drain. The PVC or latex drain shall be permanently attached to the bottom of the bag. To prevent drip, the end of the drain tube shall be placed in a permanent pocket or socket, located directly above the drain. As an alternate construction, Type II bags may employ push-pull valves with protective caps for facilitating the withdrawal of urine from the main drainage urine container (bag) and the urometer. A permanent pocket or socket would not be required with the alternate construction. A drain shall be included at the bottom of the urometer to facilitate aseptic withdrawal of urine. The material of the urometer shall be sufficiently strong to preclude breakage when the push-pull valve is tested as specified in 4.5.9. The drain tube shall be furnished with an ON - OFF clamp to seal the drain tube. The clamp shall be so designed and manufactured as to avoid danger of puncturing the bag.

3.3.8 Sterility. The contents of the plastic overwrapper shall be sterile when tested as specified in 4.5.1.

3.3.8.1 Certificate of sterility. A certificate of sterility, stating that samples from each sterilizer lot have been tested and found to be sterile, shall be submitted to the procuring activity. The certificate of sterility shall include:

a. Item identification.
b. National Stock Number.
c. Contract or purchase order number.
d. Sterility lot or control number.
e. Date of sterilization.
f. Quantity of product in sterilizer lot.
g. Type of sterilization (gas, radiation, heat, or steam).
h. Form of biologicial indicator (i.e., inoculated paper spore strip or carrier, inoculated product, or inoculated simulated product).
i. Test organism(s) and population density used.
j. Test medium, incubation temperature, incubation time, and dates of tests.
k. Number of test units.
l. Statement that samples comply with sterility tests.
m. Signature or head of laboratory or a designated qualified alternate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:09Z
Check the source to verify that this is the current version before use.

When inspection is made at destination, the certificate of sterility shall be submitted by the contractor to the Directorate of Medical Materiel, ATTN: DPSC-ATQ, Defense Personnel Support Center, 2800 South 20th Street, Philadelphia, PA 19101, and a copy forwarded with each shipment to the consignee. When inspection is made at source, certificate of sterility shall be furnished to the cognizant Government Quality Assurance Representative for submission to the procuring activity.

3.3.8.2 Sterility lot. A sterility lot is that single quantity of product subjected to the same manufacturing operation and simultaneously sterilized in the same sterilizer chamber. Each unit package shall be clearly marked with a sterility lot (control) number and date of sterilization as specified in 5.3.1. Each sterility lot shall be traceable to the pertinent manufacturing lot(s).

3.3.9 Instructions. Each individual unit shall be furnished with instructions for suspending, using and emptying of the system.

3.4 Workmanship. Workmanship shall be first class throughout. The urine collection bags shall be free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual and auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2.1 Classification of inspection. The inspection requirements specified herein are classified as follows:

a. First article inspection.
b. Quality conformance inspection.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:09Z
Check the source to verify that this is the current version before use.

4.1.2.1.1 First article inspection. First article inspection shall be performed on the bag when a first article sample is required (see 3.1). This inspection shall include examination of 4.3 and tests of 4.5.

4.1.2.1.2 Quality conformance inspection. Quality conformance inspection shall include the examination of 4.3 and tests of 4.5.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the materials may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the Item Identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Inspection lot.

4.2.1 Raw materials. An inspection lot for raw material inspection shall be all material produced under the same conditions and controlled to the degree that uniformity of product within the designated lot can be assured. Lot identification of the raw material by the manufacturer is one method of control, as cited above, which is considered acceptable.

4.2.2 End product. An inspection lot for inspection of end product shall consist of all urine bags of the same type produced under essentially the same conditions and offered for inspection at one time.

4.3 Sampling.

4.3.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, and table I. The unit of product for sampling purposes shall be one urine collection bag.

TABLE I. Sampling for examination.

	Inspection level	AQL (defects per hundred units)
For visual examination		
Major defects	II	1.0
Total defects (Major and Minor combined)	II	4.0
For dimensional examination	S-1	2.5

Source: https://assist.dla.mil -- Downloaded: 202640-21T10:09Z
Check the source to verify that this is the current version before use.

4.3.2 For tests. Sampling for tests of end item shall be conducted in accordance with MIL-STD-105 and table II. The unit of product for sampling purposes shall be the urine collection bag.

TABLE II. Sampling for tests.

Component (Unit of product)	Characteristic	Requirement	Test procedure	Inspection level	AQL
PVC Sheeting	Tensile strength	3.2.1	4.5.2	2/	1/
	Elongation	3.2.1	4.5.2	2/	1/
End Item (see 4.3.3)	Submersion	3.3.1 & 3.3.2	4.5.3	S-4	1.0 1/
	Pull	3.3.1 & 3.3.2	4.5.4	S-3	2.5
	Leakage	3.3.3	4.5.5	S-4	1.0 1/
	Indelible	3.3.3.1	4.5.6	S-3	2.5
	Volume	3.3.3.2	4.5.7, 4.5.7.1	S-3	2.5
	Burst	3.3.3	4.5.8	S-3	2.5
	Sterility	3.3.7	4.5.1	3/	1/
	Urometer valve	3.3.6	4.5.9	S-3	1.0

1/ Acceptance number shall be zero for all sample sizes.
2/ Each lot of sheeting shall be tested.
3/ Sample size shall not be less than the minimum specified in 4.5.1.

4.3.3 Test samples. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

4.4 Examination. The urine bags shall be examined to determine compliance with all requirements contained in this specification.

4.4.1 Dimensional examination. The urine bags shall be examined for defects in dimensions. Any dimension not within the tolerances specified herein shall be classified as a defect.

4.4.2 Classification of defects. Examination shall be conducted in accordance with the classification of defects listed in table III. Examination shall not be restricted to the classified possible defects listed below.

TABLE III. Classification of defects.

Major	
101	Bag not free from stains, cracks, or other imperfections.
102	Bag not properly sealed.
103	A component not sealed.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:09Z
Check the source to verify that this is the current version before use.

TABLE III. Classification of defects. - Continued

104	Heat seals not minimum width.
105	Overwrapper not free from cuts or tears, etc.
106	Urometer not integral or adequately attached to main drainage urine container (Type II bags).
107	Urometer, tubing or main drainage urine container not sufficiently clear to view flow and contents.
108	Urometer or main drainage urine container not properly identified.
Minor	
201	Drain control valve not capable of sealing without damaging the tube.
202	Marking on bag not clear.
203	Marking on urometer not clear.
204	Instructions for use not clear or complete.

4.4.3 Packaging inspection. The inspection of the unit packages, packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

4.5 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.5.1 Sterility test. The urine bags shall be tested for sterility in accordance with the method specified in the USP for nonparenteral devices using inoculated product, inoculated simulated product, or paper spore strip (carrier). Product containing a biological indicator shall be marked or labeled so that it clearly differentiates from product intended for distribution. The incubation medium and minimum time in each case shall be soybean casein digest for 7 days for inoculated product, inoculated simulated product, or paper spore strip (carrier). The incubation temperature shall be as indicated in 4.5.1.1. The minimum number of test units, regardless of lot size, shall be 10, except for inoculated carrier which shall be 20 for type I sterilization and 30 for type II sterilization including inoculated carrier and product test units. The inspection level shall be as indicated in table II.

4.5.1.1 Inoculated product or simulated product. For each type of sterilization, the organism used shall be as listed below:

Type of Sterilization	Organism	Incubation Temperature
Steam	B. stearothermophilus	55° - 65°C.
Dry Heat	B. stearothermophilus, or	55° - 65°C.
	B. subtilis var. niger (or globigii)	30° - 35°C.
Gas	B. subtilis var. niger (or globigii)	30° - 35°C.
Radiation	B. pumilus	30° - 35°C.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:09Z
Check the source to verify that this is the current version before use.

The number of viable cells shall be sufficient to provide assurance that the product has been sterilized.

4.5.1.2 Inoculated carrier. Inoculated carriers (paper spore strips) if used, shall be in accordance with MIL-S-36586. Other paper carriers shall be inoculated to be of equivalent resistance. The method of incubation and number of test units shall be as indicated in 4.5.1 and 4.5.1.1. The method of sterilization will control the organism in the spore strip to be tested as shown in 4.5.1.1. NOTE: When an inoculated carrier is chosen, product samples must also be tested as required by the USP.

4.5.1.3 Alternate method. Suppliers conducting sterility tests or designating lots in a manner other than that indicated herein shall submit proof that the alternate method provides equal assurance concerning the sterility of the product. However, all sterility tests shall be conducted using biological indicators. Requests for approval of any alternate method of testing for sterility shall be made to the Contracting Officer, and approval obtained before employment of the alternate method on a government contract.

4.5.1.4 Product testing. Testing for sterility of the product after receipt by the government shall be without a biological indicator and in accordance with the USP.

4.5.2 Tensile strength and elongation. The plastic sheeting shall be tested for tensile strength and elongation in accordance with ASTM D882.

4.5.3 Submersion test. The complete bag in its overwrapper which has been at room ambient temperature for at least 4 hours shall be submerged in hot water (90°F.), at least one-inch below the water surface. The bag shall be gently squeezed with the fingers. Observation of evolution of air bubbles shall be made at each postion of the sample. Bubbles which appear on the surface of the package but are not released or are released at a slowly decreasing rate are not to be construed as indication of failure. A steady stream or recurring succession of bubbles from any surface or seam shall be cause for rejection.

4.5.4 Pull test. The complete system shall be filled to the 2000 cc level and suspended by the catheter adapter for one-hour. The system shall be checked at each joint or connection for separation.

4.5.5 Leakage test. The urine collection bag shall be filled with 2000 cc of water at room temperature and suspended for 96 hours. The bag shall then be inspected for leakage.

4.5.6 Indelible test. Markings shall not fade or disappear when rubbed 10 complete cycles (forward and back) with a water moistened cloth.

4.5.7 Volume test. The collection bag shall be placed in its normal operating position in accordance with the manufacturer's instructions. Water at room temperature shall be added. The accuracy at 1000 cc, 1600 cc and 2000 cc shall be determined.

Source: https://assist.dla.mil -- Downloaded: 2026l0-21T10:09Z
Check the source to verify that this is the currentversion before use.

4.5.7.1 Volume test (Urometer). The Type II urine collection bag shall be placed in its normal operating position in accordance with the manufacturer's instructions. Water at room temperature shall be added to the urometer. The accuracy at 50 cc, 100 cc and 200 cc shall be determined.

4.5.8 Burst test. The urine collection bag shall be inflated with compressed air through the connecting tube. Prior to inflation, any air vents or filters shall be sealed with an adhesive to prevent loss of air. Internal pressure shall be stabilized at 20 mm Hg and then increased through a regular valve at a rate of 1 mm Hg per second. The pressure at which a leak or burst occurs shall be recorded.

4.5.9 Urometer valve. The urometer valve shall be operated forcefully ten times. At the conclusion of the last cycle the urometer shall be visually examined for evidence of breakage.

5. PACKAGING

5.1 Packaging. Packaging shall be level A or C, as speciifed (see 6.2).

5.1.1 Unit of issue. One package containing ten type II or twenty type I collection bags, as specified, constitutes one unit of issue.

5.1.2 Level A.

5.1.2.1 Individual package. Each collection bag shall be packaged in a sealed peel-open or tear type bag capable of maintaining sterility of contents unless damaged or opened. Plastic or metal components of the collection bag shall be protected in a manner to prevent damage to the collection bag or the individual package. The individual package shall not be capable of being resealed.

5.1.2.2 Unit package. Twenty type I bags or ten type II bags individually packaged as specified in 5.1.2.1 shall be packaged in a box conforming to PPP-B-636, class domestic. Closure shall be as specified in the appendix of the box specification.

5.1.3 Level C. Each collection bag shall be individually packaged as specified in 5.1.2.1. Twenty type I or ten type II individually packaged collection bags shall be further packaged in a manner that will afford adequate protection against deterioration and physical damage during shipment from the supply source to the first receiving activity.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.2).

5.2.1 Level A. Two unit packages containing collection bags of one type only shall be packed in a shipping container designed for a type 2 load and conforming to PPP-B-636, class weather-resistant. Closure, strapping and waterproofing shall be as specified in the appendix of the box specification.

Source: https://assist.dla.mil -- Downloaded: 202610-21T10:09Z
Check the source to verify that this is the current version before use.

5.2.2 Level B. Two unit packages containing collection bags of one type only shall be packed in a shipping container designed for a type 2 load and conforming to PPP-B-636, class domestic. Closure shall be as specified in the appendix of the box specification.

5.2.3 Level C. The packaged collection bags shall be packed in shipping containers in a manner that will afford adequate protection against damage during direct shipment from the supply source to the first receiving activity. These packs shall conform to the carrier's rules and regulations.

5.3 Marking.

5.3.1 Individual package. In addition to any commercial markings, each bag shall be marked in accordance with the requirements of the Federal Food, Drug and Cosmetic Act. In addition, each package shall be marked to identify contents. The lot or control number and the following or similar legends shall be shown: "STERILE" and "STERILITY GUARANTEED UNLESS PACKAGE IS DAMAGED OR OPENED".

5.3.2 Unit package. Each unit package shall be marked in accordance with MIL-STD-129 and the requirements of the Federal Food, Drug and Cosmetic Act. The lot or control number, contract or purchase order number and name of contractor shall be shown. The method of sterilization (G-Gas, R-Radiation, C-Cold, D-Dry Heat) shall be shown following the level of packaging and date of packaging. Example: "A 3/78-G". Final sterilization shall be on or after indicated packaging date.

5.3.3 Shipping container. Shipping containers shall be marked as specified in MIL-STD-129. The lot or control number and method of sterilization as specified in 5.3.2 shall be shown.

6. NOTES

6.1 Intended use. Shall be suitable for the collection and measurement of urinary output without disagreeable odor or external contamination.

6.2 Ordering data. Purchasers should select the preferred options permitted herein and include the following information in procurement documents.

a. Title, number and date of this specification.
b. Selection of applicable levels of packaging and packing (see 5.1 and 5.2).
c. Type required (see 1.2).
d. When a first article is required (see 3.1, 4.1.2.1 and 6.3).

6.3 First article. When a first article is required, it shall be tested and approved under the appropriate provisions of the Armed Services Procurement Regulations. The contract or purchase order specifies the quantity, procedure for delivery and other specific instructions for inspection and approval of the first article.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:09Z
Check the source to verify that this is the current version before use.

6.4 This specification covers the following items appearing in the Federal Supply Catalog:

National Stock Number	Item Identification	Type
6530-00-105-8649	BAG, URINE COLLECTION, Plastic, Disposable, 20s	I
6530-01-038-0731	BAG, URINE COLLECTION, Plastic, Disposable, 10s (with urometer)	II

6.5 Changes from previous issue. Asterisks are not used in this revision to identify changes with respect to the previous issue, due to the extensiveness of the changes.

Preparing activity:
DLA - DM

Project No. 6530-1444

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:09Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:09Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/ vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO

DATE

DD FORM 1 JAN 72 1426 REPLACES EDITION OF 1 JAN 68 WHICH MAY BE USED

S/N 0102-014-1802

Source: https://assist.dla.mil -- Downloaded: 202610-21T10:09Z
Check the source to verify that this is the current version before use.

FOLD

POSTAGE AND FEES PAID
DOD- 314

U.S. MAIL

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

Commander
Defense Personnel Support Center
ATTN: DPSC-E
2800 South 20th Street
Philadelphia, Pennsylvania 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:09Z
Check the source to verify that this is the current version before use.