MILITARY SPECIFICATION

BANDAGE, COTTON, ELASTIC

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE AND CLASSIFICATION

1.1 Scope. This specification covers two sizes of cotton elastic adhesive bandages suitable for use as a surgical dressing in areas where wound healing is enhanced by pressure and where drainage is required.

1.2 Classification. The elastic cotton bandages covered by this specification shall be of the following sizes, as specified (see 6.2).

Size 1 - Bandage, Cotton, Elastic, 6 inches by 5-1/2 yards (3 yards unstretched).

Size 2 - Bandage, Cotton, Elastic, 4 inches by 5-1/2 yards (3 yards unstretched).

2. APPLICABLE DOCUMENTS

2.1 Specifications and standards. The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of the specification to the extent specified herein.

SPECIFICATIONS

Federal

PPP-B-566	Boxes, Folding, Paperboard
PPP-B-585	Box, Wood, Wirebound
PPP-B-601	Box, Wood, Cleated-plywood
PPP-B-621	Box, Wood, Nailed and Lock-Corner
PPP-B-636	Boxes, Shipping, Fiberboard
PPP-B-676	Box, Set-up

FSC 6510

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:05Z
Check the source to verify that this is the current version before use.

Military

MIL-L-10547	Liners, Case and Sheet, Overwrap; Waterproof

STANDARDS

Federal

FED-STD-191	Textile Test Methods

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes
MIL-STD-129	Marking for Shipment and Storage

(Copies of specifications, standards and drawings required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Material.

3.1.1 Backing fabric. The backing fabric shall have a minimum of 74 threads per square inch, consisting of a minimum of 42 warp and 32 filler threads. The warp and filler yarns shall be combed or carded peeler. The fabric shall be all cotton, crimped weave with leno edges. As an alternate the fabric may have slit edges.

3.1.2 Adhesive mass. The adhesive mass shall be a smooth homogeneous blend of new material and synthetic rubber, plasticizers, fillers, and resins, free of toxic or skin irritating substances.

3.1.3 Crinoline. The crinoline shall have a thread count of 32 ± 3 in the warp, and 28 ± 3 in the filling.

3.2 Construction. The running length of the fabric shall be with the warp direction. The basic fabric shall be coated on one side with a pressure sensitive adhesive mass so that the minimum weight of the adhesive mass shall be four ounces per square yard. The feather edges of the leno edged fabric shall be free of adhesive. The unstretched length of each bandage shall be not less than three yards with a tolerance of minus five percent. The width of the Size 1 bandage shall be 6 inches ± 1/8 inch and the width of the Size 2 bandage shall be 4 inches ± 1/8 inch. Each bandage shall be evenly

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:05Z
Check the source to verify that this is the current version before use.

and neatly rolled, with the adhesive mass to the outside, on a paperboard tube having a core of 3/4 to 1 inch. The last revolution shall have a crinoline covering. Color of the adhesive mass shall be ivory. Color of the fabric backing shall be natural or flesh colored. The bandage shall be a continuous one piece construction.

3.3 Adhesive quality. Adhesive quality shall be not less than three pounds, when tested as specified in 4.4.4.

3.4 Stretch. Each bandage shall stretch to a minimum of 5-1/2 yards with a tolerance of 5 percent without tearing.

3.5 Delivery. Not more than six months shall have elapsed from the date of manufacture to the date of delivery.

3.6 Workmanship. The adhesive bandage shall be free from imperfections which may affect the appearance or serviceability of the bandage. The adhesive mass shall be homogeneous and uniformly spread. The bandage shall be free from ravels, extraneous matter, and excessive bumps due to imperfections in the fabric.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:05Z
Check the source to verify that this is the current version before use.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the bandages, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the item identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as indicated in Table I. The unit of product for inspection purposes shall be an individual roll of bandage.

Table I. Sampling for examination

Classification	Inspection level	Minimum sample size	AQL (percent defective)
Major A	S-2	1	1.0
Total (Major A and Minor A)	S-2	1	4.0
Major B	S-1	3	* 1/

* Unless otherwise specified, acceptance number shall be zero for all sample sizes.

1/ If any defects are found, the inspection level shall go to S-3 with an AQL of 1.5 percent defective, and the full sample shall be tested.

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and Table II. Unit of product for test purposes shall be an individual roll.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:05Z
Check the source to verify that this is the current version before use.

Table II. Sampling for tests 2/

Characteristic	Test procedure	Inspection level	AQL	Minimum sample size
Stretch.	4.4.1	S-1	1/	3
Weight, adhesive mass.	4.4.2	S-1	1/	3
Adhesive quality.	4.4.4	S-1	1/	3
Width.	4.4.3.1	S-1	1/	3
Length.	4.4.3.2	S-1	1/	3

1/ Acceptance number shall be zero for all sample sizes.
2/ Where possible, the same sample shall be used for the the test procedures are specified.

4.3 Examination.

4.3.1 Examination of bandage. The rolls of bandage shall be examined to determine compliance with all requirements contained in this specification. Examination shall be in accordance with, but shall not be limited to, the defects listed in Table III.

Table III. Classification of defects

Major A.		
	101	Bandage rolls not uniformly and smoothly wound.
	102	Crinoline backing not present.
Major B.		
	151	Adhesive separates from backing upon unwinding.
	152	Roll more than one piece.
	153	Bandage not free from holes, tears, breaks, cuts, or splits.
	154	Adhesive coating on roll not uniform, not free of breaks.
Minor A.		
	201	Bandage roll crushed to the extent of presenting an oval rather than a round profile.
	202	Color not as specified.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:05Z
Check the source to verify that this is the current version before use.

4.3.2 Examination of preparation for delivery. An examination shall be made to determine whether the packaging, packing, and marking comply with the requirements of Section 5. Defects shall be scored as specified in Table IV. Sampling shall be in accordance with MIL-STD-105. The sample unit shall be one container fully prepared for delivery. The lot shall be the number of containers offered for inspection at one time. The inspection level shall be S-2 and the acceptable quality level shall be 4.0 defects per hundred units.

Table IV. Examination of preparation for delivery

Examination	Defect
Containers	Not as specified.
Quantity	Quantity in interior packages and shipping containers not as specified.
Packaging and packing materials	Any component missing or damaged.
Workmanship	Inadequate application of components such as incomplete closure of container flaps, tape improperly applied, or containers distorted.
Markings	Omitted, incorrect, illegible, improper size location, sequence, or method of application.

4.4 Tests.

4.4.1 Stretch. Stretch shall be determined by stretching manually.

4.4.2 Weight of adhesive mass. The weight of a specimen 20 square inches in area of the bandage, as received, shall be determined. The sample shall be placed adhesive mass side down on a glass plate and the backing fabric wetted with petroleum ether, U.S.P. After 30 seconds, the backing shall be stripped from the glass plate leaving the bulk of the adhesive mass on the glass plate. The remaining adhesive mass shall be removed from the backing by two immersions of 30 seconds

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:05Z
Check the source to verify that this is the current version before use.

each into the petroleum ether. The fabric backing shall then be air dried for 30 minutes at room temperature and the weight determined. The weight of the adhesive mass shall be the original weight of the adhesive bandage minus the weight of the fabric backing. The weight of the adhesive mass shall be the average of the results obtained from five specimens tested from five sample rolls.

4.4.3 Dimensions.

4.4.3.1 Width. Measure the width of the bandage at five points evenly spaced throughout the length of the roll. The average of the five measurements shall be the width of the bandage.

4.4.3.2 Length. The bandage shall be rolled out on a flat surface and measured under minimum tension to prevent wrinkles.

4.4.4 Adhesive quality. The adhesive quality of the bandage shall be determined by cutting strips one inch wide and approximately six inches long from each sample unit selected for test. Two square inches (one by two inches) of these strips shall be applied to a bakelite surface by means of a rubber roller under pressure of 30 ounces. The roller shall pass over the strips twice at the rate of 12 inches per minute. The temperature of the bakelite surface and the strips shall be approximately 37° centigrade when tested. The strip shall be pulled (sheared) from the board by means of the machine specified in method 5100 of FED-STD-191, the jaws separating at the rate of 12 + 0.5 inches per minute. The adhesive quality shall be the average of the results obtained from not less than five specimens from each sample roll.

4.4.5 Thread count of cloth backing. The total number of yarns per square inch (warp and filling) shall be counted on five specimens from each sample roll. The average of the results obtained on the five specimens shall be the thread count for the sample roll.

5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be level A or C as specified (see 6.2).

5.1.1 Unit of issue. One package of size 1, containing six (6) bandages or one package of size 2, containing twelve (12) bandages as specified, constitute one unit of issue.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:05Z
Check the source to verify that this is the current version before use.

5.1.1.1 Level A.

5.1.1.1.1 Immediate package. Each bandage shall be securely wrapped in glassine paper or other suitable material acceptable to Food and Drug Administration (FDA) for the product contained therein and shall be of such compatibility that neither the contents nor the container are altered in any way by each other. At the option of the contractor, each bandage may be further packaged in a box of appropriate size conforming to PPP-B-566 or PPP-B-676. Closure of the box shall be adequate to prevent accidental opening.

5.1.1.1.2 Unit package. Six (6) immediate packages for size 1 or twelve (12) immediate packages for size 2 shall be packaged in a box of appropriate size constructed in accordance with PPP-B-636, Class domestic. Closure shall be adequate to prevent accidental opening.

5.1.2 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.2).

5.2.1 Level A. Six (6) units of issue of size 1 and size 2 shall be packed in a box designed for a type 2 load, and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2 or PPP-B-636, type CF, class weather-resistant. Closure and strapping shall be in accordance with the appendix of the applicable box specification.

5.2.1.2 Case liner. Each Level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes. Fiberboard boxes shall be waterproofed in accordance with PPP-B-636.

5.2.2 Level B. Six (6) units of issue of size 1 and size 2 shall be packed in a box designed for a type 2 load, and constructed in accordance with PPP-B-585, class 2, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, type CF, class domestic. Closure shall be as specified in the appendix of the applicable box specification. Special requirements of applicable table shall apply for PPP-B-636 box.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:05Z
Check the source to verify that this is the current version before use.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Immediate container. Each immediate container shall be marked in accordance with the requirements of the Federal Food, Drug and Cosmetic Act. In addition to any commercial markings, the date of manufacture shall also be shown.

5.3.2 Unit package. In addition to any commercial markings, each unit package shall bear the following information:

(a) National Stock Number
(b) The item identification
(c) Date of manufacture
(d) Lot (control) number
(e) Name and address of the manufacturer
(f) The legend: "STORE IN A COOL PLACE"
(g) Contract or purchase order number

5.3.3 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129. Lot (control) number shall be shown. The date of manufacture shall be shown in lieu of the date packed. Marking shall include the legend: "STORE IN A COOL PLACE".

6. NOTES

6.1 Intended use. The elastic cotton bandage covered by this specification is intended for use as a surgical dressing in areas where wound healing is enhanced by pressure and where drainage is required.

6.2 Ordering data. Procurement documents should specify the following:

(a) Title, number and date of this specification.
(b) Size required (see 1.2)
(c) Selection of applicable level of packaging and packing (see 5.1 and 5.2).
(d) National Stock Number (NSN).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:05Z
Check the source to verify that this is the current version before use.

6.3 This specification is intended to cover the following items listed in the Federal Supply Catalog:

National Stock Number	Item Identification
6510-00-935-4093	Bandage, Cotton, Elastic, 6 inches by 5-1/2 yards, 6s
6510-00-935-4094	Bandage, Cotton, Elastic, 4 inches by 5-1/2 yards, 12s

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6510-0601
cs

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:05Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

MIL-B-37163 BANDAGE, COTTON, ELASTIC

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426 REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:05Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, Pa. 19101

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304



HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:05Z
Check the source to verify that this is the current version before use.