MIL-C-37187A
19 August 1976
SUPERSEDING
MIL-C-37187 (DSA-DM)
18 September 1975

MILITARY SPECIFICATION

CHART, RECORDING INSTRUMENT

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

* 1.1 This specification covers one type of chart suitable for use with Automatic Audiometer (see 6.1), calibrated to ISO Standard R389/ANSI Standard S3.6.

2. APPLICABLE DOCUMENTS

2.1 Issues of documents. The following documents of the issue in effect on date of invitation for bids or request for proposals form a part of the specification to the extent specified herein:

SPECIFICATIONS

FEDERAL

TT-I-563	- Ink, Writing, for Fountain and Dip Pens
PPP-B-566	- Boxes, Folding, Paperboard.
PPP-B-636	- Boxes, Shipping, Fiberboard.
PPP-B-676	- Boxes, Setup.

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Headquarters, Defense Personnel Support Center, ATTN: Directorate of Medical Materiel, Code ATT, 2800 South 20th Street, Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 202640-21T10:08Z
Check the source to verify that this is the current version before use.

STANDARDS

MILITARY

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes

MIL-STD-129 - Sampling for Shipment and Storage.

DRAWING

Defense Personnel Support Center (DPSC)

23307 Chart, Recording Instrument, Audiometer

(Copies of specifications, standards, drawings, and publications required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposal shall apply.

American National Standards Institute (ANSI)

S3.6 Specification for Audiometers

International Organization for Standardization (Standards) ISO

R389 Standard Reference Zero for the Calibration of Pure-Tone Audiometers.

(Application for copies should be addressed to the American National Standards Institute, 1430 Broadway, New York, NY 10018.)

Joint Committee on Printing, Congress of the United States

Government Paper Specification Standards - Part 2

(Application for copies should be addressed to the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20402.)

Source: https://assist.dla.mil -- Downloaded: 202640-21T10:08Z
Check the source to verify that this is the currenversion before use.

Technical Association of the Pulp and Paper Industry (TAPPI)

T409 Machine Direction of Paper

(Application for copies should be addressed to the Technical Association of the Pulp and Paper Industry, 1 Dunwoody Park, Atlanta, GA 30341.)

3. REQUIREMENTS

* 3.1 First article. When specified in the procurement document, the contractor shall furnish sample unit(s) for first article inspection and approval (see 4.1.2.1 and 6.2).

3.2 Material.

3.2.1 Paper. The paper shall be white tag, 100% bleached sulphate. No fluorescent (optical) brighteners shall be added to the pulp or paper during manufacture, but minimal trace amounts are permissible. The paper surface shall be free from lint, fuzz, or any particles which will pick, lift, fluff or pile under normal conditions. Paper acidity shall be not less than a pH of 4.4. The paper weight, 500 sheets, 24 x 36 inches, shall be 125 pounds ± 5 percent. The paper thickness shall be 0.0092 inches ± 5 percent. The paper shall be tested as specified in 4.4.2. When tested as specified in 4.4.3, the paper shall have a good printing and writing surface and shall show no feathering.

3.2.2 Ink. The ink shall be black or green lithographic ink of either air drying oxidizing type or flash drying type utilizing a heater for drying. Ink shall bind to paper so that, when thoroughly dry and ready for shipment, there shall be no smudges when the chart is subjected to the rubbing test in 4.4.1.

3.3 Design. The chart shall be lithographed in black or green ink, cut to size and have two punched holes. Chart size shall be 5 inches by 8 inches. Design and construction shall be in accordance with DPSC drawing 23307. The reverse (back) side of the chart shall be lithographed in accordance with the instructions on the drawing.

3.4 Grain direction. The long grain shall run parallel to the 5 inch dimension of the chart.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:08Z
Check the source to verify that this is the current version before use.

3.5 Identification marking. Each chart shall be clearly and legibly marked with the manufacturer's name or registered trademark, and part number, if applicable.

3.6 Workmanship. Workmanship shall be first class throughout. Charts shall be cut square and edges shall be parallel with the lines printed thereon. Charts shall be free from defects which detract from their appearance and which may impair their intended use and serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual and auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

* 4.1.2.1 Classification of inspection. The inspection requirements specified herein are classified as follows:

a. First article inspection.
b. Quality conformance inspection.

* 4.1.2.1.1 First article inspection. First article inspection shall be performed on the charts when a first article sample is required (see 3.1). This inspection shall include examination of 4.3. and tests of 4.4.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:08Z
Check the source to verify that this is the current version before use.

* 4.1.2.1.2 Quality conformance inspection. Quality conformance inspection shall include the examination of 4.3 and tests of 4.4.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the materials may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the Item Identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

* 4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, and table I. The unit of product for sampling purposes shall be the individual chart.

TABLE I. Sampling for examination

	Inspection level	AQL (defects per hundred units)
For visual examination		
Major defects	S-1	1.0
Minor defects.	S-1	4.0
For dimensional examination	25 Charts	1/

1/ Acceptance number shall be zero.

4.2.2 For tests. Sampling for tests shall be conducted in accordance with table II and MIL-STD-105. The acceptance number shall be zero.

TABLE II. Sampling for tests.

Requirement	Test Method	Inspection level
Paper (3.2.1)	4.4.2	(see 4.4.2)
Paper surface (3.2.1)	4.4.3	S-1
Ink (3.2.2)	4.4.1	S-1
Grain (3.4)	4.4.4	S-1

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:08Z
Check the source to verify that this is the current version before use.

4.3. Examination. The charts shall be examined to determine compliance with all requirements contained in this specification.

4.3.1 Classification of defects. Examination shall be conducted in accordance with the classification of defects listed in table III. Examination shall not be restricted to the classified possible defects listed below.

TABLE III. Classification of defects

Category	Defect
Major	
101	Chart values incorrect.
102	Specified printing missing, incomplete, illegible.
Minor	
201	Holes not cleanly punched.
202	Grid line coloring not as specified.
203	Chart not clean.
204	Chart not free of fuzz, lint or loose particles.

4.3.2 Dimensional examination. The charts shall be examined for defects in dimensions. Any dimension not within the tolerances specified herein shall be classified as a defect.

4.4 Tests.

4.4.1 Rubbing test. A clean piece of bond paper shall be manually rubbed four times over different areas of the selected sample. The bond paper and chart shall then be visually examined for any smudge.

4.4.2 Paper. The paper shall be tested for the applicable characteristics for each lot presented for examination. The characteristics tested, number of determinations, acceptance criteria and applicable test procedures shall be as specified in Government Paper Specification Standards-Part 2. For those quality characteristics cited in paragraph 3.2.1, all sample units must pass for the lot to be considered acceptable.

4.4.3 Surface. Lines or characters shall be drawn or written on both sides of the chart with a smooth medium-point dip pen using an ink conforming to TT-I-563. To be considered satisfactory, there shall be no feathering of the inked lines or characters.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:08Z
Check the source to verify that this is the current version before use.

4.4.4 Grain. The grain direction shall be determined on each sample unit in accordance with TAPPI method T409. Report the grain direction in relation to the 5 inch dimension of the card.

4.5 Packaging inspection. The inspection of the unit package, packing and marking for shipment and storage shall be in accordance with quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

5. PACKAGING

5.1 Packaging. Packaging shall be level A or C as specified (see 6.2).

5.1.1 Unit of issue. One package containing five hundred charts, as specified, constitutes one unit of issue.

5.1.2 Level A.

5.1.2.1 Unit package. Five hundred charts shall be packaged in a snug fitting box constructed in accordance with PPP-B-566, PPP-B-676 or PPP-B-636, class domestic. Closure shall be adequate to prevent accidental opening.

5.1.3 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be Level A, B or C as specified (see 6.2).

5.2.1 Level A. Four unit packages shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-636, class weather-resistant. Closure, strapping and waterproofing shall be as specified in the appendix of the box specification.

5.2.2 Level B. Four unit packages shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-636, class domestic. Closure shall be as specified in the appendix of the box specification.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:08Z
Check the source to verify that this is the current version before use.

5.3 Marking.

5.3.1 Unit packages. Each unit package shall be marked in accordance with MIL-STD-129. Contract or purchase order number and name of contractor shall

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Intended use. The chart covered by this specification is intended for use with the von Bekesy type recording Audiometers, Model 1703, manufactured by Grason-Stadler Co., Inc. and supplied under NSN 6515-00-536-5056. Not all Bekesy type recording audiometers will accept the chart design specified herein.

6.2 Ordering data. Purchasers should select the preferred options permitted herein and include the following information in procurement documents.

(a) Title, number and date of this specification.
(b) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).
(c) When a first article is required (see 3.1, 4.1.2.1 and 6.3).

* 6.3 First article. When a first article is required, it shall be tested and approved under the appropriate provisions of 7-104.55 of Armed Services Procurement Regulation. The contract or purchase order specifies the quantity, procedure for delivery and other specific instructions for inspection and approval of the first article.

6.4 This specification covers the following item appearing in the Federal Supply Catalog.

National Stock Number	Item Identification
6515-00-536-5059	CHART, RECORDING INSTRUMENT, Audiometer for ANSI or values, 500s

Source: https://assist.dla.mil - Downloaded: 2026Ø-21T10:08Z
Check the source to verify that this is the currentversion before use.

* 6.5 Changes from previous issue. The margins of this specification are marked with an asterisk to indicate where changes (additions, modifications, corrections, deletions) from the previous issue were made. This was done as a convenience only and the Government assumes no liability whatsoever for any inaccuracies in these notations. Bidders and contractors are cautioned to evaluate the requirements of this document based on the entire content irrespective of the marginal notations and relationship to the last previous issue.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
DSA-DM

Project No. 6515-2000

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:08Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

MIL-C-37187A CHART, RECORDING INSTRUMENT

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426 REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:08Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, PA 19101

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304



OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 202640-21T10:08Z
Check the source to verify that this is the current version before use.