MILITARY SPECIFICATION

DRESSING, FIRST AID, FIELD, INDIVIDUAL TROOP

This amendment forms a part of Military Specification MIL-D-37165, dated 3 June 1975, and is approved for use by all Departments and Agencies of the Department of Defense.

Page 4

3.3 Dressing Design. last line, after " ....compress" add the following: "(Type I only)".

3.5.1 Inner wrapper. Delete instruction "c" entirely and substitute:

"c. Apply correct side of pad to wound.
(for Type I only) 1/

Apply white side of pad to wound.
(for Type II only) 1/"

At the bottom of the page add: " 1/ The phrases (for Type I only) and (for Type II only) shall not be printed on the wrapper."

Page 6

3.5.4 Leakage. Delete entirely and substitute: "The outer wrapper (bag) shall not contain any holes, rips, tears, cuts, or open seams. When tested as specified in 4.4.3 the outer wrapper of the wrapped dressing shall not leak."

Page 11

4.4.3 Leakage test. Delete entirely and substitute: "The leakage test shall be performed in accordance with ASTM D3078-72 except the vacuum shall be 12 inches of Hg. Wire cloth having a mesh of 2x2, or 3x3, or 4x4 may be substituted for the transparent plate in the vacuum chamber."

FSC 6510

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:06Z
Check the source to verify that this is the current version before use.

Page 16

Figure (1) upper right hand corner delete "or on a suitable label". After "(in 1/2 inch red bold face letters on the bandage surface located opposite the absorbent side of the compress)" add "(for Type I only)"

Page 18

Figure (3) upper right hand corner delete "apply white side of pad to wound." and substitute:

"Apply correct side of pad to wound. (for Type I only).

"Apply white side of pad to wound. (for Type II only)."

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6510-0621

Source: https://assist.dla.mil-- Downloaded: 2026-10-21T10:06Z
Check the source to verify that this is the current version before use.