MILITARY SPECIFICATION

INSERT CABINET, MEDICAL INSTRUMENT AND SUPPLY SET CHEST, 5 DRAWER

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers a metal insert of drawers, suitable for insertion in the standard No. 3 and No. 6 medical service aluminum field chests.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

QQ-A-200/8	Aluminum Alloy Bar, Rod, Shapes, Tube, and Wire, Extruded and Structural Shapes, 6061
QQ-A-200/9	Aluminum Alloy Bar, Rod, Shapes, Tubes, and Wire, Extruded, 6063
QQ-A-250/8	Aluminum Alloy 5052, Plate and Sheet
QQ-P-416	Plating, Cadmium (Electrodeposited)
QQ-S-766	Steel Plate, Sheet, And Strip-Corrosion-Resisting
TT-E-527	Enamel, Alkyd, Lustreless
TT-P-1757	Primer Coating, Zinc Chromate, Low Moisture Sensitivity
TT-T-306	Thinner, Synthetic Resin, Enamels
PPP-B-585	Boxes, Wood, Wirebound
PPP-B-601	Boxes, Wood, Cleated Plywood
PPP-B-621	Boxes, Wood, Nailed and Lock Corner
PPP-B-636	Boxes, Shipping, Fiberboard

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:10Z
Check the source to verify that this is the current version before use.

Military

MIL-C-5541	Chemical Conversion Coatings on Aluminum and Aluminum Alloys
MIL-A-8625	Anodic Coatings, for Aluminum and Aluminum Alloys
MIL-L-10547	Liners, Case, and Sheet, Overwrap, Water-Vaporproof or Waterproof, Flexible
MIL-P-15328	Primer (Wash), Pretreatment, Blue (Formula No. 117-B For Metals)

STANDARDS

Federal

FED-STD-151	Metal, Test Methods
FED-STD-595	Color

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes
MIL-STD-129	Marking for Shipment and Storage
MS 16686	Medical Services Symbol

DRAWING

Defense Personnel Support Center (DPSC)

10018	Insert Cabinet, Medical Instrument and Supply Set Chest, 5 Drawer.

(Copies of specifications, standards, and drawing required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Component parts. The metal insert cabinet shall consist of a drawer retaining frame and five drawers. The components of each drawer shall be body and sides, a drawer front, a drawer back, and two female elements of the drawer slides. There shall be partitions in drawers No. 2, 3, 4, and 5, and partitioned trays in drawers Nos. 1 and 2. There shall be a removable locking bar for securing the drawers. The drawer retaining frame assembly shall consist of two similar halves joined by two metal brackets, and ten female elements of the drawer slides, and ten inner parts of the drawer slides.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:10Z
Check the source to verify that this is the current version before use.

3.2 Materials. Except as specified herein, materials shall conform to the applicable portions of the specifications, standards, and drawing (with Bill of Materials) referenced herein, and shall be capable of meeting the requirements shown therein. In addition, the metals shall conform to the chemical requirements of their applicable specifications when tested as specified in Section 4. When materials to be used are not specifically designated, those used shall be of the highest commercial quality, and suitable for the purpose intended.

3.2.1 Aluminum sheet. Aluminum alloy sheet for drawer backs, part No. 3; drawer hole shields, part No. 5; upper plate frame, part No. 22; and frame brackets, part No. 23 shall conform to QQ-A-250/8, and shall be temper H32. Aluminum alloy sheet for drawer body and front, part Nos. 1 and 2; drawer partitions, part Nos. 6 and 7; long, left hand rest and short rest, part Nos. 8 and 9; trays and tray partitions, part Nos. 10, 11, 12, 13, 14, 15, and 16; and locking bar and bracket, part Nos. 17 and 18, shall conform to QQ-A-250/8, and shall be temper H34. As an alternate, aluminum alloy for the locking bar shall be composition 6061-T6 conforming to QQ-A-200/8.

3.2.2 Aluminum channels. Aluminum alloy channels for the frame, part No. 20, shall be 6063-T5 in accordance with QQ-A-200/9.

3.2.3 Steel, corrosion-resisting. Corrosion-resisting steel for the cam, part No. 19, shall conform to QQ-S-766, class 302 or 304.

3.2.4 Primer. Primer shall conform to color Y of TT-P-1757.

3.2.5 Enamel. Enamel shall conform to TT-E-527. Color of the enamel shall approximate 34087 (lusterless olive drab) as shown in FED-STD-595, except that for marking, as specified in 3.5, color shall approximate 37038 (lusterless black) of the same standared (see 6.3).

3.3 Construction. Construction shall conform in all details to the requirements specified herein and as shown on applicable portions of the Drawing (with Bill of Materials). The construction shall be sound and durable in all respects. The assembly shall be welded on jigs capable of maintaining required tolerance on finished inserts. The insert cabinet shall fit into a No. 3 medical field chest conforming to MIL-C-16775 (see 6.4), when tested as specified in 4.3.2.

3.3.1 Welding. Welding shall be durable, sound in every detail and in accordance with latest practices recommended by the American Welding Society. All welds shall be produced by qualified welding operators with satisfactory welding equipment using satisfactory filler metal or electrode where applicable to provide the best possible weld.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:10Z
Check the source to verify that this is the current version before use.

Surfaces to be welded shall be free from grease, oil, oxides or any other foreign matter; any cleaning method which will remove these substances and leave the surface clean is satisfactory. All weld joints shall blend into the adjacent metal in gradual, smooth curves. These welds shall be continuous with no visible voids, pin holes, or cracks. All exposed spot welds shall be smoothed sufficiently so as to match adjacent area. Welds shall not be oxidized.

3.4 Finish. After welding has been completed, but before final assembly, all accessible surfaces of the insert cabinet, except the working surfaces of the drawer slides, shall be finished with enamel as specified herein. Prior to application of the finish all surfaces shall be thoroughly cleaned free from dirt, oil, grease, welding flux and loose scale with solvents, detergents, or other processes which have no deleterious effect on the surfaces. After thorough cleaning and degreasing, surfaces to be enameled shall be prepared for topcoating by one of the following methods:

(1) Chemical treatment in accordance with MIL-C-5541.
(2) Anodic treatment in accordance with MIL-A-8625.
(3) Priming treatment, using primer specified in 3.2.4, thinned and applied by spray in accordance with the directions specified under "Preparation for Delivery" of TT-P-1757, and air-dried for not less than 30 minutes before top-coating.
(4) Preprimining with a wash-primer coat in accordance with MIL-P-15328, and then treated as in (3) above.

Topcoating shall be with one coat of enamel as specified in 3.2.5. The enamel shall be reduced by taking four parts by volume of enamel with one part by volume of thinner in conformance with TT-T-306. The reduced enamel shall be applied by spraying and cured by a heat treatment equivalent to baking at 250° Fahrenheit for not less than 45 minutes. The resultant finish shall be hard, smooth, and highly resistant to abrasion, chipping, cracking, acid, alkali, and solvents.

3.4.1 Film thickness of paint. Contractor shall endeavor to control film thickness so that the thickness of a single dried coat of paint is 0.3 to 0.7 mil for primer and 1.0 to 1.5 mil for enamel.

3.4.2 Steel drawer slides. The female drawer slides, part numbers 4 and 24, shall be cadmium plated in accordance with QQ-P-416, before riveting.

3.5 Identification markings. The insert cabinet shall be permanently and legibly marked with enamel (3.2.5) as specified on appropriate portion of the drawing, and as follows:

a. Centrally located on the front of the second drawer from the top, a one inch Medical Services Symbol as shown for type I of MS16686.
b. The following information shall be centrally located on the upper surface of the bracket (Part No. 23) at the top of the frame:

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:10Z
Check the source to verify that this is the current version before use.

6545-00-663-1548	3/8" High
INSERT CABINET, MEDICAL INSTRUMENT AND SUPPLY SET CHEST, 5 DRAWER	3/8" High
Name or registered trademark of the manufacturer	1/4" High
Contract Number	1/4" High

3.6 Workmanship. Workmanship shall be first class throughout. Frame and drawers shall be square and free from twist. Drawers shall operate smoothly and easily. All burrs shall be removed and all sharp corners broken. Completed insert cabinets shall be free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during performance of the contract and for a period of three (3) years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the aluminum sheet and channels, and corrosion-resisting steel, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the item identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:10Z
Check the source to verify that this is the current version before use.

4.2 Quality conformance inspection. The examinations and tests required to assure conformance of those items or lots of items to be offered for acceptance are classified as specified in tables I and II (see 4.2.1, 4.2.2) and 4.2.3.

4.2.1 Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The inspection levels and acceptable quality levels shall be as indicated in table I. Unit of product for examination purposes shall be one insert cabinet.

Table I. Sampling for Examination

Type of defect	Inspection level	AQL (percent defective)
Major	II	1.0
Minor A	II	4.0
Minor B	S-1	1.0

4.2.2 Examination.

4.2.2.1 Insert cabinet. The insert cabinet shall be examined for defects including, but not limited to, those listed in table II:

Table II. Classification of Defects

Categories	Defects
Critical	None defined.
Major	
101	Unit not complete with all specified component parts.
102	Any component part not free of a crack, tear, rupture, or deformed surface.
103	Any exposed or accessible edge not free of a burr and sharpness.
104	Any weld not sound and smoothly finished.
105	Any drawer does not completely close.
106	Finish not smooth, even, and adherent.
107	Unit identification markings not complete and correct.
Minor A	
201	Identification marking not completely legible.
202	Enameled surfaces not free of excessively chipped or scratched areas.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:10Z
Check the source to verify that this is the current version before use.

Table II. Classification of Defects (cont'd)

Categories	Defects
203	Color of enamel not as required.
204	Movable parts do not operate smoothly.
205	Not clean, free of grease, oil, dirt, other foreign matter.
Minor B	
251	Dimensions not within tolerances shown on the drawing.

4.2.2.2 Preparation for delivery. An examination shall be made to determine whether the packaging, packing, and marking comply with the requirements of Section 5. Defects shall be scored as specified in Table III. Sampling shall be in accordance with MIL-STD-105. The sample unit shall be one container fully prepared for delivery. The lot shall be the number of containers offered for inspection at one time. The inspection level shall be S-2 and the acceptable quality level shall be 4.0 (defects per hundred units).

Table III. Examination for Preparation for Delivery

Examination	Defect
Containers	Not as specified.
Quantity	Quantity in interior packages and shipping containers not as specified.
Packaging and packing materials	Any component missing or damaged.
Workmanship	Inadequate application of components, such as incomplete closure of container flaps, tape improperly applied, or containers distorted.
Markings	Omitted, incorrect, illegible, improper size, location, sequence, or method of application.

4.2.3 Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, Inspection Level S-1, except that the acceptance number shall be zero for all sample sizes. Unit of product for test purposes shall be one insert cabinet.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:10Z
Check the source to verify that this is the current version before use.

4.2.3.1 Sampling for chemical testing of metals. A sufficient amount of each lot of each metal used in the construction of the cases shall be selected for the chemical analysis as specified in 4.3.1.

4.3 Tests. Tests shall be conducted to determine compliance with the requirements of this specification.

4.3.1 Chemical analysis of metals. The samples selected in accordance with 4.2.3.1 shall be analyzed in conformance with methods 111 or 112 of FED-STD-151. In case of dispute, method 111 shall be used for referee analysis.

4.3.2 Fit test. The frame of the insert cabinet shall be positioned in a No. 3 medical field chest (see 6.4) and secured in place with the four locking screws. The chest shall be set on end, and each of the 5 drawers shall be inserted in the frame to their full depth.

5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be level A or C as specified (see 6.2).

5.1.1 Level A.

5.1.1.1 Unit preparation. All exterior finished surfaces shall be protected by wrapping with kraft paper and cushioning with flexible corrugated fiberboard or other suitable materials. Wrapping and cushioning material shall be secured by tying with cotton cord or tape, or by sealing with kraft gummed paper tape. Work shall be accomplished in such a manner as to insure against marring, scratching, or damage to the unit.

5.1.2 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.2).

5.2.1 Level A. Each unit, prepared as specified in 5.1.1.1, shall be overpacked in an exterior container designed for type 3 load, and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B621, class 2 or PPP-B-621, class or PPP-B-636, class weather-resistant. Closure and strapping shall be as specified in the appendix of the applicable box specification.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:10Z
Check the source to verify that this is the current version before use.

5.2.1.2 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to Specification MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes. Fiberboard box shall be waterproofed as specified in the appendix of PPP-B-636.

5.2.2 Level B. Each unit prepared as specified in 5.1.1.1, shall be overpacked in an exterior container designed for a type 3 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure of boxes shall be as specified in the appendix of the applicable box specification.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Intended use. The completed insert cabinets are intended to hold part of the contents of Stock No. 6545-00-299-8592, DENTAL EQUIPMENT SET, PROSTHETIC, FIELD, Part A, or the contents of Stock No. 6545-00-299-8343, DENTAL EQUIPMENT SET, PROSTHETIC, FIELD, Part C, or the contents of Stock No. 6545-00-935-6570 MEDICAL SUPPLY SET, AEROMEDICAL EVACUATION, Part A.

6.2 Ordering data. Procurement documents should specify the following:

(a) Title, number, and date of this specification.
(b) National Stock Number and Item Identification.
(c) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.3 Availability of color chips. Individual chips of the colors specified in 3.2.5 are available from:

Commanding Officer
U.S. Naval Publications and Forms Center
5801 Tabor Avenue
Philadelphia, PA 19120

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:10Z
Check the source to verify that this is the current version before use.

6.4 Sample chest. A 6545-00-914-3480, No. 3 Medical Instrument and Supply Set Field Chest may be obtained on a loan basis upon application to the Contracting Officer and payment of the required deposit.

6.5 The insert cabinet covered by this specification is listed in the Federal Supply Catalog as follows:

National Stock Number	Item Identification
6545-00-663-1548	Insert Cabinet, Medical Instrument and Supply Set Chest, 5 Drawer.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project number: 6545-0021
is

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:10Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

MIL-I-37195 Insert Cabinet, Medical Instrument and Supply Set Chest. 5 drawer

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426

REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:10Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, PA 19101

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304

U.S.MAIL

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:10Z
Check the source to verify that this is the current version before use.