MILITARY SPECIFICATION

INSERT CABINET, MEDICAL INSTRUMENT AND SUPPLY SET CHEST, OPHTHALMIC LENS, 10 DRAWER

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1 SCOPE

1.1 Scope. This specification covers a metal insert of drawers suitable for insertion in the standard No. 3 medical service aluminum field chest. The drawers are partitioned to accommodate various ophthalmic lenses.

2 APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

QQ-A-200/9	Aluminum Alloy Bar, Rod Shapes, Tube and Wire, Extruded, 6063
QQ-A-250/8	Aluminum Alloy 5053, Plate and Sheet
TT-E-527	Enamel, Alkyd, Lustreless
TT-P-1757	Primer Coating, Zinc Chromate, Low Moisture Sensitivity
TT-T-306	Thinner, Synthetic Resin, Enamels
PPP-B-585	Boxes, Wood, Wirebound
PPP-B-601	Boxes, Wood, Cleated Plywood
PPP-B-621	Boxes, Wood, Nailed and Lock Corner
PPP-B-636	Boxes, Shipping, Fiberboard

Military

MIL-C-5541	Chemical Conversion Coatings on Aluminum and Aluminum Alloys
MIL-A-8625	Anodic Coatings, for Aluminum and Aluminum Alloys

FSC 6545

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:10Z
Check the source to verify that this is the current version before use.

MIL-L-10547	Liners, Case, and Sheet, Overwrap, Water-Vaporproof or Waterproof, Flexible
MIL-P-15328	Primer, (Wash), Pretreatment, Blue (Formula No. 117-B for Metals)

STANDARDS

Federal

FED-STD-151	Metal, Test Methods
FED-STD-595	Color

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes
MIL-STD-129	Marking for Shipment and Storage
MS 16686	Medical Services Symbol

DRAWING

Defense Personnel Support Center, (DPSC).

10045	Insert Cabinet, Medical Instrument and Supply Set Chest, Ophthalmic Lens, 10 Drawer

(Copies of specifications, standards, and drawing required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Component parts. The metal insert of drawers comprises a drawer retaining frame assembly and ten identical drawers, a drawer shield, and four screws. The drawer retaining frame assembly consists of two similar two-piece halves (a front and rear half), a top and bottom frame plate, two stiffeners, four tie pieces, twenty female elements of the drawer slides, and twenty inner parts of the drawer slides. The components of each drawer are, a one-piece drawer bottom and sides, a drawer front to which is secured a card holder, a drawer back, two drawer pull covers, two female elements of the drawer slides and five drawer partitions.

3.2 Materials. Except as specified herein, materials shall conform to the applicable portions of the specifications, standards, and drawing (with Bill of Materials) specified herein, and shall be capable of meeting the requirements specified therein. In addition, the metals shall conform to the chemical requirements of their applicable specifications when tested as specified in Section 4. When materials to be used are not specifically designated, those used shall be of the highest commercial quality, and entirely suitable for the purpose intended.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:10Z
Check the source to verify that this is the current version before use.

3.2.1 Aluminum sheet. Aluminum alloy sheet shall be alloy 5052, and shall conform to QQ-A-250/8. Sheet metal parts shall be temper H34 except that shields (drawer and hole), drawer back, and partitions shall be temper H32.

3.2.2 Aluminum channels. Aluminum channels shall be fabricated of alloy 6063 and shall conform to QQ-A-200/9, temper T5.

3.2.3 Primer. The primer shall conform to TT-P-1757, and shall be color Y.

3.2 4 Enamel. Enamel shall conform to TT-E-527 Color of the enamel shall approximate 34087 (lusterless olive drab) as shown in FED-STD-595, except that for marking as specified in 3.5, color shall approximate 37038 (lusterless black) of the same standard (see 6.3)

3 3 Design and construction. The design and construction of the insert cabinet shall be in accordance with the referenced drawing (and Bill of Materials) and as specified herein The assembly shall be welded on jigs capable of maintaining required tolerances on finished inserts The insert cabinet shall fit into a No 3 medical field chest conforming to MIL-C-16775 (see 6.4) when tested as specified in 4.3.2.

3 3.1 Welding. Welding shall be durable, sound in every detail and in accordance with latest practices recommended by the American Welding Society All welds shall be produced by qualified welding operators with satisfactory welding equipment using satisfactory filler metal or electrode where applicable to provide the best possible weld Surfaces to be welded shall be free from grease, oil, oxides or any other foreign matter; any cleaning method which will remove these substances and leave the surface clean is satisfactory All weld joints shall blend into the adjacent metal in gradual, smooth curves These welds shall be continuous with no visible voids, pin holes, or cracks. All exposed spot welds shall be smoothed sufficiently so as to match adjacent area. Welds shall not be oxidized.

3 4 Finish After welding has been completed, but before final assembly and attachment of cardholders to drawer fronts, all accessible surfaces of the insert cabinet, except the working surfaces of the drawer slides shall be finished with enamel as specified herein. Prior to the application of the finish, all surfaces shall be thoroughly cleaned free from dirt, oil, grease, welding flux and loose scale with solvents, detergents, or other processes which have no deleterious effect on the surfaces After through cleaning and degreasing, surfaces to be enameled shall be prepared for topcoating by one of the following methods

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:10Z
Check the source to verify that this is the current version before use.

(1) Chemical treatment in accordance with MIL-C-5541.
(2) Anodic treatment in accordance with MIL-A-8625.
(3) Priming treatment, using primer as specified in 3.2 3, thinned and applied by spray in accordance with directions specified under "Preparation for Delivery" of TT-P-1757, and air-dried for not less than 30 minutes before topcoating.
(4) Prepriming with a wash-primer coat in accordance with MIL-P-15328, and then treated as in (3) above

Topcoating shall be with one coat of enamel as specified in 3.2.4. The enamel shall be reduced by taking four parts by volume of enamel with one part by volume of thinner in conformance with TT-T-306 The reduced enamel shall be applied by spraying and cured by a heat treatment equivalent to baking at 250° Fahrenheit for not less than 45 minutes. The resultant finish shall be hard, smooth, and highly resistant to abrasion, chipping, cracking, acid, alkali, and solvents.

Cardholders shall be attached to the drawer fronts and topcoating shall then be accomplished with one coat of enamel, as specified above.

3 4.1 Film thickness of paint. Contractor shall endeavor to control film thickness so that the thickness of a single dried coat of paint is 0.3 to 0.7 mil for primer and 1 0 to 1.5 ml for enamel

3.5 Identification markings. The insert cabinets shall be permanently and legibly marked with enamel as specified in 3.2 4, as follows

Suitably located on the front of the top drawer, there shall be a 1-inch-Medical Services Symbol as shown for type I in Standard MS16686

The following information shall be centrally located on the upper surface of the top frame plate

6545-00-299-9503	3/8 high
INSERT CABINET, MEDICAL INSTRUMENT AND SUPPLY SET CHEST, OPHTHALMIC LENS, 10 DRAWER	3/8 high
Name or registered trademark of the manufacturer.	1/4 high
Contract Number	1/4 high

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:10Z
Check the source to verify that this is the current version before use.

3.6 Workmanship. Workmanship shall be first class throughout. Frame and drawers shall be square and free from twist. Drawers shall operate smoothly and easily All burrs shall be removed and all sharp corners broken. Completed insert cabinets shall be free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements

4 1 1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate

4.1 2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4 1 3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the aluminum sheet and channels, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the item identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4 2 Quality conformance inspection The examinations and tests required to assure conformance of those items or lots of items to be offered for acceptance are classified as specified in Tables I and II (see 4.2.1, 4.2.2) and 4.2 3

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:10Z
Check the source to verify that this is the current version before use.

4.2 1 Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The inspection levels and acceptable quality levels shall be as indicated in table I Unit of product for examination purposes shall be one insert cabinet

Table I. Sampling for Examination

Type of defect	Inspection level	AQL (percent defective)
Major	II	1.0
Minor A	II	4.0
Minor B	S-1	1.0

4 2.2 Examination

4.2.2.1 Insert cabinet The insert cabinet shall be examined for defects including, but not limited to, those listed in table II.

Table II Classification of Defects

Categories	Defects
Critical	None defined
Major	
101	Unit not complete with all specified component parts.
102	Any component part not free of a crack, tear, rupture, or deformed surface
103	Any exposed or accessible edge not free of a burr and sharpness
104	Any weld not sound and smoothly finished
105	Any drawer does not completely close.
106	Finish not smooth even, and adherent.
107	Unit identification markings not complete and correct
Minor A	
201	Identification marking not completely legible
202	Enamelled surfaces not free of excessively chipped or scratched areas
203	Color of enamel not as required
204	Moveable parts do not operate smoothly.
205	Not clean, free of grease, oil, dirt, or other foreign matter
Minor B	
251	Dimensions not with tolerances shown on the drawing

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:10Z
Check the source to verify that this is the current version before use.

4.2.2.2 Preparation for delivery. An examination shall be made to determine whether the packaging, packing, and marking comply with the requirements of Section 5 Defects shall be scored as specified in Table III Sampling shall be in accordance with MIL-STD-105 The sample unit shall be one container fully prepared for delivery The lot shall be the number of containers offered for inspection at one time The inspection level shall be S-2 and the acceptable quality level shall be 4 0 (defects per hundred units)

Table III Examination for Preparation for Delivery

Examination	Defect
Containers	Not as specified
Quantity	Quantity in interior packages and shipping containers not as specified
Packaging and packing materials	Any component missing or damaged
Workmanship	Inadequate application of components, such as incomplete closure of container flaps, tape improperly applied, or containers distorted
Markings	Omitted, incorrect, illegible, improper size, location, sequence, or method of application

4.2.3 Sampling for tests Sampling for tests shall be conducted in accordance with MIL-STD-105, Inspection Level S-1, except that the acceptance number shall be zero for all sample sizes Unit of product for test purposes shall be one insert cabinet

4.2.3.1 Sampling for chemical testing of metals. A sufficient amount of each lot of each metal used in the construction of the cabinets shall be selected for the chemical analysis as specified in 4.3.1

4.3 Tests Tests shall be conducted to determine compliance with the requirements of this specification

4.3.1 Chemical analysis of metals The samples selected in accordance with 4.2.3.1 shall be analyzed in conformance with methods 111 or 112 of FED-STD-151 In case of dispute, method 111 shall be used for referee analysis

4.3.2 Fit test The frame of the insert cabinet shall be positioned in a No 3 medical field chest (see 6.4) and secured in place with the four locking screws The chest shall be set on end, and each of the ten drawers shall be inserted in the frame to their full depth

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:10Z
Check the source to verify that this is the current version before use.

5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be level A or C as specified (see 6.2)

5.1.1 Level A

5.1.1.1 Unit preparation. All exterior finished surfaces shall be protected by wrapping with kraft paper and cushioning with flexible corrugated fiberboard or other suitable materials. Wrapping and cushioning material shall be secured by tying with cotton cord or tape, or by sealing with kraft gummed paper tape, Work shall be accomplished in such a manner as to insure against marring scratching, or damage to the unit

5.1.2 Level C Units shall be packaged in standard commercial containers of the size and kind commonly used which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.2)

5.2.1 Level A. Each unit, prepared as specified in 5.1.1.1 shall be overpacked in an exterior container designed for type 3 load, and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type, PPP-B-621, class 2 or PPP-B-636, class weather-resistant. Closure and strapping shall be in accordance with the appendix of the applicable box specification

5.2.1.2 Case liner Each level A wood box shall be lined with a waterproof case liner conforming to Specification MIL-L-10547 Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes. Fiberboard box shall be waterproof in accordance with the appendix of PPP-B-636.

5.2.2 Level B. Each unit prepared as specified in 5.1.1.1 shall be overpacked in an exterior container designed for a type 3 load and constructed in accordance with PPP-B-585 class 1, style 3 PPP-B-601, domestic type PPP-B-621, class 1, or PPP-B-636, class domestic Closure shall be as specified in the appendix of the applicable box specification

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers at the lowest rate to point of delivery called for in the contract or purchase order

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:10Z
Check the source to verify that this is the current version before use.

5.3 Marking

5.3.1 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Intended use. The completed insert cabinets are intended to hold various standard ophthalmic lenses having a maximum diameter of 2 9/16 inches.

6.2 Ordering data. Procurement documents should specify the following

a. Title, number, and date of this specification.
b. National Stock Number and Item Identification.
c. Selection of applicable levels of packaging and packing (see 5.1 and 5.2)

6.3 Availability of color clips. Individual chips of the colors specified in 3.2.4 are available from.

Commissioner
Federal Supply Service
General Services Administration
Standardization Division, Paints Branch
Washington, DC 20406

6.4 Sample chest. A 6545-00-914-3480 No. 3 Medical Instrument and Supply Set Field Chest may be obtained on a loan basis upon application to the Contracting Officer and payment of the required deposit

6.5 The insert cabinet covered by this specification is listed in the Federal Supply Catalog as follows

National Stock Number	Item Identification
6545-00-299-9503	Insert Cabinet, Medical Instrument and Supply Set Chest, Ophthalmic Lens, 10 Drawer

Custodians
Army - MD
Navy - MS
Air Force - 03

Preparing activity
Defense Supply Agency - DM

Project Number: 6545-0022
is

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:10Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No 22 R255

INSTRUCTIONS The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay or will otherwise enhance use of the document DoD contractors, government activities, or manufacturers, vendors who are prospective suppliers of the product are invited to submit comments to the government Fold on lines on reverse side, staple in corner and send to preparing activity Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements Attach any pertinent data which may be of use in improving this document If there are additional papers, attach to form and place both in an envelope addressed to preparing activity

DOCUMENT IDENTIFIER AND TITLE
MIL-I-37196 INSERT CABINET, MEDICAL INSTRUMENT AND SUPPLY SET CHEST, OPHTHALMIC LENS, TO DRAWER

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1 HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A GIVE PARAGRAPH NUMBER AND WORDING

B RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2 COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3 IS THE DOCUMENT RESTRICTIVE?

☐ YES ☐ NO *(If "Yes", in what way?)*

4 REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO

DATE

DD FORM 1 JAN 72 1426

REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:10Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, PA 19101

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD 304

U.S. MAIL

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:10Z
Check the source to verify that this is the current version before use.