MIL-I-37200A
27 March 1975
SUPERSEDING
MIL-I-37200
13 January 1975

MILITARY SPECIFICATION

INHALER, MEDICAMENT, PLASTIC

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers a Chromolyn Sodium Inhaler Device suitable for use with Chromolyn Sodium Capsules, 20 mg per Capsule (see 3.5).

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposals, form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

PPP-B-566	Boxes, Folding, Paperboard Boxes.
PPP-B-636	Box, Shipping, Fiberboard.
PPP-B-676	Box, Set-up.

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

DRAWING

Defense Personnel Support Center (DPSC)

No. 24355	INHALER, MEDICAMENT, PLASTIC

(Copies of specifications, standards and drawing required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:10Z
Check the source to verify that this is the current version before use.

2.2 Other publications. The following document forms a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposals shall apply.

U.S. DEPARTMENT OF HEALTH, EDUCATION, AND WELFARE
FOOD AND DRUG ADMINISTRATION

Federal, Food, Drug and Cosmetic Act and
General Regulations for its Enforcement.

(Application for copies should be addressed to the Food and Drug Administration, U.S. Department of Health, Education and Welfare, Washington, DC 10016.)

3. REQUIREMENTS

3.1 Material.

3.1.1 Plastic components. Plastic components of the inhaler shall be fabricated of 11 Amino-Undecanoic Acid (Nylon 11).

3.1.2 Metal components. The washers on the capsule piercing device shall be fabricated from aluminum. All other metal components of the inhaler shall be fabricated from stainless steel.

3.2 Style and design. The inhaler shall consist of a body with a sliding sleeve, a capsule piercing device, a mouth piece and permanently affixed spindle, and a propeller. The sliding sleeve on the body shall cause the capsule to be pierced, when operated in accordance with the instructions. The propeller component shall securely accommodate the capsules. All components shall be securely assembled. Shall be in accordance with Defense Personnel Support Center Drawing (DPSC) No. 24355.

3.2.1 Instructions for use. Each inhaler shall be provided with an instruction sheet containing ample illustrations and clear, concise, instructions for proper use of the inhaler. The instructions shall include a statement that the sliding sleeve shall be operated to pierce each capsule only once.

3.3 Weight loss on inhalation. Chromolyn Sodium capsule content shall have an average weight loss of 90 percent, minimum, when tested according to the method of 4.4.1. No more than 5 percent of the samples tested shall have less than 75% of their weight expelled.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:10Z
Check the source to verify that this is the current version before use.

3.4 Compliance with the Federal Food, Drug, and Cosmetic Act. The items furnished under this specification must meet the requirements of the Federal Food, Drug, and Cosmetic Act. The contractor must submit evidence to the purchasing agency prior to award of contract that the items he proposes to furnish under this specification meet the requirements of the Federal Food, Drug and Cosmetic Act, including an effective New Drug Application on file with the Food and Drug Administration.

3.5 Government furnished material. The Government shall furnish (see 6.4) the contractor the Chromolyn Sodium capsules (NSN 6505-00-133-5421) needed to perform the weight loss on inhalation test specified in 4.4.1.

3.6 Identification marking. The inhaler shall be permanently and legibly marked with the name or registered trademark of the contractor.

3.7 Workmanship. Workmanship shall be first class throughout. The inhalers shall be free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:10Z
Check the source to verify that this is the current version before use.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The inspection levels and acceptable quality levels shall be as indicated in Table I. The unit of product for examination shall be one inhaler.

Table I. Sampling for Examination

Type of defect	Inspection level	AQL (percent defective)
Major	II	1.0
Minor	II	2.5

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105. Inspection level shall be S-3, with an AQL of 0.65.

4.3 Examination. The inhalers shall be examined to determine compliance with all requirements contained in this specification. Examination shall be conducted in accordance with MIL-STD-105 and the following classification of defects. Examination shall include, but not limited to, the defects listed in Table III. Any inhalers in the sample containing one or more defects shall be rejected, and if the number of defective inhalers exceeds the acceptance number for the sample, the lot represented by the sample shall be rejected.

Table III. Classification of Defects

Categories	Defects
Critical	None defined.
Major	
101	Style and design not as specified.
102	Sliding sleeve does not operate smoothly.
103	Inhaler not clean and free of loose particles.
104	Inhaler not free of flash.
105	Washer not securely attached to piercing device.
Minor	
201	Exterior surface of mouthpiece not smooth.
202	Instructions missing, incomplete or illegible.
203	Contractor's name or registered trademark missing, incomplete or illegible.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:10Z
Check the source to verify that this is the current version before use.

4.4 Tests.

4.4.1 Inhalation test. An apparatus shall be constructed so that the inhaler is attached by its mouthpiece to an aspirational action. The volume of air aspirated through the loaded inhaler (containing one capsule of Chromolyn Sodium (see 3.5) shall be at the rate of 60 liters/minute, 1-1/2 seconds duration, and repeated 4 times within 10 seconds. The weight of the capsule contents shall be recorded to the nearest tenth of milligram both before and after the aspirational exposure and the weight loss percentages calculated.

4.5 Examination for preparation for delivery. An examination shall be made to determine whether the packaging, packing, and marking comply with the requirements of Section 5. Defects shall be scored as specified in Table IV. Sampling shall be in accordance with MIL-STD-105. The sample unit shall be one container fully prepared for delivery. The lot shall be the number of containers offered for inspection at one time. The inspection level shall be S-2 and the acceptable quality level shall be 4.0 defects per hundred units.

Table IV. Examination of Preparation for Delivery

Examination	Defect
Containers	Not as specified.
Quantity	Quantity in interior packages and shipping containers not as specified.
Packaging and packing materials	Any component missing or damaged.
Workmanship	Inadequate application of components, such as incomplete closure of container flaps, tape improperly applied, or containers distorted.
Markings	Omitted, incorrect, illegible, improper size, location, sequence, or method of application.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:10Z
Check the source to verify that this is the current version before use.

5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be level A or C as specified (see 6.2).

5.1.1 Unit of issue. Each inhaler, as specified, constitutes one unit of issue.

5.1.2 Level A.

5.1.2.1 Unit package. Each inhaler shall be thoroughly cleaned, wrapped in tissue paper and packaged, along with instructions for use, in a snug fitting box constructed in accordance with PPP-B-566 or PPP-B-676. Closure of the box shall be secured. As an alternate, a plastic or metal container may be used in lieu of the tissue wrap.

5.1.2.2 Intermediate package. Twelve unit packages shall be packaged in a snug fitting box constructed in accordance with PPP-B-566 or PPP-B-676. Closure shall be adequate to prevent accidental opening.

5.1.3 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.2).

5.2.1 Level A. Twenty four intermediate packages (288 units) shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-636, class weather-resistant. Closure, waterproofing and strapping shall be as specified in the appendix of the box specification.

5.2.2 Level B. Twenty four intermediate packages (288 units) shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-636, type CF, class domestic and having a minimum bursting strength of 200 pounds. Closure of the exterior container shall be as specified in the box specification.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:10Z
Check the source to verify that this is the current version before use.

5.3 Marking.

5.3.1 Interior packages. In addition to any commercial markings supplied by the contractor, each unit package and each intermediate package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Intended use. The inhalers are intended for use with capsules of Cromolyn Sodium powder as an adjunct in the treatment of respiratory diseases.

6.2 Ordering data. Purchasers should exercise any desired options offered herein, and the procurement documents should specify the following:

(a) Title, number, and date of this specification.
(b) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.3 This specification is intended to cover the following item appearing in the Federal Supply Catalog:

National Stock No.	Item Identification
6515-00-133-5426	INHALER, MEDICAMENT, Plastic

It does not include all types or sizes as indicated by the title nor all types which are commercially available.

6.4 Government furnished material. The Chromolyn Sodium Capsules required to perform the testing in paragraph 3.5 will be furnished by the Defense Personnel Support Center, 2800 S. 20th Street, Philadelphia, PA 19101 within 30 days of date of contract.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6515-1936

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:10Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:10Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE: MIL-I-37200A INHALER, MEDICAMENT, PLASTIC

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?
   A. GIVE PARAGRAPH NUMBER AND WORDING.

   B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

   ☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426 REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

☆ GPO 1974 - 603-680/6126 2 1

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:10Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, PA 19101

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
D-D - 304

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:10Z
Check the source to verify that this is the current version before use.