MILITARY SPECIFICATION

PUMP AND FILLER, PRESCRIPTION BOTTLE

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 This specification covers one type of pump and bottle filler for use in a MUST (Medical Unit Self-Contained Transportable) pharmacy.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitatio for bids or request for proposal, form a part of the specification to the extent specified herein.

SPECIFICATIONS

Federal

PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.

Military

MIL-M-7298	Manuals, Technical: Commercial Equipment.
MIL-P-116	Preservation-packaging, Methods of
MIL-T-152	Treatment, Moisture and Fungus-resistant, Of Communications, Electronic, and Associated Electrical Equipment.

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

DRAWING

Defense Personnel Support Center (DPSC)

24251 Pump and filler, Prescription Bottle

FSC 6530

Source: https://assist.dla.mil -- Downloaded: 202640-21T10:10Z
Check the source to verify that this is the currenversion before use.

(Copies of Military specifications, standards, and drawings required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Materials. Materials shall be in accordance with Defense Personnel Support Center (DPSC) Drawing 24251, and as specified herein. Material shall be of the highest quality throughout and suitable in all respects for the intended purpose.

3.1.1 Moisture and fungus resistance. The motor shall be treated in accordance with the requirements specified in MIL-T-152.

3.2 Construction. Unit shall be constructed in accordance with DPSC Drawing 24251 and as specified herein. The unit shall consist of a pump and motor assembly and a bottle filler assembly. The total weight shall not exceed 40 pounds.

3.2.1 Pump and motor assembly. The pump and motor assembly shall consist of a single stage, rotary, sliding vane pump and motor mounted on a base. The pump and motor shall be interconnected by means of a flexible coupling. The pump shall be equipped with sealed, permanently lubricated ball bearings and carbon graphite vanes. An inlet filter and outlet muffler for the pump shall be provided. A three conductor, type SO power cord, a minimum 8 feet in length, terminated in a twist type three prong male plug capable of mating with the 60 Hertz MUST receptacle, shall be connected to the motor switch. A strain relief shall be provided. The base shall be equipped with four rubber feet capable of preventing the pump and motor from creeping during operation. The motor shall be rated at 1/4 horsepower, 1725 revolutions per minute and shall operate on 115 volt, 60 Hertz (Hz) electric power. The overall dimensions of this assembly shall not exceed 7 by 17 by 10 inches.

3.2.2 Bottle filler assembly. The bottle filler assembly shall consist of a portable three spout filler handle, two overflow pieces, one short and one long as shown on the drawing, and a minimum of 25 feet of polyvinyl chloride tubing. The three spout handle shall have double tube spouts equipped with adjustable collars and rubber sealing gaskets. The tubing shall be suitable for interconnecting the filler handle, the overflow pieces, and the pump. The tubing shall have a 3/8 inch inside diameter, 1/8 inch wall thickness and shall have a sufficient durometer hardness to preclude collapse during use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:10Z
Check the source to verify that this is the current version before use.

3.3 Performance. When operating on 115 volt 60 Hz electric power the unit shall develop the following minimum output characteristics:

Pressure - 10 pounds per square inch gage (psig)

Vacuum - 20 inches of mercury-continuous
24 inches of mercury-intermittent

Rate - 3.5 cubic feet per minute at 0 psig (without filters or tubing connected).

The unit shall operate satisfactorily after exposure to the following environmental conditions:

High temperature - 155 degrees Farenheit (F.) for 4 hours.

Low temperature --minus 65 degrees F. for 12 hours.

There shall be no damage or degradation of materials following exposure to the above conditions when tested as specified in 4.4.3.

The unit shall be capable of pumping a minimum of 6 liters per minute against a 30 inch head when tested as specified in 4.4.4.

3.4 Safety. The unit shall be constructed so as to afford protection from electrical shock or mechanical injury. The unit shall be capable of withstanding without breakdown the dielectric test specified in 4.4.2.

3.5 Finish. All metal parts shall be finished to protect against corros All parts shall be thoroughly cleaned and degreased prior to finishing. The motor, pump, and base shall be finished in an enamel in accordance with best commercial practice. These parts shall be of the same color. Hardware shal be plated.

3.6 Identification marking. Each unit shall be permanently and legibly marked with the following information:

Name or registered trademark of the manufacturer.

Manufacturer's model number.

Electrical rating.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:10Z
Check the source to verify that this is the current version before use.

3.7 Service Data. The contractor shall furnish, with each unit, two copies each of a technical manual which contains complete instructions for installation, operation, and maintenance and lists of component parts. The manual shall comply with the requirements of MIL-M-7298. As soon as practicable after award of the contract or purchase order, the contractor shall furnish to the Contracting Officer, via the cognizant Government inspector, two preliminary copies of his proposed technical manual for review and comment by the Directorate of Medical Materiel, Defense Personnel Support Center. The preliminary manuals shall be submitted sufficiently early to permit adequate review by the Directorate of Medical Materiel, Defense Personnel Support Center (based on a maximum period of 30 days for such review) and to allow incorporation by the contractor of any required revisions and corrections in the final manual without delaying delivery under the contract, or purchase order. In addition, the Contractor shall furnish 15 copies of the approved technical manual to the Contracting Officer for distribution. Should the supplier have furnished acceptable manuals for the identical item within two years preceding the date of the contract and propose to furnish manuals identical with those previously accepted, then the requirements for review of drafts and submission of 15 additional copies may be waived upon request to the Contracting Officer. Waiver of submission of drafts and submission of 15 additional copies shall be at the option of the Government. Service data shall include specific instructions for dismantling the pump assembly for cleaning and disinfecting purposes.

3.8 Workmanship. Workmanship shall be first class throughout. Units shall be free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual, and auditory, investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 202610-21T10:10Z
Check the source to verify that this is the current version before use.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the component or material may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The Certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the Item Identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, and Table I.

Table I Sampling for Examination

	Inspection level	AQL (defects per hundred units)
For visual examination		
Major defects	S-4	1.0
Minor defects	S-4	4.0
For dimensional examination	S-1	2.5

4.2.2 For tests. Sampling for tests shall be conducted in accordance with Table II and MIL-STD-105. Acceptance number shall be zero for tests.

Table II Sampling for Tests

Test	Inspection level
Operation test	100%
Dielectric test	100%
Environmental test	*
Pumping Test	S-1

* To be conducted on the first article only.

Source: https://assist.dla.mil -- Downloaded: 202610-21T10:10Z
Check the source to verify that this is the currentversion before use.

4.3 Examination. The units shall be examined to determine compliance with all requirements contained in this specification.

4.3.1 Classification of defects. Examination shall be conducted in accordance with the following classification of defects, as shown in Table III.

Table III Classification of Defects

Categories	Defects*
Major	
101	Unit not complete with all components.
102	Pump and motor not securely fastened to base.
103	Pump not free of cracks and breaks.
104	Line cord not held securely by strain relief.
Minor	
201	Identification marking not complete, not present.
202	Unit not clean, not free of dirt, grease, oil.
203	Unit not free of sharp edges or burs.
204	Painted surfaces not free of runs, sags, or thin areas.

*Examination shall not be limited to the classified possible defects listed above.

4.4 Tests. Tests shall be performed to determine compliance with requirements.

4.4.1 Operation test. The unit shall be connected to a 115 volt, 60 Hz power source and operated for a minimum of 30 minutes. Pressure and vacuum measurements shall be made during operation.

4.4.2 Dielectric test. Immediately following the operation test, while the motor is still in a heated condition, the unit shall be subjected to a dielectric test. Voltage shall be applied between live and dead metal parts and gradually raised from 0 to 1000 volts and maintained at that value for one minute.

4.4.3 Environmental test. The entire unit shall be placed in a test chamber where the temperature shall be gradually increased to 155 degrees F. and maintained at that temperature for four hours. At the end of this period the temperature shall be gradually reduced to minus 65 degrees F. and maintained at this tempeature for 12 to 20 hours. The equipment shall then be gradually returned to room temperature,examined and operated per 4.4.1.

4.4.4 Pumping test. The unit shall be completely set-up and connected to a 115 volt, 60 Hertz electric power source. The vacuum manifold of the bottle filler assembly shall be attached to the vacuum inlet of the pump through a trap bottle using the transparent vinyl tubing. One end of a six foot length of vinyl tubing shall be attached to the liquid manifold of the bottle filler assembly and the other end shall be placed in a 5 gallon carboy containing at

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:10Z
Check the source to verify that this is the current version before use.

least 3 gallons of water. The level of the water shall be approximately 30 inches below the bottle filler assembly. Three prescription bottles, not less than 16 ounce capacity each, shall be filled and the time shall be measured. The bottle filling shall be repeated 5 times. The pumping rate shall be determined for each test and the average obtained for the five tests.

4.5 Examination for preparation for delivery. An examination shall be made to determine whether the packaging, packing, and marking comply with the requirements of Section 5. Defects shall be scored as specified in Table IV. Sampling shall be in accordance with MIL-STD-105. The sample unit shall be one container fully prepared for delivery. The lot shall be the number of containers offered for inspection at one time. The inspection level shall be S-2 and the acceptable quality level shall be 4.0 defects per hundred units.

Table IV Examination for Preparation for Delivery

Examination	Defect
Containers	Not as specified
Quantity	Quantity in interior packages and shipping containers not as specified.
Packaging and packing materials	Any component missing or damaged.
Workmanship	Inadequate application of components, such as incomplete closure of container flaps, tape improperly applied, or containers distorted.
Markings	Omitted, incorrect, illegible, improper size, location, sequence, or method of application.

5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be level A or C as specified (see 6.2).

5.1.1 Level A.

5.1.1.1 Unit package. Pump and filler shall be packaged in accordance with MIL-P-116, method IC-2. Preservative shall not be applied.

5.1.2 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.2).

5.2.1 Level A. Two (2) units of issue shall be packed in a box designed for a type 2 load and constructed in accordance with PPP-B-601, overseas type, or PPP-B-621, class 2. Closure and strapping shall be as specified in the appendix of the applicable box specification.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:10Z
Check the source to verify that this is the current version before use.

5.2.2 Level B. Two (2) units of issue shall be packed in a box designed for a type 2 load and constructed in accordance with PPP-B-601, domestic type, or PPP-B-621, class 1. Closure shall be as specified in the appendix of the applicable box specification.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Unit packages. Each unit package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.
(b) Selection of applicable levels of packaging and packing (see 5.1 and 5.2)

6.2 This specification covers the following item appearing in the Federal Supply Catalog:

National Stock Number	Item Identification
6530-00-480-6007	PUMP AND FILLER, PRESCRIPTION BOTTLE, 115 Volt, 60 Hz, AC

6.3 This specification does not cover all types, classes, grades, or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the types which are normally procured to meet requirements.

Custodians:
Army - Md
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DSA-DM

Project No. 6530-1335

☆ U. S. GOVERNMENT PRINTING OFFICE: 1975 -603-766/1085

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:10Z
Check the source to verify that this is the current version before use.