MILITARY SPECIFICATION

TEST KIT, PREGNANCY DETERMINATION

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 This specification covers a Test Kit, Pregnancy Determination 100 tests. (See 6.2)

2. APPLICABLE DOCUMENTS

2.1 The following documents, of the issue in effect on date of invitation for bids or request for proposals, form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

PPP-C-186	Containers, Packaging and Packing for Drugs, Chemicals, and Pharmaceuticals. (See 6.3)

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

(Copies of specifications and standards required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposals shall apply.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:08Z
Check the source to verify that this is the current version before use.

AMERICAN PHARMACEUTICAL ASSOCIATION PUBLICATION

National Formulary.

(Application for copies should be addressed to the Mack Publishing Company, Easton, Pa. 18042.)

U.S. DEPARTMENT OF HEALTH, EDUCATION, AND WELFARE, FOOD AND DRUG ADMINISTRATION

Federal Food, Drug, and Cosmetic Act and Regulations Promulgated Thereunder.

(Application for copies should be addressed to the Food and Drug Administration, Washington, D.C. 20204.)

U.S. PHARMACOPEIAL CONVENTION, INC.

Pharmacopeia of the United States.

(Application for copies should be addressed to the Mack Publishing Company, Easton, Pa. 18042.)

3. REQUIREMENTS

3.1 Material. Shall be as specified hereafter for the specified component. The kit shall consist of the following components necessary for performing a minimum of 100 slide tests for pregnancy.

3.1.1 Components. Each kit shall contain the following components:

(a) One or two bottles of Antiserum Reagent.
(b) One or two bottles of Antigen Reagent.
(c) A minimum of 100 disposable pipettes in a suitable container or holder.
(d) Two reusable rubber bulbs.
(e) A minimum of 100 wooden/plastic stirrers or applicator sticks in a suitable container or holder.
(f) Glass slide.

3.1.2 Suitability. Shall be suitable for pregnancy determination discernible at two minutes when performed in accordance with the manufacturer's directions.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:08Z
Check the source to verify that this is the current version before use.

3.2 Description of components.

3.2.1 Antiserum Reagent. Each kit shall be furnished with one or two bottles with a dropper cap containing sufficient amount of Anti-Human Chorionic Gonadotropic Serum (Rabbit) for a minimum of 100 tests. Shall contain 0.1% Sodium Azide or 0.01% Thimerosal as a preservative.

3.2.2 Antigen Reagent. Each kit shall be furnished with one or two bottles with a dropper cap containing sufficient amount of HCG coated or chemically bonded latex suspension for a minimum of 100 tests. Shall contain 0.1% Sodium Azide or 0.01% Thimerosal as a preservative.

3.2.3 Glass slide. Shall be a glass slide with three circles of suitable size for use in the determination of test results.

3.3 Sensitivity test. Shall be capable of detecting 2.5 I.U. Human Chorionic Gonadotropin (HCG) or less when tested as specified in 4.5.2.

3.4 Package insert. The package insert shall explain the utilization of the kit and the interpretation of the results.

3.5 Ingredients. All ingredients entering into the reagents and controls shall be of U.S.P. or N.F. quality or, if not included in either of these compendia, the ingredients shall be of the highest pharmaceutical quality.

3.6 Potency period. The reagents shall have an expiration dating period (potency period) of 12 months from the date of manufacture. The date of manufacture is defined as the date of filling the final container, or the date of manufacturer's or contractor's final quality approval of the final containers, which shall be not later than one month after the date of filling.

3.6.1 Delivery. Not less than 10 months of the expiration dating period (potency period) shall remain at the time of delivery to the Government.

3.7 Workmanship. Workmanship shall be first class throughout. The kit and its components shall be free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Supplier responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:08Z
Check the source to verify that this is the current version before use.

Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 No company supplying any ingredient(s) to the contractor will be considered an acceptable facility for the performance of any inspection requirements specified herein.

4.2 Lot. For purpose of this specification, a lot, batch, or control is that single, uniform, and homogeneous quantity of reagents, produced from one formulation, subjected to the same compounding and manufacturing operation, and filled into final containers.

4.2.1 Inspection. Inspection shall be conducted, as necessary, to determine compliance with the requirements of this specification.

4.3 Sampling shall be conducted in accordance with MIL-STD-105.

Table I Sampling

For visual examinations	Inspection level	AQL (percent defective)	Unit of product
Major	S-2	1.0	Each (Unit package)
Minor	II	2.5	Each (Unit package)

4.3.1 For end item testing. Noncompliance with any of the following is cause for rejection.

Source: https://assist.dla.mil -- Downloaded: 202610-21T10:08Z
Check the source to verify that this is the current version before use.

Table II End Item Testing

Characteristic	Requirement Paragraph	Test Paragraph	Inspection level*	Minimum Sample
Components	3.1.1	--	S-2	5
Sensitivity test	3.3	4.5.2	S-2	5
Reaction test	---	4.5.1	S-2	5

* Where possible, the same sample shall be used for two or more tests.

4.4.4 Examination. The Pregnancy Test Kit shall be examined to determine compliance with all requirements of this specification. Nonconformance will be permitted to the extent indicated in 4.3 and 4.3.1.

4.4.1 Classification of defects. Examination will be conducted in accordance with the following classification of defects:

Categories and defects*

Major

101 Anti-Human Chorionic Gonadotropic Serum sufficient to perform 100 tests not included with each kit.
102 Human Chorionic Gonadotropic coated or chemically bonded latex suspension sufficient to perform 100 tests not included with each kit.
103 A minimum of 100 disposable pipettes not included with each kit.
104 Two reusable rubber bulbs not included with each kit.
105 A minimum of 100 stirrers or applicator sticks not included with each kit.
106 A glass slide not included with each kit.

Minor

201 Exterior of kit not free from contents or foreign matter.

* Inspection is not restricted to the classified possible defects listed above.

Source: https://assist.dla.mil -- Downloaded: 202610-21T10:08Z
Check the source to verify that this is the current version before use.

Table III. Examination for Preparation for Delivery

Examination	Defect
Containers	Not as specified.
Quantity	Quantity in interior packages and shipping containers not as specified.
Packaging and Packing Materials	Any component missing or damaged.
Workmanship	Inadequate application of components, such as incomplete closure of container flaps, tape improperly applied, or containers distorted.
Markings	Omitted, incorrect, illegible, improper size, location, sequence, or method of application.

5. Preparation for Delivery

5.1 Preparation for Delivery shall be in accordance with all applicable requirements of PPP-C-186, and as specified herein.

5.2 Packaging. Packaging shall be level A or C as specified (see 6.1).

5.2.1 Unit of issue. One each test kit as specified, constitutes one unit of issue.

5.2.2 Level A.

5.2.2.1 Immediate containers. Immediate containers for the reagents shall be commercially available immediate containers considered acceptable by the Food and Drug Administration (FDA) for the product contained and shall be of such compatibility that neither the contents nor the container are altered in any way by each other.

5.2.2.2 Unit package. Each unit shall be packaged as specified in 5.2.5 of PPP-C-186.

Source: https://assist.dla.mil -- Downloaded: 202610-21T10:08Z
Check the source to verify that this is the current version before use.

5.4.2 Intermediate package. Each intermediate package shall be marked as specified in 5.5.3 of PPP-C-186. Type I shelf life marking as specified in MIL-STD-129 shall be shown. Marking shall include the legends:

"STORE BETWEEN 2° - 8°C. (35° - 46° F.)"

and

"DO NOT PERMIT TO FREEZE"

5.4.3 Exterior container. Exterior container shall be marked as specified in 5.5.3 of PPP-C-186. Type I shelf life marking as specified in MIL-STD-129 shall be shown. Marking shall include the legends:

"STORE BETWEEN 2° - 8°C. (35° - 46° F.)"

and

"DO NOT PERMIT TO FREEZE"

Note: The unrefrigerated shipping time for this item shall be not more than 72 hours.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.
(b) the National Stock Number.
(c) Selection of applicable levels of packaging and packing (See 5.2 and 5.3).

6.2 This specification covers the following item(s):

National Stock No.	Item Identification
6505-00-137-6055	TEST KIT, PREGNANCY DETERMINATION, 100 Test

6.3 All references to PPP-C-186 refer to Interim Federal Specification PPP-C-00186B which is mandatory for use until superseded by a revision.

Source: https://assist.dla.mil -- Downloaded: 202640-21T10:08Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, Pa. 19101

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304

U.S. MAIL

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:08Z
Check the source to verify that this is the current version before use.