INCH-POUND

A-A-54853
16 April 1993
SUPERSEDING
A-A-54213
31 July 1990
and
A-A-53584
29 July 1988

COMMERCIAL ITEM DESCRIPTION

MOULAGE SET, WAR WOUNDS

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for MIL-M-36278B, which is cancelled.

This Commercial Item Description covers a set of wound and other anatomical training moulages, suitable for the demonstration of procedures and techniques encountered in administering first aid.

Salient Characteristics:

Materials and construction.

Moulages and straps. Moulages and straps shall be made from a vinyl plastisol, or other suitable material, and shall be formulated to produce a finished product which is resilient, has good dimensional stability, retains its finish without chipping, and is capable of withstanding repeated use.

Adaptors. Adaptors shall be injection-molded vinyl.

Reservoirs, pumps, and tubing assembly. The assembly shall consist of a bag (reservoir), upper section tubing with a sliding occluding device, a pump bulb with valving, and lower section tubing with moulage connector at bottom end. As an alternate to the pump and reservoir specified, an integral reservoir and vinyl hand pump with attached tubing and connector shall be acceptable. Eight complete assemblies are required for each set of 19 moulages.

AMSC N/A FSC 6910

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

The bag shall be fabricated from 12 mil vinyl sheet conforming to Federal Standard L-P-375, Type II, Class II. The bag shall have a volume capacity of, at least, 16 fluid ounces. There shall be a flap at the top of the bag with 2 elongated eyelets for use to secure the bag by means of a one inch wide belt, strapping, or a 5/16 inch diameter hook. The eyelets shall be appropriately positioned to facilitate suspension of the bag by belting/strapping which is secured to a mock victim. The bag shall have a water tight valve port subassembly near the top for filling purpose. The filling valve cap shall not be removeable from the bag.

The tubing shall be clear commercial quality vinyl, approximately 1/8 inch ID. Both sections shall be 30 inches (plus or minus 2 inches) long. The combined length of upper tubing section, pump bulb, and lower tubing section shall be at least 5 feet long.

The sliding occluding device shall be made from either aluminum or plastic material and shall not be removable from the upper section tubing. The device shall be capable of being set to occlude the tubing passageway, and set to allow full passage of fluid through the tubing.

The pump bulb shall be fabricated from clear vinyl. The pump bulb shall be appropriately shaped for hand squeezing/pumping and shall be capable of pumping one fluid ounce per squeeze. The pump bulb shall have a valving arrangement to suck simulated blood from the bag and expel it from the moulage set connector. The valving shall allow gravity feed of fluid from the bag out of the moulage connector. The pump shall be capable of manual operation by the wearer of the moulage. Either an arrow shall be on the bulb to indicate direction of flow or the inlet shall be marked "R" and the outlet marked "M".

The moulage connector shall be approximately 3/16 inch OD by approximately 3/32 inch ID by approximately 1/2 inch long (beyond tubing end). The connector shall be a straight shank which may have ring depressions to assist securing a sealed connection to a moulage. The connector shall be fabricated of brass or anodized aluminum.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

The reservoir and pump assembly shall be unitized. When the bag is secured to a hook by a single eyelet, the bag shall be capable of supporting 5 pounds, attached to the bottom of the bag without causing damage. The connection points of tubing to bag, pump bulb, and moulage connector shall be capable of withstanding a tension force of 5 pounds without causing evidence of damage/joint failure.

The filling port salient characteristics shall be: The uncapped port opening shall be approximately 1 inch across to facilitate refilling without spilling. The cap shall be capable of being dislodged from a closed condition when subjected to an opening force not to exceed 3 pounds. The cap shall maintain a sealed closed condition when a full (of water) bag is dropped (freefall) from a height of 4 feet onto a steel or concrete surface. When the valve cap is open, its connection to the bag shall be capable of withstanding a tension force of 5 pounds between the valve cap and either one of the eyelets. (See note #1 at end of specifications).

Carrying case. The top, bottom and sides of the carrying case shall be a minimum of 3/8 inch clear beech, birch, or maple lumber, dressed both-sides (D2s), kiln dried, with a moisture content not to exceed 12 percent. Front and back panels of carrying case shall be 3-ply D2s, 1/8 inch, birch or maple plywood. The plywood shall be thoroughly kiln dried. Wood shall be free of cracks and knotholes, and any knots present shall be sound.

Corner guards and hinge. Corner guards and hinge for the carrying case shall be made of brass, or brass plated steel. The hinge shall be piano type.

Safety hasp. Safety hasps shall be of brass or steel, and nickel or chrome plated. Safety hasps shall be 3-1/4 inches long by 1-1/2 inches wide.

Handle and bumpers. Handle shall be made of leather and bumpers of rubber.

Design and construction of components. The moulage set, including all components, shall be produced in accordance with Government furnished sample and the requirements of this specification. Except as specified herein, all materials, details, colors and dimensions shall match those found in the Government furnished prototype.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Moulages. Moulages shall be of the bleeding wound, sucking wound, non-bleeding wound, and hypodermic technique types. Each moulage shall be a life-size model of the wound or affected area of the anatomy. Material thickness at any point of the moulage shall be 1/16 inch plus 1/16 minus 1/32 inch, except for wound areas, where it may be less. The blood vessels shall be permanently and securely attached along their entire lengths to the underside of the moulages. Moulages shall show no evidence of damage to the structure or finish, when tested as specified herein.

Bleeding wound type. Each bleeding moulage shall be capable of bleeding by means of simulated blood pumped through simulated blood vessels, located in the moulages. The following shall be bleeding moulages:

a. Amputation of Leg.
b. Compound Fracture of the Femur.
c. Compound Fracture of the Humerus.
d. Compound Fracture of the Lower Leg.
e. Gunshot Wound of the Hand (palm).
f. Laceration of the Scalp.
g. Shrapnel Wound of the Abdominal Wall with Protruding Intestine.
h. Shrapnel Wound of the Lower Jaw with Partial Loss of Jaw.

Sucking wound type. Sucking wound of the chest shall be a blood-suffused moulage which, when air is pumped to the sponge in the bullet hole, produces a frothing action, and when the bulb is released, produces a sucking sound. This moulage shall have its own pump. This pump shall force the contents of the bulb (air and simulated blood) to and from the sponge in the sucking wound of the chest. The sucking wound of the chest bulb adapter shall have a single outlet, and be so designed that a plastic tube will connect the wound to the pump.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Non-bleeding wound type. The following shall be non-bleeding moulages:

a. Phosphorus Burn of the Hand.
b. Second and Third Degree Burns of the Forearm.
c. Atomic Burn of Chest.
d. Atomic Burn of Back.
e. Atomic Burn of Hand.
f. Atomic Burn of Face.
g. Frostbite of Foot.
h. Trench Foot.
i. Face in Shock.

Hyodermic technique type.

Moulage, Upper Arm Skin, Hypodermic Insertion, Adult. This moulage shall be used for hypodermic needle insertion training. It shall be a simulated upper arm, molded plastic with skin-like texture, resiliency and resistance to puncture. Shall be secured by hooks and two permanently fixed retaining straps, located to comfortably secure the moulage to the upper arm. Straps shall be adjustable to four positions of tension by suitable quick release fasteners, such as snaps.

Simulated blood powder. Each set of 19 moulages shall contain five heat-sealed vinyl bags of simulated blood powder. The simulated blood powder shall consist of a uniformly mixed dry powder containing the ingredients listed below to simulate the appearance and consistency of human blood when mixed with water. The dye used in the formulation of the powder shall be certified by the Food and Drug Administration for at least external cosmetic use. The formulation of the powder shall be as follows:

Dye and coloring powder	56.0 percent
Methyl cellulose	33.0 percent
Sodium Benzoate	11.0 percent

A tolerance of plus or minus 3 percent shall apply to the formulation. Excluding wool, dyes selected shall wash off all clothing, skin and moulages.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Thirty-seven (37) grams plus or minus 10 percent of the powder shall be contained in a heat sealed 8-gage vinyl film plastic bag. Each bag shall have the following information legibly printed on one side.

POWDER, SIMULATED BLOOD

For Preparation of Simulated Blood

1. To one quart of hot water, add the package contents very slowly, stirring/agitating constantly.
2. Stir/agitate until the powder appears to be wetted out and without lumps.
3. Add the remaining water at room temperature or colder to make one gallon.
4. Agitate thoroughly until solution is complete.

This formula will wash off skin, moulages and clothing except wool. NOT INTENDED FOR HUMAN CONSUMPTION. DO NOT PUT IN MOUTH.

The following information shall also be included on each bag:

Name and address of manufacturer
Date of manufacture

Adjustable straps. Each moulage, except those of the hand and foot, shall be provided with an adjustable strap. One end of each strap shall be permanently attached to the moulage. A sufficient number of straps shall be provided to effect a good fit by means of fasteners. The straps shall be adjustable so as to fit a male, standard coat sizes from 38 to 42. Each strap shall have non-corrosive metal fasteners.

Hand moulages. Moulages simulating injuries of the hand shall readily fit a male who wears a standard size 9 glove.

Face moulages. Moulages simulating injuries or condition of the face shall readily fit a male adult's face.

Foot moulages. Moulages simulating injuries of the foot shall readily fit a male adult who wears a size 10 shoe.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Carrying case. The carrying case for each complete set shall be adequately compartmented for each moulage, for the pumps, reservoirs, tubes, and blood powder. Each moulage compartment shall be provided with vinyl plastic straps and snap fasteners for securing the moulages therein. Carrying case may be smaller in size than the prototype, providing all other conditions specified in this specification are met. The frame shall be of a lock-corner construction, and glued. Wire brads may be used along with the glue. The separators shall be of rabbeted or dado construction, and glued. Wire brads may be used along with the glue. The leather handle shall be secured to the case with appropriate hardware using roundhead bolts, and nuts. Corner guards shall be nailed or screwed on all corners, when the case is closed. There shall be two nails or screws on side surfaces contacted by the corner guard. The piano hinge shall be attached to the carrying case by means of flathead screws. Hinge shall have holes spaced on 2-inch centers. Holes shall be countersunk so that the head of screws, when inserted, will be flush with the hinge. Safety hasps shall be secured with roundhead screws. Front and back panels shall be glued. Wire brads may be used along with the glue. Six rubber bumpers shall be placed at the bottom of the case, three on each half of the case. All wood shall be sanded smooth.

Instruction sheets. The contractor shall furnish two of the following instruction sheets in each case. Instructions shall be laminated between two sheets of clear plastic as follows:

INSTRUCTIONS FOR THE CARE AND USE OF MOULAGE SET,

WAR WOUNDS, NSN 6910-00-540-6378

One set of 19 moulages appropriately labeled and compartmented.
Eight sets of reservoir, pump, tubing assemblies for use with the bleeding moulages.
One pump, tubing, sponge assembly for use with sucking wound of chest.
Five packets of powdered, simulated blood, each packet sufficient to make one gallon.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

NOTE: Simulated blood can be removed from most clothing by good laundering but may leave permanent stains on woolen clothing.

Assembly and operation.

1. Dip all bloodline connectors in water before using. Fill plastic reservoir through the 1 inch filling port. Cap and squeeze reservoir until all air is exhausted through bloodline port. Insert bloodline and connect to pump port "R". Connect pump port "M" to moulage. (If bulb is marked with an arrow indicating direction of flow, "R" is the tail and "M" is the front end of the arrow.)

2. Squeeze bulb several times until blood fills bulb and bloodline to moulage. Gentle, intermittent squeezing of the bulb will simulate arterial bleeding.

3. The sucking wound chest moulage does not have a reservoir. It is connected directly to the single bulb which has been initially filled with sufficient blood to moisten the wound. The bulb is then operated in a position to pump a mixture of blood and air creating a sucking and frothing effect.

Handling and cleaning.

1. Where necessary, roll, DO NOT FOLD, for storing.
2. Wash with a mild soap and warm water; do not scrub or use detergents or abrasive cleaners.
3. Clean reservoirs, bloodlines, and moulage capillary lines IMMEDIATELY AFTER USING. CAUTION: DO NOT USE BOILING WATER UNDER ANY CIRCUMSTANCE.

NOTE: The contents of this set are designed to perform under ambient conditions of temperature ranging for 20°F to 130°F, and relative humidity ranging up to 95 percent.

(See other side for replacement items).

The reverse side of the instruction sheet shall have the following:

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Replacement parts.

NSN 6910-00-729-6161	SIMULATED BLOOD, POWDER, 37 Gm, 5s
NSN 6910-00-809-6713	SUCKING WOUND OF CHEST WITH PUMP
NSN 6910-00-809-6724	ATOMIC BURN OF CHEST
NSN 6910-00-809-6725	LACERATION OF FOREHEAD (SCALP)
NSN 6910-00-809-6760	AMPUTATION OF LEG
NSN 6910-00-809-6761	COMPOUND FRACTURE OF LOWER LEG
NSN 6910-00-809-6773	ATOMIC BURN OF BACK
NSN 6910-00-809-6799	ATOMIC BURN OF HAND
NSN 6910-00-809-6801	PHOSPHORUS BURN OF HAND
NSN 6910-00-809-6802	SECOND AND THIRD DEGREE BURNS OF THE FOREARM
NSN 6910-00-809-6807	SHRAPNEL WOUND OF LOWER JAW
NSN 6910-00-809-6808	SHRAPNEL WOUND OF ABDOMINAL WALL
NSN 6910-00-809-6816	COMPOUND FRACTURE OF HUMERUS
NSN 6910-00-809-6819	COMPOUND FRACTURE OF FEMUR
NSN 6910-00-809-6828	ATOMIC BURN OF FACE
NSN 6910-00-809-6835	FROSTBITE OF FOOT
NSN 6910-00-809-6880	TRENCH FOOT
NSN 6910-00-809-6881	GUNSHOT WOUND OF HAND
NSN 6910-00-809-6882	FACE IN SHOCK
NSN 6910-00-892-1942	MOULAGE, UPPER ARM SKIN, HYPODERMIC INSERTION, Adult
NSN 6910-01-129-3194	RESERVOIR ASSEMBLY, SIMULATED BLOOD

The reverse side shall also include the following information:

Manufacturer name or registered trademark
Contract or purchase order number
Date of manufacture

Finish.

Moulages. The moulages shall be finished with a matte coating which will best simulate human skin, and which will form a durable bond with the material used in making the moulages. All colors used shall be those which accurately simulate the color of the particular skin, flesh, bone and tissue of the particular area of the human anatomy. Moulages in the Government-loaned sample moulage set shall be used as a guide for matching colors. The finish shall be durable and shall not chip, soften, crack, become sticky or tacky, when tested as specified herein.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Carrying case. The carrying case shall be painted on the outside only. The finish shall be an air drying, semi-gloss olive drab paint applied over a white primer coat. The coating system shall be smooth, evenly applied, and cured to a hard, durable finish. (See note #2 at end of specification).

Identification marking.

Moulages. All moulages shall be labeled as to type of wound and affected anatomical part of the body. Label shall be either of durable material firmly attached to the underside of the moulage or strap, or "hot-stamped" on the strap. As an alternate, labeling shall be applied to the underside of moulage or strap in an ink which is indelible when applied to vinyl plastic.

Carrying case. In the center of both side panels, the carrying case shall be marked with the letters "U.S." a minimum of 2 inches in height. Centered below the letters "U.S." shall appear the Federal stock number and below that, the item identification in block letters not less than 1 inch high. Numerals and letters shall be yellow or white. Compartments shall be stamped with the name of the moulages they contain.

Government-loaned property. The following property shall be loaned by the Government:

a. Moulage molds (one set of 19 molds)

b. War Wound Moulage Set (set of 19 moulages) with carrying case, reservoir-pump-tubing assemblies, and powdered simulated blood).

NOTE: The contractor shall be responsible for the care, maintenance and return of all Government property in the same condition as when originally supplied by the procuring activity. Return of all Government-loaned property shall be at the expense of the contractor. If the procurement document citing this specification is not for a complete set of nineteen (19) moulages, NSN: 6910-00-540-6378, but for separate component(s) only, the Government may elect to provide only the mold(s) and the component(s) being procured.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Workmanship. The reservoir assembly shall be free from defects which detract from its appearance or impair its serviceability. Particular attention shall be given to detail in finishing and coloring the moulage surfaces. For the moulages the following shall apply:

a. Surfaces shall be smoothly finished. Any pits and blisters present on outer surfaces, thickenings and carbonized vinyl present in or on the surfaces of the moulage shall be no greater than those present in the Government furnished prototype.

b. Moulages shall be free of holes and cuts. Shapes shall be clean without ragged edges attached thereto. Identification markings shall be uniformly dark and completely legible.

c. The top and bottom of the closed case shall be in complete alignment. Hinges and corner guards shall be securely fastened to the case. Corner guard nails and other sharp metal points shall not project above the inner surface.

Unit. Set (SE). One package consisting of a carrying case containing all items specified herein. Those items are: one set of nineteen moulages, eight bags of simulated blood powder, eight sets of reservoir- pump- tubing assemblies for use with the bleeding moulages, and one pump- tubing- sponge assembly for use with sucking wound of chest moulage.

Replacements when ordered separately, shall have the following units of issue:

Unit. Each (EA). One amputation of leg moulage, as specified, constitutes one unit.

Unit. Each (EA). One compound fracture of the femur moulage, as specified, constitutes one unit.

Unit. Each (EA). One compound fracture of the humerus moulage, as specified, constitutes one unit.

Unit. Each (EA). One compound fracture of the lower leg moulage, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Unit. Each (EA). One gunshot wound of the hand (palm) moulage, as specified, constitutes one unit.

Unit. Each (EA). One laceration of the forehead (scalp) moulage, as specified, constitutes one unit.

Unit. Each (EA). One shrapnel wound of the abdominal wall with protruding intestines moulage, as specified, constitutes one unit.

Unit. Each (EA). One shrapnel wound of the lower jaw with partial loss of jaw moulage, as specified, constitutes one unit.

Unit. Each (EA). One sucking wound of the chest moulage (and pump), as specified, constitutes one unit.

Unit. Each (EA). One phosphorus burn of the hand moulage, as specified, constitutes one unit.

Unit. Each (EA). One second and third degree burn of the forearm moulage, as specified, constitutes one unit.

Unit. Each (EA). One atomic burn of chest moulage, as specified, constitutes one unit.

Unit. Each (EA). One atomic burn of back moulage, as specified, constitutes one unit.

Unit. Each (EA). One atomic burn of hand moulage, as specified, constitutes one unit.

Unit. Each (EA). One atomic burn of face moulage, as specified, constitutes one unit.

Unit. Each (EA). One frostbite of foot moulage, as specified, constitutes one unit.

Unit. Each (EA). One trench foot moulage, as specified, constitutes one unit.

Unit. Each (EA). One face in shock moulage, as specified, constitutes one unit.

Unit. Package (PG). Simulated blood, powder. One package containing 5 bags of powdered blood, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Unit. Package (PG). Reservoir assembly, simulated blood. One package containing eight reservoir assemblies, as specified, constitutes one unit.

Unit. Each (EA). One hypodermic needle insertion moulage, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Certificates of quality. When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

Sampling. Sampling for examination and testing, including dimensional tests, shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of II. The unit of inspection for examination and testing purposes shall be one moulage set.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Tests. Tests shall be conducted to determine compliance with specification requirements. Tests shall include, but shall not be limited to the following:

Drop test. Each moulage, including pump, full reservoir, and tubing, shall be dropped from a height of 48 inches onto a concrete floor.

Resistance to temperature and humidity tests.

Resistance to cold. Moulage shall be subjected to a temperature of 20°F plus or minus 5°F for a period of not less than two hours.

Resistance to heat and humidity. Moulage shall be subjected to a temperature of 130°F plus or minus 5°F, and a minimum of 95 percent relative humidity for a period of not less than four hours.

Solubility of simulated blood powder test. One gallon of simulated blood shall be prepared as instructed herein. The resulting solution shall be free of foreign and undissolved matter which would clog the apparatus during use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

NOTES: 1. Halkey Roberts' filling port, Aqua Drain Part No. 650-AD valve meets the specified characteristics for the cap of the Reservoir, Pump, and Tubing Assembly.

2. The color of the government furnished prototype carrying case is No. X24087 of FED STD 595 and is applied to the outside only conforming to class A of TT-E-491. A white primer coat conforming to TT-P-659 is applied initially.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

3. This specification covers the following National Stock Numbers:

NSN	Item name
NSN 6910-00-540-6378	MOULAGE SET, WAR WOUND, SET OF 19
NSN 6910-00-729-6161	SIMULATED BLOOD, POWDER, 37 Gm, 5s
NSN 6910-00-809-6713	SUCKING WOUND OF CHEST
NSN 6910-00-809-6724	ATOMIC BURN OF CHEST
NSN 6910-00-809-6725	LACERATION OF FOREHEAD
NSN 6910-00-809-6760	AMPUTATION OF LEG
NSN 6910-00-809-6761	COMPOUND FRACTURE OF LOWER LEG
NSN 6910-00-809-6773	ATOMIC BURN OF BACK
NSN 6910-00-809-6799	ATOMIC BURN OF HAND
NSN 6910-00-809-6801	PHOSPHORUS BURN OF HAND
NSN 6910-00-809-6802	BURN OF FOREARM
NSN 6910-00-809-6807	SHRAPNEL WOUND OF LOWER JAW
NSN 6910-00-809-6808	SHRAPNEL WOUND OF ABDOMINAL WALL
NSN 6910-00-809-6816	COMPOUND FRACTURE OF HUMERUS
NSN 6910-00-809-6819	COMPOUND FRACTURE OF FEMUR
NSN 6910-00-809-6828	ATOMIC BURN OF FACE
NSN 6910-00-809-6835	FROSTBITE OF FOOT
NSN 6910-00-809-6880	TRENCH FOOT
NSN 6910-00-809-6881	GUNSHOT WOUND OF HAND
NSN 6910-00-809-6882	FACE IN SHOCK
NSN 6910-00-892-1942	MOULAGE, UPPER ARM SKIN, HYPODERMIC INSERTION, Adult
NSN 6910-01-129-3194	RESERVOIR ASSEMBLY, SIMULATED BLOOD, 8s

MILITARY INTERESTS:

Custodians:

Air Force - 03
Army - MD
Navy - MS

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQAS
USPHS
VA - OSS

Project Number: 6910-0113
Location: MOULGC1D.DOC

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.