INCH-POUND

A-A-54886
30 June 1993

COMMERCIAL ITEM DESCRIPTION

ENVELOPE, RADIOGRAPHIC FILM MOUNT, DENTAL

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Military Specification MIL-E-37816, which is cancelled.

This Commercial Item Description covers envelopes for dental radiographic film mounts, suitable for filing and preserving dental X-ray film.

Salient characteristics:

Classification. The envelopes shall be of the following sizes, as specified:

Size 1	-	5-1/2 by 9-1/4 inches
Size 2	-	3-1/8 by 5-1/2 inches

Material. The envelopes shall be fabricated of smooth, light-colored heavyweight No. 1 Kraft paper, composed of 100 percent unbleached sulfate pulp. No groundwood pulp shall be present in the finished paper. The fiber content shall be determined by Technical Association of the Pulp and Paper Industry (TAPPI) Standard T401. Recycled paper meeting specification requirements is acceptable.

AMSC N/A FSC 7530

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Physical prop sical properties of the paper shall be as s he size indicated.

properties

Size	Basis weight, (17 by 22 - 500 sheets) tolerance, each direction, minimum	Folding endurance, double folds	Bursting strength, average (minimum)
1	28	450	Points 60
2	20	300	36

Basis weight in accordance with TAPPI Standard T4 rance shall be tested in rd T423. Bursting strength accordance e with TAPPI Standard T403. shall be te

Acidity (pH e paper for the envelopes shall be not less ested in accordance with TAPPI Standard T4

Stability. folding endurance of the paper for the en not more than 30 percent when tested as s

Accelerate elope shall show no signs of deteriorat film shall show no signs of fading, da loration when tested as specified below.

Constructi hall be made in the center seam, open-end s ll be an open end along one short dimension. and dimensions of Size 1 envelope shall be s e 1. Style, design, and dimensions ope shall be as shown on Figure 2. The inside hall be gummed. The gum shall be of such q subjected to accelerated aging test it s tacky and unusable. The seams shall be by adhesive extending the whole length of

Printing. e envelopes shall be as shown on the figur able size.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Ruling and writing properties. Lines ruled and characters written on the envelopes shall be sharp, clear, and legible, and shall not feather when tested as specified below.

Workmanship. The envelopes shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Hundred (HD). One hundred envelopes of one size only, as specified, constitutes one unit.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified. Unit of product for examination purposes shall be one envelope as specified. Sampling shall be inspection level II with an Acceptable Quality Level (AQL) of 1.0.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified. Unit of product for test purposes shall be one envelope. Sampling shall be inspection level S-2 with an AQL of 1.0.

Atmospheric conditions. For the basis weight, bursting strength, folding endurance, and acidity tests, samples shall be conditioned in accordance with TAPPI Standard T402.

Stability test. The envelope shall be heated in dry air for 72 hours at 100° C and then subjected to the folding endurance test. The envelope shall not be conditioned for this test in accordance with TAPPI Standard T402.

Accelerated aging test.

Test film. The film used for test shall be processed in standard commercially available developer and fixer. The film shall be either dental radiographs or dental films exposed and processed to give a density of 0.5 to 1.5.

Procedure. The test films shall be inserted in a dental film mount of the proper size and inserted in the test envelope. The envelope shall then be bent back slightly and tied with string in order to insure contact between the inner surface of the envelope and the film and the mount.

Accelerated aging shall be achieved by incubating the test material at 50° ± 3° C and 74 ± 2 percent relative humidity for 10 days. These test conditions can be achieved by placing the specimen in a sealed chamber containing a saturated solution of sodium chloride, with temperature maintained as specified. A fan to circulate the air within the chamber will insure an even distribution of humidity.

Ruling and writing properties test. Using a 1-percent solution of either Color Index 18050, C.I. Acid Red, or Color Index 42085 C.I. Acid Green 3, apply ruling and written characters to the envelopes with a fine printing pen.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026l0-26T09:39Z
Check the source to verify that this is the currentversion before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

NOTE: The following National Stock Numbers and their respective Item Identifications are covered by this document:

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION
7530-00-890-1015	ENVELOPE, RADIOGRAPHIC FILM MOUNT, DENTAL, 5-1/2 by 9-1/4 inches
7530-00-890-1016	ENVELOPE, RADIOGRAPHIC FILM MOUNT, DENTAL, 3-1/8 by 5-1/2 inches

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:


DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project Number: 7530-0212
Location: c:\word\doc\licht\envelope\sg

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.




NSN 7530-00-890-1015

NOTE:
ALL DIMENSIONS IN INCHES.

FIGURE 1 ENVELOPE, SIZE 1

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.



NSN 7530-00-890-1016

NOTE:

ALL DIMENSIONS ARE IN IN

FIGURE 2 ENVELOPE, SIZE 2

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-54886	930630

3. DOCUMENT TITLE
ENVELOPE, RADIOGRAPHIC FILM MOUNT, DENTAL

4. NATURE OF CHANGE *(Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)*

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)
	e. DATE SUBMITTED (YYMMDD)

7. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code) (1) Commercial	(2) AUTOVON
Defense Personnel Support Center ATTN: DPSC-MSE	(215)7372870	444-2870
c. ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

DD Form 1426, OCT 89 — *Previous editions are obsolete.*

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:39Z
Check the source to verify that this is the current version before use.