A-A-51729A
15 June 1991
SUPERSEDING
A-A-0051729(DM)
24 October 1986

COMMERCIAL ITEM DESCRIPTION

BANDAGE, GAUZE, (COTTON, ELASTIC, SPUN RAYON, AND POLYESTER)

The General Services Administration has authorized the use of this Commercial Item Description.

Salient Characteristics:

Shall be an individually sealed, sterile, rolled, rubberless, elastic gauze bandage, 4-1/2 yards (4.1 meters) minimum average length, stretched, suitable for securing surgical dressings. The minimum unstretched length of a roll shall be 58 inches.

Bandage shall be nontoxic, nonirritating and virtually lint free and shall be additionally suitable for use as a dressing.

Bandage shall be elastic in the warp direction and shall be capable of causing compression when applied. Bandage shall have a minimum stretch (elongation) of 140 percent in the warp direction.

Bandage shall be fabricated of individual yarns of cotton, rayon and polyester. Warp yarns shall be of alternating rows of polyester yarns and spun rayon. Filler shall be of 2-ply cotton yarn for absorption of fluids. Bandages shall have a minimum of 9 filling yarns per inch in the stretched condition and a minimum of 30 warp yarns per inch unstretched.

Finished bandage shall be one continuous piece with finished woven edges, free of loose threads or ravelings and shall be rolled with the edges evenly aligned.

Bandages shall not slip, loosen or bind up during use.

SIZE	DIMENSIONS (Stretched)	PER ROLL (GRAMS)
1	3 inches (± 1/4") by 4-1/2 yards (± 5%)	7
2	4 inches (± 1/4") by 4-1/2 yards (± 5%)	9-1/2

AMSC\NA FSC 6510

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

The material components of the bandages shall be, 42% cotton, 34% rayon and 24% polyester.

Workmanship. The bandages shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Package (PG). Each bandage shall be individually sealed in an easy peel-open immediate container capable of maintaining sterility of contents unless package is damaged or opened. One package containing 12 individually sealed bandages of one size only, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplied and services conform to prescribed requirements.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for examination purposes shall be one bandage. Sampling shall be Inspection level II with an acceptable quality level (AQL) of 2.5 (percent defective).

Examination. Examination shall be conducted to determine compliance with specification requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

Salient characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for test purposes shall be one bandage. Sampling shall be Inspection level S-1 with an AQL of 2.5 (percent defective).

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric Products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerance specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION	
6510-01-230-8702	BANDAGE, GAUZE, COTTON, ELASTIC, Spun Rayon, and Polyester Yarns, Lint Free, 3 inch by 4-1/2 yards (7.6 cm by 4.1 meters), 12s	1
6510-01-171-5130	BANDAGE, GAUZE, COTTON, ELASTIC, Spun Rayon, and Polyester Yarns, Lint Free, 4 inch by 4-1/2 yards (10.2 cm by 4.1 meters), 12s	2

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:
DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS
PHS
FDA-MPQAS

PGC 07028

Project No. 6510-0780

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8 In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-51729A	15 JUNE 1991

3. DOCUMENT TITLE
BANDAGE, GAUZE, (COTTON, ELASTIC, SPUN RAYON, AND POLYESTER)

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code) (1) Commercial	(2) AUTOVON
efense Personnel Support Center	(215)737-2870	444-2870
c. ADDRESS (Include Zip Code) ATTN: DPSC-MSE 2800 S. 20th Street Philadelphia, PA 19101	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

DD Form 1426, OCT 89 Previous editions are obsolete. 198-290

Source: https://assist.dla.mil -- Downloaded: 202640-26109:00Z
Check the source to verify that this is the current version before use.