A-A-53262
1 March 1988

COMMERCIAL ITEM DESCRIPTION

PAD, NONADHERENT

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Military Specification MIL-P-36861 which is cancelled.

This Commercial Item Description covers sterile, nonadherent pads.

Salient Characteristics:

The nonadherent pads shall be of the following sizes and absorbencies.

SIZE	DIMENSIONS (INCHES)	Minimum Pad Absorbency (Grams of WATER)
1	2 ± 1/8 by 3 ± 1/8	4.7
2	3 ± 1/8 by 4 ± 1/8	8.0
3	3 ± 1/8 by 8 ± 1/8	15.0

Shall be a sterile, nonadherent, absorbent surgical dressing pad suitable for use as a dry dressing for wounds. The absorbent filler shall be completely enclosed by a cover except that two opposite edges may be cut. The cover shall be attached to the absorbent filler in such a way as to prevent the filler from sliding out of its cover during use. For testing the finished pad absorbency, weigh each of five dry pads, then immerse each pad in distilled water for 20 seconds, then pick each pad up by one corner with tweezers and allow to drain for 20 seconds and then reweigh each pad. The difference in weight is the amount of water absorbed. The test result is the average of the individual results of the five pads.

AMSC/NA FSC 6510

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

PREPARING ACTIVITY:

The absorbent filler shall be nontoxic and shall be composed of absorbent cotton, absorbent rayon or absorbent cellulose or a combination thereof. The filler shall be a nonwoven fabric or shall be needle loomed. The absorbency rate of the filler shall not exceed 15 seconds when tested in accordance with Purified Cotton, USP, and the filler shall not disintegrate or fray after standing in water for 24 hours. The filler shall have a minimum weight of 75 grams per square yard when tested in accordance with method 5041 of Federal Test Method Standard No. 191 and the filler shall have a minimum thickness of 30 mils when measured in accordance with method 5030 of Federal Test Method Standard No. 191.

The cover shall be a foraminous plastic film or a foraminous nonwoven fabric. The cover shall have openings comprising a minimum open area of 15 percent of the cover area. The openings shall be closely spaced and uniformily distributed over the entire cover area. Wound exudates shall readily pass through the cover and be absorbed by the filler.

The pad cover shall be smooth, inert, nontoxic, insoluble in the wound and shall not adhere to the wound.

Workmanship. The Nonadherent Pads shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Package (PG).

Each pad shall be individually packaged in a sealed peel-open type envelop capable of maintaining sterility of the pad unless package is damaged or opened.

For Size 1 and Size 2, one package containing 100 individually sealed pads of one size only, as specified, constitutes one unit.

For Size 3, one package containing 50 individually sealed pads, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:17Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item. Testing and examination shall be conducted to determine compliance with all specification requirements. Test Methods shall be suitable, accurate and reproducible.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for examination shall be one nonadherent pad. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5 maximum percent defective.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:17Z
Check the source to verify that this is the current version before use.

Metric Products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION	SIZE
6510-00-018-6184	Pad, Nonadherent, 2 by 3 inches, 100s	1
6510-00-111-0708	Pad, Nonadherent, 3 by 4 inches, 100s	2
6510-00-986-2942	Pad, Nonadherent, 3 by 8 inches, 50s	3

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PGC 05208

Project No. 6510-0696

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-53262	PAD, NONADHERENT

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR ☐ USER ☐ MANUFACTURER ☐ OTHER *(Specify)*:
b. ADDRESS *(Street, City, State, ZIP Code)*	

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-A[redacted] RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.