A-A-53882
10 August 1989

COMMERCIAL ITEM DESCRIPTION

ELEVATOR, EAR, Heavy, Curved Blade

The General Services Administration has authorized the use of this Commercial Item Description.

This commercial item description covers a heavy, curved ear elevator for use in surgical procedures.

Salient Characteristics:

Shall be a heavy, curved elevator suitable for use during ear surgical procedures. Shall have longitudinal grooves on the handle with a thumb rest to provide for a positive grip.

Style, design and dimensions of the elevator shall be in accordance with Defense Personnel Support Center (DPSC) Drawing No. 22318. The elevator shall be of two-piece construction, consisting of an operative end (blade) and handle. The operative end, with shank, shall be securely welded to the handle.

Material. The operative end (blade) and shank shall be fabricated of class 4, Type 420, martensitic corrosion-resisting steel, as specified in ASTM F899 for composition only.

The elevator handle shall be fabricated of Class 3, Type 302 or 304, austenitic corrosion-resisting steel, as specified in ASTM F899 for composition only.

Hardness. After proper stress relieving, the Rockwell hardness of the operative end shall be C45 to C54, as specified herein. The hardness of the remainder of the instrument shall be suitable for the purpose intended.

Corrosion resistance. The entire elevator shall be subjected to the copper sulfate and boil tests, as specified herein.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

Finish. The operative end shall have a clean, sharp, beveled edge on inside of curved blade. Finish of the handle and shank shall be matte or bright. The matte finish shall exhibit a uniform nondirectional surface texture equivalent to that produced by lightly blasting the prepared surface with fine abrasives under pressure.

Identification marking. The elevator handle shall be marked with the manufacturer's registered trademark and part number. Elevator of foreign manufacture shall bear the name of the country of origin.

Workmanship. The elevator shall be free from coarse grind mark, feathers, pits, cracks or any other defects which may detract from its appearance or impair its serviceability.

Unit. Each (EA). One elevator, as specified, constitutes one unit.

Unit package. Each elevator shall be wrapped in neutral tissue or other suitable anti-corrosive material and shall be enclosed in a close fitting container. The elevator shall be so packaged to prevent damage to the operative end.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection, examination and test purposes shall be one elevator. Unless otherwise specified, sampling for examination shall be conducted at inspection level, S-II, except for dimensions which shall be conducted at inspection level S-1, and sampling for tests (when applicable) shall be conducted at inspection level S-1. For tests, acceptance number shall be zero (0) for all sample sizes.

Examination. Examination shall include but shall not be limited to the following defects.

Defects.

Style, design and dimensions of the elevator not as shown on applicable drawing.
Operative end not free from nicks, burrs, feathers, pits, or corrosion.
Finish on operative end not as specified.
Instrument not free from cracks.
Finish on handle not as specified.
Identification marking not present or not permanent, legible, and complete.

Tests.

Chemical analysis. In the absence of certificates of quality, the chemical analysis of the instrument shall be determined in accordance with method 111 or 112 of FED-STD-151. In the event of a conflict, method 111 shall govern.

Hardness. The hardness shall be determined in accordance with ASTM E18 or E92.

Corrosion resistance. The boil test and copper sulfate test shall be determined in accordance with ASTM F1089.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerance specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable . The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling etc., involving shipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document.

National Stock Number	Item Identification
6515-01-139-8925	ELEVATOR, EAR, Heavy Curved Blade, 7-1/2 mm by 25 mm

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

AGENT:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

52203

Project No. 6515-4626

Location- NCR Disk 1

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.