A-A-53886
19 June 1989

COMMERICAL ITEM DESCRIPTION

PHOSPHORIC ACID GEL, DENTAL

The General Services Administration has authorized the use of this Commerical Item Description.

This commercial item description covers a dental phosphoric acid gel.

Salient Characteristics:

Material. The gel shall be 34 ± 1.5 percent phosphoric acid when tested as specified herein.

General. The phosphoric acid gel shall be designed for use in the acid etch dental procedure.

Physical properties. The gel shall be water soluble such that it can be completely removed from the tooth enamel by rinsing with water. The gel shall be a thixotropic gel and shall be colored so as to markedly contrast with tooth enamel.

Flow test. The gel shall not change shape or flow for one minute when tested as specified herein.

Consistency. The gel shall have a consistency of not less than 20 mm and not more than 30 mm when tested as specified herein.

Stability. The shelf life of the gel shall be 24 months from the date of manufacture and no less than 20 months shall remain from the date of delivery to the Government. Each unit package shall be marked with the expiration date.

Instructions. An instruction sheet detailing the proper use of the gel shall be provided.

Workmanship. The phosphoric acid gel shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing one or two bottles for a total of 12 ml (minimum) of phosphoric acid gel as specified, constitutes one unit. Each package shall be furnished with instructions for use.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of II (percent defective) and an inspection level of 2.5. The unit of product for examination purposes shall be one bottle as specified in this description.

Examination. Examination shall include but shall not be limited to the following:

Defects

Gel does not flow when pressure is applied to the bottle during emptying.
Gel flows without being stirred or extruded from bottle.
Color not as specified.
The gel shall be free of all dirt, foreign matter or other contaminants.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

Sampling for tests. Sampling for tests, shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of S-2. The unit of product for test purposes shall be one bottle.

Tests. Prior to performing the flow and consistency tests the gel and equipment used shall be conditioned at 73° ± 2° F and 52.5 ± 2.5% relative humidity for 10 hours.

Flow test. An amount of gel sufficient to cover a tooth shall be applied to a smooth vertical surface and observed for one minute.

Consistency. Extrude 0.5 ml of gel into a glass tube containing an extruder. Extrude the material onto a 2"x 2" cellulose acetate sheet. Cover the material with a second 2"x 2" cellulose acetate sheet. Place a 70 gram weight onto the sheet for one minute. Remove the weight and the plate and measure the major and minor diameters of the gel. The average of the measurements shall be the consistency of the gel.

Composition. Take a 1.5 gram sample of the gel and mix with 60 ml of deionized water. Titrate the solution with 1 N NaOH to a violet end point using phenolphthalein as an indicator.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pounds units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerance specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling etc., involving shipment from the supply source to the receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

National Stock Number	Item Identification
6520-01-282-2246	PHOSPHORIC ACID GEL, DENTAL

Ordering data. Intermediate/exterior package quantities, labeling and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.

CIVIL AGENCY COORDINATNG ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project Number: 6520-2225

Location: NCR DISK 1

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:26Z
Check the source to verify that this is the current version before use.